

Mass Spectrum SmartFormula Report

Analysis Info


Analysis Name E:\Data2\Kilian\KC-mNH2(K)pCF3000001.d
Method Tune_low_pos.m
Sample Name KC-mNH2(k)pCF3
Comment

Acquisition Date 2017-09-04 15:23:07

Operator Carin Larsson
Instrument / Ser# microTOF 125

Acquisition Parameter

Source Type	ESI	Ion Polarity	Positive	Set Nebulizer	0.3 Bar
Focus	Not active			Set Dry Heater	180 °C
Scan Begin	50 m/z	Set Capillary	4500 V	Set Dry Gas	4.0 l/min
Scan End	1000 m/z	Set End Plate Offset	-500 V	Set Divert Valve	Source



Meas. m/z	#	Formula	Score	m/z	err [mDa]	err [ppm]	mSigma	rdb	e ⁻	Conf	N-Rule
266.0783	1	C 14 H 11 F 3 N O	100.00	266.0787	0.4	1.5	12.5	8.5	even		ok