

# Mass Spectrum SmartFormula Report

## Analysis Info


Analysis Name E:\Data2\Kilian\oNH2KCy000001.d  
Method Tune\_low\_pos.m  
Sample Name oNH2KCy  
Comment

Acquisition Date 2018-02-15 16:20:33

Operator Carin Larsson  
Instrument / Ser# micrOTOF 125

## Acquisition Parameter

Source Type	ESI	Ion Polarity	Positive	Set Nebulizer	0.3 Bar
Focus	Not active			Set Dry Heater	180 °C
Scan Begin	50 m/z	Set Capillary	4500 V	Set Dry Gas	4.0 l/min
Scan End	1000 m/z	Set End Plate Offset	-500 V	Set Divert Valve	Source



Meas. m/z	#	Formula	Score	m/z	err [mDa]	err [ppm]	mSigma	rdb	e <sup>-</sup>	Conf	N-Rule
226.1190	1	C 13 H 17 N Na O	100.00	226.1202	1.2	5.3	26.5	5.5	even		ok