

# Mass Spectrum SmartFormula Report

## Analysis Info


Analysis Name E:\Data2\Kilian\pBrKCy000001.d  
Method Tune\_low\_pos.m  
Sample Name pBrKCy  
Comment

Acquisition Date 2018-02-19 10:47:49

Operator Carin Larsson  
Instrument / Ser# microTOF 125

## Acquisition Parameter

Source Type	ESI	Ion Polarity	Positive	Set Nebulizer	0.3 Bar
Focus	Not active			Set Dry Heater	180 °C
Scan Begin	50 m/z	Set Capillary	4500 V	Set Dry Gas	4.0 l/min
Scan End	1000 m/z	Set End Plate Offset	-500 V	Set Divert Valve	Source



Meas. m/z	#	Formula	Score	m/z	err [mDa]	err [ppm]	mSigma	rdb	e <sup>-</sup>	Conf	N-Rule
289.0190	1	C 13 H 15 Br Na O	100.00	289.0198	0.8	2.9	63.8	5.5	even		ok