

Mass Spectrum SmartFormula Report

Analysis Info


Analysis Name E:\Data2\Kilian\KC-10-052000001.d
Method Tune_low_pos.m
Sample Name haloperidol-frag
Comment

Acquisition Date 2019-02-07 10:27:59

Operator Carin Larsson
Instrument / Ser# micrOTOF 125

Acquisition Parameter

Source Type	ESI	Ion Polarity	Positive	Set Nebulizer	0.3 Bar
Focus	Not active			Set Dry Heater	180 °C
Scan Begin	50 m/z	Set Capillary	4500 V	Set Dry Gas	4.0 l/min
Scan End	1000 m/z	Set End Plate Offset	-500 V	Set Divert Valve	Source



Meas. m/z	#	Formula	Score	m/z	err [mDa]	err [ppm]	mSigma	rdb	e ⁻	Conf	N-Rule
205.0642	1	C 10 H 11 F Na O 2	100.00	205.0635	-0.7	-3.3	19.1	4.5	even		ok