

Display Report

Analysis Info

Analysis Name E:\Data2\Kilian\KC-10-82000001.d
Method Tune_low_pos.m
Sample Name Ketoprofen13C
Comment

Acquisition Date 2019-02-06 16:24:21

Operator Carin Larsson
Instrument / Ser# microTOF 125

Acquisition Parameter

Source Type	ESI	Ion Polarity	Positive	Set Nebulizer	0.3 Bar
Focus	Not active			Set Dry Heater	180 °C
Scan Begin	50 m/z	Set Capillary	4500 V	Set Dry Gas	4.0 l/min
Scan End	1000 m/z	Set End Plate Offset	-500 V	Set Divert Valve	Source

