



Luce  
Dipartimento Di ingegneria  
Università di Ferrara  
Via Saragat, 1  
44122 Ferrara

## Dielectric strength test

REF. FILE							
OPERATOR	Eng. Daniele Pareschi				DATA	27-04-2020	
TEMP.	23	°C	HUMIDITY	48	%	P.ATMOSPHERIC	1011 hPa

Test Site: ☒ LAB1  
E.U.T.: RESPIRATORE AUTOMATICO  
Manufacturer: Centro IT- UNIFE  
Serial number: -  
Standard reference: EN 60601-1:2006, EN 60601-1:2006 /A1:2013, EN 60601-1:2006 /A1:2013/AC:2014, EN 60601-1:2006/A12:2014  
Test equipment: ☒ LAB-064a ☒ APP-022  
Remark: -

Operator's signature

Application points	Test current type	Test voltage	Rising time	Test duration	Verdict
Between 230 V plug and outlet plug of the power supply	AC 50 Hz	3000 V	10 s	60 s	PASS
Between 230 V plug and the accessible metal parts of the device	AC 50 Hz	3000 V	10 s	60 s	PASS