

Mass Spectrum SmartFormula Report

Analysis Info


Analysis Name E:\Data2\Raul AM\rmm176000001.d
Method Tune_low_pos.m
Sample Name rmm176
Comment

Acquisition Date 2016-04-01 12:04:28

Operator Carin Larsson
Instrument / Ser# micrOTOF 125

Acquisition Parameter

Source Type	ESI	Ion Polarity	Positive	Set Nebulizer	0.3 Bar
Focus	Not active			Set Dry Heater	180 °C
Scan Begin	50 m/z	Set Capillary	4500 V	Set Dry Gas	4.0 l/min
Scan End	1000 m/z	Set End Plate Offset	-500 V	Set Divert Valve	Source



Meas. m/z	#	Formula	Score	m/z	err [mDa]	err [ppm]	mSigma	rdb	e ⁻	Conf	N-Rule
253.0657	1	C ₁₄ H ₁₄ NaO ₂ S	100.00	253.0658	0.0	0.1	206.4	7.5	even		ok

