

# Mass Spectrum SmartFormula Report

## Analysis Info


Analysis Name C:\Users\amlab\Desktop\Kilian\2KC-5-51Ket000001.d  
Method APCI\_2015\_Tune\_low\_pos.m  
Sample Name KC-5-51Ket  
Comment

Acquisition Date 2017-03-28 17:24:19

Operator Carin Larsson  
Instrument / Ser# micrOTOF 125

## Acquisition Parameter

Source Type	APCI	Ion Polarity	Positive	Set Nebulizer	2.0 Bar
Focus	Not active			Set Dry Heater	200 °C
Scan Begin	50 m/z	Set Capillary	3000 V	Set Dry Gas	8.0 l/min
Scan End	1600 m/z	Set End Plate Offset	-500 V	Set Divert Valve	Source



Meas. m/z	#	Formula	Score	m/z	err [mDa]	err [ppm]	mSigma	rdb	e <sup>-</sup> Conf	N-Rule
189.1260	1	C 13 H 17 O	100.00	189.1274	1.4	7.5	10.3	5.5	even	ok