in passing the cystoscope. It gives perfect security against extraneous scratches, and it permits the removal of the wax-tipped catheter for examination without previously withdrawing the cystoscope. The latter is a great advantage. Several control waxed tips may be passed one after the other, thus confirming the findings or accurately measuring the distance of the stone from the ureteral orifice. At the same time, and through the same cystoscope, an ordinary ureteral catheterization followed by functional study, ureteropyelogram or even both may be made, thus obtaining complete information in the case at the one cystoscopy.

Fig. 4.—Catheter sheath fastened in place by rubber tip. The elevator of the instrument has been raised, showing how the rubber sheath protects the waxed tip.

The operating cystoscope (Brown-Buerger is used, but the method is applicable with any operating instrument) is passed into the bladder as for ordinary cystoscopy, the obturator withdrawn, the bladder washed out (the irrigating fluid should not be too warm, as it may soften or melt the wax), and the operating sheath of the instrument then inserted. The wax-tipped catheter has been provided with a soft rubber sheath, which should easily pass through the operating barrel of the instrument, as shown in Figures 1 and 3. The usual rubber tip (Fig. 2 A) for the irrigating instrument is too small to cover the orifice of the telescope when the slit ends of the rubber sheath are turned back. A very satisfactory tip may be quickly made, as indicated in Figure 2 B, C, D and E. In fact, this tip will be found to be much more satisfactory for all work with the operating cystoscope than the hard, friable tips on the market. The method of passing the wax-tipped catheter is indicated in Figure 4. The sheath with its catheter inside (to within 1 inch of the end) is threaded into the instrument until the end of the sheath appears in the visual field just beyond the end of the elevator of the instrument. The slit portions of the other end of the sheath are then spread to the opposite sides and covered by the rubber tip, which thus securely holds the sheath in place and insures dryness. The waxed tip is now carefully pushed into the field (Fig. 5), the elevator, which strikes the rubber sheath and not the catheter is raised, and the catheter then pushed on up safely by the beak of the instrument into the ureter (Fig. 4).

Fig. 5.—Waxed tip and rubber sheath before raising the elevator of the cystoscope.

In removing the wax-tipped catheter, after passing it up the ureter as far as desired, the lever of the instrument should be kept partially elevated; the waxed end can then be withdrawn into its sheath with no danger of striking the beak of the instrument. It is preferable to withdraw the sheath with the catheter in it, rather than to withdraw the catheter through the entire length of the sheath, for examination. In this way the waxed tip will have passed only a very short distance through its sheath, being originally retrograded into it to a distance of about 1 inch, and, before examination and after having passed up the ureter, not being withdrawn into it to a greater distance than this.

Physicians Building.

A NEW LID EVERTER

SIDNEY ISRAEL, M.D., HOUSTON, TEXAS

This small lid everter is the outcome of an endeavor to find an instrument that would give a large exposure of the entire upper lid and culdesac. Its most salient features are simplicity of design, ease of manipulation, and freedom from causing pain in introduction. It has been the means of revealing several cases of trachoma that had been overlooked by ordinary methods.

In introducing the instrument, it is grasped at its distal end between the tips of the thumb and index finger. With the patient looking down, the lashes of the upper lid are held in the tips of the thumb and index finger of the left hand. The knobbed ends of the instrument are then placed gently on the upper lid about one-half inch from the ciliary margin with the instrument in the vertical position. With the fingers of the left hand holding the lashes, the lid is retracted over the knobbed ends, and with gentle pressure on the distal end, the entire under surface of the upper lid, including the culdesac, is freely brought into view. In this position the knobbed ends can be slid from the outer to the

Fig. 1.—New lid everter.

Fig. 2.—Method of using the lid everter.

inner canthus, exposing any part desired. In this position also, with the under surface of the lid exposed, the hold on the eyelashes by the fingers of the left hand can be released, and with the tip of the index finger of the right hand on the distal end still maintaining the position of the instrument in the vertical meridian, the thumb of the right hand can be dropped down to hold the lashes in place to maintain a complete eversion of the lid, while with the left hand free, any foreign body that is present on the under surface of the lid is easily removed.

The instrument can be readily carried in the pocket, and because of facility of introduction, can be used universally by the general practitioner and the ophthalmologist.

IMPROVED FORM OF BRAIN EXPLORER

HAROLD GIFFORD, M.D., OMAHA

In the *Laryngoscope*, January, 1914, I described a brain explorer consisting essentially of two knife blades held together with a spring, to be used in cases of suspected brain abscess in place of the exploring needle or scalpel. As is well known, each of the latter instruments may pass through the wall of an abscess, and either on account of thickness of the wall or thickness of the pus, may not reveal its presence. The practice of passing in dressing forceps and opening them here and there in the brain can hardly be called either accurate or elegant. With this brain explorer,