DISCUSSION.

Mr. E. Hatschek asked if the author had seen a paper by Friedlaender¹ on the segregation of liquid mixtures near the critical temperature. This work was carried out in a thermostat in which the temperature could be maintained constant within 0.01°C .

Dr. Lewis said he knew that such work had been done, but he had not read the paper in question.

Mr. Hatschek asked whether toluene had been found satisfactory.

Dr. Lewis said he had found it perfectly satisfactory.

Mr. Hatschek said he had found calcium chloride more successful than toluene.

Dr. Lewis said there was one thing to be said with regard to toluene. He was using for other purposes the purest toluene, free from thiophene, which he could obtain commercially and he re-distilled it two or three times, and some of this was used for filling the regulator. That might have something to do with the fact that it proved so satisfactory. In any case, there had been no trouble whatever.

Dr. Pring said there was a point in connection with the relay. The thermostat he had been using had a relay of similar construction to that shown, with the exception that in place of a mercury break he had graphite plates 8 mm. in diameter with surface slightly convex and with a small condenser across the spark gap. In that way, in place of an arc he got a small high tension spark which did not cause any disintegration and it worked very satisfactorily. He would like to know if Dr. Lewis had tried that device and how it compared with the one he had shown.

Dr. Lewis said he had not tried that.

The President suggested that iron terminals might be better than copper. Graphite points with the condenser across seemed a good idea.

Dr. Borns said the author had mentioned trouble with the heaters; he did not see why that should be so.

Dr. Lewis said there was considerable trouble with the heaters at first. The first suggestion with regard to them came from Professor Philip of South Kensington. He had a small heater which was the shape and size of a lunch-biscuit and he believed it worked very well; but in Professor Philip's case the thermostat was very much smaller and it was used under

¹ *Zeitschr. f. Phys. Chem.*, **38**, 385, 1901.

DISCUSSION

the influence of only ordinary temperatures. These heaters lasted very well when used for only comparatively small currents. In his own case, however, he was working under great pressure both as to time and temperature and these elements broke down rather rapidly. The chief reason, perhaps, was that the uralite or asbestone or asbestos, whichever was used, became dried, possibly altered in composition and then broke up, and then the wires touched. He believed the chief difficulty had reference to that. He then had some plates of fused quartz made by the Thermal Syndicate and they were fairly satisfactory. However, they unfortunately broke and then he was in trouble again. Eventually, he devised the plan of having a strip of copper between strips of uralite, winding the wire round that, and putting a casing of uralite outside that. In the newer ones, the copper projected at each end and was screwed on to little nuts which were soldered on to the inside of the casing. These had given no trouble at all,