obtained by the mastic and Wassermann reactions on the 200 cases.

In fourteen cases there was a positive mastic and a negative Wassermann reaction. Eleven of these had increased cell counts together with all the clinical signs of paresis and were so diagnosed. Two were diagnosed as cerebrospinal syphilis. The other case giving a positive mastic and a negative Wassermann reaction was that of a man of 50, who two and a half years before entering the hospital had had a "stroke," leaving him with a right-sided paralysis. On entering the hospital he had a mitral insufficiency with casts and albumin in the urine, and died one month later as a result of the heart and kidney complication. He gave a definite history of alcohol with an indefinite history of syphilis. The spinal fluid examination showed a cell count of 10 with an increase of globulin. In no case was there a positive Wassermann and a negative mastic reaction. In two cases in which the mastic test was always positive, the Wassermann test was at first negative, but later became positive.

Of the seven cases of epilepsy, two showed a very slight precipitation in the first tube. Two of the fifty cases of dementia praecox showed the same reaction,

TABULATION OF TWO HUNDRED MASTIC EXAMINATIONS

Classification	Number of Cases	Mastic	Wassermann
Paresis	68	+	+
Paresis	11	+	+
Cerebrospinal syphilis	2	+	+
Cerebrospinal syphilis	2	+	+
Tabs	1	+	+
Dementia praecox	50	—	—
Manic-depressive psychosis	31	—	—
Arteriosclerotic brain disease	6	—	—
Alcoholic psychosis	6	—	—
Senile psychosis	5	—	—
Involution melancholia	2	—	—
Paranoia	3	—	—
Mental defective	2	—	—
Autointoxication	1	—	—
Epilepsy	7	—	—
Traumatic psychosis	1	—	—
Alcoholism	1	—	—
Cerebral hemorrhage	1	±	—

but these were not to be compared with the precipitation in a positive case. In one patient diagnosed as having cerebral hemorrhage, there was a ± in the first two tubes.

Thus far my work has been confined to psychiatric patients. What the action will be in such diseases as poliomyelitis and tuberculous meningitis must be further investigated.

SUMMARY

The mastic test, when taken together with the history and with the cell count, is of undoubted aid in the diagnosis of syphilis of the nervous system.

By the addition of potassium carbonate, the degree of positiveness can be graded very accurately.

By first incubating and then centrifugalizing, the test can be completed in two hours.

It parallels quite closely the colloidal gold reaction, and is more easily interpreted and much more easily and quickly performed.

In eighty-four cases of syphilis of the nervous system, the mastic test was uniformly positive.

Workmen's Compensation.—Thirty-seven states, the territories of Alaska, Hawaii, and Porto Rico have enacted compensation laws, the United States adopting a like principle for the relief of its civilian employees under date of Sept. 7, 1916.—National Industrial Conference Board.

Clinical Notes, Suggestions, and
New Instruments

METHODS FOR THE COLLECTION AND TRANSPORT OF
BODY FLUIDS FOR BACTERIOLOGIC INVESTIGATION

ALFRED H. CAULFEILD, M.B.,* TORONTO, ONT.
Captain, Canadian Army Medical Corps

The exigencies of war have made the bacteriologic investigation of the body fluids even more imperative than during times of peace, while at the same time the accommodations

Fig. 1.—A, trocar and cannula; B, collection tube or flask of quartz glass.

surrounding the soldier-patient are frequently less adaptable to bacteriologic technic than they are in civilian practice. Consequently it seems desirable at the present time to record any advantageous modifications in the instruments and apparatus used for the collection and transport of such fluids. Improvements to be of value should simplify the technic at the bedside, should insure safety and sterility during transport, and should not necessitate elaborate methods of preparation. Indeed, it is most desirable that the value and the result of the final investigation should not be dependent on any special care or expertness on the part of the nurse or lay attendant responsible for the original sterilization or for the transport of the material. In this connection it may be stated that devices by which sterilization may be carried out in the dry oven or autoclave, and which will leave without further manipulation the apparatus or instrument ready for transport, and convenient for use at the bedside, are the most satisfactory and trustworthy. Particularly is this true for those enjoying the facilities of organized hospitals and institutes, as it relieves them of all irksome detail or improvisation. For such metal articles as needles, trocar-cannulas, etc., dry sterilization is preferable,

Fig. 2.—Syringe.

in that any possibility of rusting and blocking of the bore is avoided. It is along these lines and with these points in view that I have attempted to develop the technic.

The general satisfaction I have experienced with the method herein described (which is an amplification of those previously used by me) is to a great extent due to the use of quartz glass for certain parts of the apparatus; and as the use of this material, so far as my knowledge or information goes, is not at all general with the profession, I venture

* Detailed for special investigation from No. 4 Canadian General Hospital (University of Toronto).

to present these methods in full, rather than serve them piecemeal in possible future publications. The particular advantage of quartz glass or silica lies in the fact that it has no coefficient of expansion, and consequently will not break or crack in extremes of heat and cold.

COLLECTION OF MATERIAL (SUCH AS CEREBROSPINAL FLUID)
BY MEANS OF TROCAR AND CANNULA

The danger of contamination in catching fluid, as it flows from the open end of a straight needle or cannula, as well as the difficulty of avoiding hand contamination incident to the puncture operation, led me to adopt the form of trocar and cannula shown in Figure 1 A. The proximal opening of this instrument is ground to fit the nozzle of a syringe, in case the amount of fluid to be drawn is small and not under pressure.

The collection tube or flask of quartz glass is shown in Figure 1 B. In my experience tubes of this nature, made of ordinary glass, fail to give satisfaction, and sooner or later break during sterilization. Preparation for sterilization consists in fitting and wiring to the upper arm a short piece of pressure rubber tubing, the open end of which is lightly packed with cotton, and in attaching to the lower arm a short piece of ordinary rubber tubing, the open end of which is plugged with a glass rod. This will later facilitate handling at the bedside. At the same time an extra piece of pressure tubing should be similarly packed at the ends with cotton. The collection flask thus assembled and the separate piece of pressure tubing should be properly wrapped and sterilized in the autoclave. As four pressure tubing clamps

Fig. 3.—Needle holders.

will be required after the tube is filled at the bedside, it is advantageous to tie these to one of the sterilized articles.

Technic at the Bedside.—The cannula and flask are carefully unwrapped, and the soft rubber tubing slipped on the side arm of the former. Any possibility of contamination can be avoided by the use of an alcohol flame. Puncture is made with the collecting tube connected in this fashion to the cannula; then the trocar is withdrawn and the tap turned off. If fluid has appeared at the tap arm opening, it will readily run through the closed passageway into the tube. In case the amount of fluid is so small that it does not flow, a sample may be obtained by fitting a syringe into the cannula.

When the puncture is made, the collecting tube, so to speak, takes care of itself; afterward it should be held upright. When the cannula is withdrawn, the tubing is pinched with the fingers, and afterward separated from the arm of the tube. The opening of the arm with any content of fluid thus exposed should be thoroughly flamed, and while hot slipped into the spare piece of pressure tubing, which can be additionally secured by copper wiring, if this is thought necessary. Finally, by tightly screwing down two clamps on each piece of tubing, these passageways are effectively blocked. The tube is now packed or carried in any position, and is proof against contamination taking place during transport.¹

COLLECTION OF MATERIAL (SUCH AS BLOOD) BY
HYPODERMIC SYRINGE AND NEEDLE

The general satisfaction and convenience that would be afforded, were it possible to sterilize in the dry oven syringe

1. Cleaning of these tubes is easily accomplished, as any particles or clots not easily washed out can be completely carbonized over a Bunsen or blast flame.

and needles already packed in their cases, led me to devise the syringe, needle holder and case illustrated.

That the parts of the syringe (Fig. 2) need not be separated during sterilization, the proximal end of the graduated quartz glass barrel is cemented to a metal collar corresponding in length to the plunger, and of slightly larger diameter than the plunger to allow for expansion. The collar thus provides, so to speak, a safety expansion chamber for the plunger, which is withdrawn into the chamber during the process of sterilization.

Fig. 4.—Case, and adjustment for sterilization.

As shown in the illustration of the needle holder (Fig. 3), the removal of either protection cap exposes the head of one of two needles. By this means adjustment of the needle of the nozzle of the syringe can safely be made at the bedside. The remaining cap allows one to handle the needle holder with ease and without risk of contaminating the remaining needle or the needle adjusted. In one of the caps there is a small aperture which before sterilization should be adjusted to a similar opening in the inner sheath of the holder to prevent the cap being forced off.

The illustration of the case, and adjustment for sterilization (Fig. 4) shows the syringe with piston extended through an opening in the end of the case, so that the plunger rests in the safety chamber. It is advisable to pack cotton around the needle holder and syringe before replacing the top and sterilizing. As soon as cooling has taken place after dry sterilization, the piston is shoved home into the barrel with

Fig. 5.—Flask for transport of blood in fluid state.

FLASK FOR TRANSPORT OF BLOOD IN FLUID STATE
To the side arm of an ordinary filter flask (containing citrate solution) is attached and wired a short piece of pressure rubber tubing, the open end of which is plugged with cotton (Fig. 5). The mouth of the flask is stoppered with a well fitting rubber cork, which is additionally secured by crossed copper wire. After sterilization in the autoclave, the tubing is blocked by two pressure tubing clamps and the cork paraffined.

Technic at the Bedside.—As everything within the syringe case is sterile, the syringe intact and ready for use, and as the head of the needle can be exposed with safety and ease,

the adjustment and use of the syringe and needle are very simple. Any possibility of contamination is fully excluded by the use of an alcohol flame. To discharge the blood from the syringe, the needle, on withdrawal from the vein and while still attached to the syringe, is heated and pressed through the cork of the flask. Both clamps on the rubber tubing should be released before depositing the blood, and afterward retightened. The flask may now be packed for transport.²

FLASK FOR DISTRIBUTION OF STERILE ENRICHMENT FLUIDS IN QUANTITATIVE AMOUNTS TO CULTURE TUBES

The subsequent bacteriologic investigation of the body fluids is so closely connected with their collection and transport that a brief reference to a device dealing with this

aspect may be regarded as within the scope and intention of this paper.

The purpose of the flask is to provide a sterile receptacle so that enriching fluids can easily and without risk of contamination be distributed quantitatively to tubes of the ordinary laboratory mediums. Some such method is almost a necessity if the work is not routine in character. In this part of the work, as well as in the collection of body fluids, simplification in the technical detail is the surest method of incorporating bacteriologic examinations with the practice of medicine.

The arrangement of the apparatus consists in attaching to a large filter flask a graduated tube by means of pressure rubber tubing as shown in Figure 6. The tightly fitting rubber cork of the flask is pierced by a cotton-plugged air vent, and the glass tubing (for siphonage) which is connected to the graduated cylinder. This cylinder plugged with

Fig. 6.—Flask for distribution of sterile enrichment fluids in quantitative amounts to culture tubes.

cotton at the top leads by means of pressure tubing to a hand revealed the Velpau, Sayre and Dessault dressings, and a few modifications of these, as the only hope which could be offered the patient for the correction of his injury.

2. Besides the individually desired materials for skin sterilization, very few articles need be carried to the bedside when the methods described above are used. Thus for the first type of operation the only necessary articles are: (1) wrapped sterile trocar and cannula; (2) wrapped sterile flask and extra piece of tubing, and (3) spirit lamp. For the second type: (1) sterile case of syringe and needles; (2) sterile flask containing citrate solution, and (3) spirit lamp. My experience with these two types of collecting flasks might be summed up by stating that all specimens of cerebrospinal fluid and blood, collected as described from seven patients, showed no contamination after transport from the North of Ireland to London; and that material shaken and immediately incubated remained sterile.

means, the flask thus fully assembled is sterilized before being filled through the side arm, after which this side passageway is properly blocked by a heated rod or tube and clamped. Once the flask is filled, siphonage is obtained by forcing air through the air vent and releasing the clamp to the graduated chamber. As the hooded pipet is made of quartz glass, there is no danger of cracking during the necessary use of the Bunsen flame before and after the withdrawal of the fluid.³

THE RATIONAL AND COMFORTABLE TREATMENT OF FRACTURE OF THE CLAVICLE BY MEANS OF A CRUTCH SPLINT

WALTER KELTON, M.D., SEATTLE

Not long ago there came under my observation a case of fracture of the clavicle wherein the Sayre dressing, or Velpau, could not be utilized. The patient, in addition to sustaining a fracture of the clavicle, had also received a severe trauma of the upper arm and forearm of the same side, causing great swelling at the elbow. The arm was fleshy, and this, together with the swelling, made it impossible to use any of the conventional dressings, as in order to do so it would have been necessary to keep the patient continually narcotized and to take chances of interference with the circulation.

A diligent search of the literature for some kind of apparatus or dressing which would take care of the case in

Fig. 1.—Side view of fracture crutch.

Fig. 2.—Rear view of patient with straps adjusted.

hand revealed the Velpau, Sayre and Dessault dressings, and a few modifications of these, as the only hope which could be offered the patient for the correction of his injury.

I constructed an apparatus which would perform the dual service of correcting the injury and alleviating the suffering.

The apparatus consists of a T shaped steel bar, on the horizontal or upper part of which rests a pneumatic tube covered over with chamois skin and which fits into the axilla (Fig. 1 A). The upright steel bar fits the side of the chest and extends downward to the hip (Fig. 1 B), where it is attached to an aluminum plate covered also with chamois skin and shaped to fit snugly over and rest on the hip. (Fig. 1 C). Holes placed equidistant in the upright bar (Fig. 1 D) render the apparatus adjustable to any adult. The bolt (Fig. 1 E) fastening the aluminum plate to the

3. A properly filled flask may quite safely be forwarded, provided all three exits are additionally clamped, and of course breakage guarded against by careful packing. Flasks of 1,000 c.c. of filtered horse serum have provided sterile enrichment fluid for culture tubes until emptied. This means that small amounts have been easily withdrawn on about 200 occasions, without the stock fluid becoming contaminated. The following precautions are suggested on the avenues of infection that might be overlooked: 1. In refilling the graduated tube, care should be taken that the fluid is not allowed to reach the cotton of the air vent. 2. The tube to which the enriching fluid is to be added should be thoroughly flamed before being inserted within the hood, so that there is no possibility of the discharging tip becoming contaminated from the inner wall of the tube. The test tube protecting the tip and blocking the mouth of the hood should be treated in the same way before being reinserted.