JOURNAL

OF THE

FRANKLIN INSTITUTE

OF THE STATE OF PENNSYLVANIA,

FOR THE PROMOTION OF THE MECHANIC ARTS.

VOL. CXXVI. SEPTEMBER, 1888.

No. 3.

THE FRANKLIN INSTITUTE is not responsible for the statements and opinions advanced by contributors to the JOURNAL.

WEIGHING MACHINES.

BY WM. KENT, M.E.

[*A lecture delivered before the FRANKLIN INSTITUTE, February 10, 1888.*]

MR. KENT was introduced by the Secretary of the INSTITUTE and spoke as follows:

MEMBERS OF THE INSTITUTE, LADIES AND GENTLEMEN:

I had the pleasure of hearing the lecture last Monday evening in this hall, and the lecturer said that it was customary to carry the subjects of lectures back to the book of Genesis and to Homer. My subject, I think, may be carried back to a still greater antiquity, for among the earliest questions asked by man of his fellow-man must have been "How much does it weigh?" And the answer would have been given after reference to the first weighing machine, which is man's

WHOLE NO. VOL. CXXVI.—(THIRD SERIES, Vol. xcvi.)

own hand, or his muscular strength. The first example of barter and sale may have included an offer like this from the seller: "For that piece of silver I will give you as much grain as you can lift from the ground, or for that smaller piece I will give as much as you can raise shoulder high." No doubt the weighing was done by the muscles of man before any more accurate machine was invented for the purpose.

A weighing machine is any machine or contrivance used to ascertain the heaviness of a body.

Scientifically speaking, *weight* is the result of the attraction of gravitation upon a body, and as the force of gravity is not the same at different parts of the earth's surface, so the weight of any piece or body, if by weight we mean the effect of gravity upon it, differs according to the place at which it is weighed. Thus the writers on natural philosophy tell us that a mass of iron which weighs 1,000 pounds at the Equator would weigh 1,005 pounds at the Pole; 500 pounds at a point 2,000 miles below the surface or 1,650 miles above it; 2,600 pounds on the surface of Jupiter, and 28,000 pounds on the surface of the sun.

But as ordinarily used, weight does not mean the absolute heaviness of a body or the effect of gravity upon it, but the relative heaviness; that is, its heaviness as compared with that of a certain piece of metal which is taken as a standard, and weighed at the same place and under the same conditions.

The accepted standard in English speaking countries is the pound avoirdupois, which is the weight of a certain piece of metal kept in the custody of the warden of the standards in London. The standard for France and for all countries which have adopted the French system of weights and measures is the kilogram, the weight of a certain piece of metal kept in a government office in Paris.

A weighing machine then is not, generally speaking, used to determine the force of gravity upon a body, but to determine what is its heaviness as compared with that of certain pieces of metal which are assumed to be standards.

For the comparison of standard weights with each other,

for governmental and scientific purposes, instruments of the utmost possible refinement are required, so as to make the error of ascertainment of their relative weights as small as possible. For commercial purposes such refinement is not attempted, but it is desirable that the error of weighing be so small that it can create no dissatisfaction in the mind of either buyer or seller, or user of the weights for any purpose. Thus in weighing tea or sugar in a grocery store, an error in weighing would be allowable if it was not large enough to cheat either buyer or seller by the amount of one cent in the calculation of the total price of the article weighed. In a drug store, the prescription balance should be so accurate that the largest possible error that could be made in weighing the strongest poison known to the Pharmacopœia could not produce the effect of an overdose and endanger the life of a patient. In chemical analysis, the largest allowable error must be less than the limit of accuracy of the process used by the chemist, or say $\frac{1}{1000}$ of one per cent. So the refinement or accuracy necessary in a weighing machine depends chiefly upon the purpose for which the weighing is done.

The mechanical principles upon which weighing machines are based are quite various, but in general they all have one idea in common, that of opposing to the force of gravitation, which acts upon the body to be weighed, some resistance the amount of which can be determined and expressed in units of the accepted standard. I will now enumerate the several mechanical principles which have been adopted or proposed for weighing machines, and show some examples of their application.

(1) *The elasticity of metals or other substances.*

Probably the simplest illustration of this principle is seen in this cheap weighing machine (*Fig. 1*), consisting of a flexible steel strip *A*, rigidly fastened at one end to a firm pedestal *B*, and carrying at its outer end a pan *C*. A one-pound weight placed in the pan *C* will bring the pointer down to the mark 1 on the graduated standard, two pounds brings it down to the mark 2, and so on. By finely graduating the index plate and by reading carefully the

position of the pointer upon it after it ceases vibrating, a fairly accurate weighing machine for a limited range of purposes is obtained.

Weighing on this crude machine is done by the method known as "weighing by substitution," which consists in substituting, for the substance on the scale pan whose weight is to be determined, standard weights sufficient to produce exactly the same effect upon the scale. The sum of the standard weights so used is the weight of the substance. For example, suppose we put on the scale pan of this machine a piece of metal of unknown weight, it depresses the pointer to a spot on the graduated plate, which spot we mark with great exactness. We now carefully remove the piece of metal and substitute standard weights sufficient to

FIG. 1.

bring the pointer to exactly the same position. If it requires just one pound, one ounce and one-eighth of an ounce to bring the pointer to the marked spot, then that is the weight of the piece of metal. This method of substitution is employed in the very finest weighing for scientific purposes, because it enables us to avoid many of the errors common to the ordinary methods of balancing a weight on one side of a scale beam against a weight on the other.

Another form of machine, using the flexibility of this same metal strip, might be made by fastening or supporting the strip at each end on a separate pedestal, and noting the deflection of the middle point. A railroad bridge might thus be used as a weighing machine to weigh the trains of cars that cross it, the heavier the train, the greater of course being the deflection.

The most common illustration of the use of the principle of the elasticity of metals is seen in the ordinary spring balance, which consists simply of a spirally wound wire, which is held at one end and the weight to be weighed suspended from a hook at the other. These wire coils may be used either in compression, like the coiled spring of a street car, or in extension, like the ordinary spring balance. (Specimen shown by the lecturer.) The common form of balance, with a pointer rotating on a dial, is just the same, with the addition of a small rack and pinion to give the rotary motion.

Instead of the elasticity of metal, we may use that of other substances, such as cork, or India-rubber, or we may use the elasticity of air. Suppose a cabinet organ bellows, or a blacksmiths' bellows, perfectly free from leaks, and well inflated with air. Place a pound weight on the upper plate of the bellows, and it will be depressed slightly. Two pounds will depress it twice as much, or nearly so. Thus an index scale of depressions corresponding to certain weights might be constructed, and the bellows thus become a weighing machine. The common gas-holder of the gas works may be used as a weighing machine. It consists of an inner tank inverted in an outside tank, which is filled with water. Gas is introduced into the inner tank, raising it to a certain height. Weights are now placed on the top of the inner tank, depressing it in proportion to the amount of weight. The amount of the depression, recording the pressure to which the contained gas is subjected, is also a record of the weight applied, and thus the gas-holder becomes a weighing machine.

(2) *The buoyancy of liquids.*

I have one illustration of this principle here, in the common hydrometer, used for determining the density of liquids, *Fig. 2*. It may also be used to determine the weight of small bodies. I have it here balanced in strong brine, and a little pan on top of the stem, in which I place this ten-grain weight. It depresses the stem a considerable distance, so that within a range of from one-tenth grain to twenty or thirty grains this hydrometer makes a fair weighing machine.

On the same principle a boat or any vessel floating in water may be used as a weighing machine. When the vessel is unloaded, the water line is at a certain mark. As the vessel is loaded, it sinks deeper and deeper in the water, and successive marks showing the amount of water the vessel is drawing, also measure the weight of the cargo placed on board.

(3) *The chain balance.*

What is called the chain balance is described in the books, but I am not aware that it is ever used in practice.

FIG. 2.

I have here a rough model of one, which is chiefly interesting to show how a crude weighing machine might be made if no better one was at hand. It consists simply of two upright posts (*c d*, *Fig. 3*), between which is loosely stretched a cord or chain, *a b*. To two points in this chain are attached the scale pans, *e f*. When the scale pans are unloaded, the chain will take a definite position, depending upon the position of the points of attachment of the scale pans. If one pan is loaded, that pan will take a lower position, while the other one will rise, the angles between the ends of the chain and the posts changing accordingly. If

an equal load be placed in the other pan, the chain will return to its original position.

(4) *The pendulum or bent-lever balance.*

This is a balance which has had rather limited application hitherto, the most commonly-known form being a cheap and not very accurate letter balance, for weighing a few ounces.

FIG. 3.

I have here a rough model, *Fig. 4*, illustrating the principle, and also a rather complicated modification, a housekeeper's cheap scale, weighing up to ten pounds. It consists, in its simplest form, as the model shows, of a bent arm *a*, pivoted near one end to an upright support, and carrying at the end of the vertical portion a weight *e*, and at the end of the

FIG. 4.

horizontal portion the scale pan *d*, suspended from a pivot. It is evident that as weights are placed in the scale pan, the weight *e* will be caused to move outwards from the supporting pillar, until the leverage of the arm and weight on one side of the pillar counterbalances the effect of the weight in the scale pan. The balance may be standardized by

marking on the graduated index the position at which the arm comes to rest under the application of successive weights.

(5) *The hydrostatic balance.*

The most elementary form of this balance is shown in the sketch on the blackboard (*Fig 5*). It is based on the principle of the hydrostatic press, or common hydraulic jack. There are two communicating cylinders, one very much larger than the other, each fitted with a piston. Leakage and friction being left out of the account, a weight placed on the piston of the smaller cylinder will balance a weight on the larger cylinder, which is as many times greater as the area of the larger cylinder is greater than

FIG. 5.

that of the smaller. By substituting for the pistons flexible diaphragms, the objections of leakage and piston friction are avoided, and upon this principle are constructed the hydraulic-weighing apparatus used in the Emery testing machine, which will be referred to later.

(6) *The even balance.*

The weighing machines based on the principles already referred to are of quite restricted application. We now come to the simplest, most ancient, and most universally used form of weighing machine, the even balance. I have here the rudely constructed parts of a model, consisting of two pieces of board, three feet long and five inches wide, a supporting pillar and two upright pan supports, and two pans with strings for suspending them, also a number of

different kinds of pivots, with which I will construct several rough models illustrating the even balance in many of the forms it assumes in use. No doubt the earliest form of the even balance was the "see-saw," used by children in their play, a plank placed across a log, and a child on each end balancing one another. As an accurate weighing machine, however, the "see-saw" is not a success, on account of the friction and variableness of position of the middle bearing. The first improvement on the see-saw was probably made by turning the plank up edgewise to give it greater resistance to bending, and boring a hole in the middle of it, and inserting a pin in the hole, supporting the pin also on a bifurcated support. This is shown in the form now presented in the model (*Fig. 6*). By suspending the scale pans

FIG. 6.

from the ends of the board by means of strings, we have a fairly good suspended-pan even-balance weighing machine. You observe that a pound weight in one pan balances a pound weight in the other.

Removing the pans, and considering the action of the beam itself, we may learn some important facts concerning the construction of the even balance. First we notice that the pin has a certain amount of friction in its bearings, which brings the beam to rest after two or three oscillations, and prevents its being set in motion except by the addition of quite an appreciable weight on one side or the other. This is very objectionable in a weighing machine, and teaches us that a perfect balance should have a frictionless pivot. Let us make the pivot of polished steel, of small diameter, so as greatly to reduce the friction, and we dis-

cover another peculiar feature of the even balance. If the pin is placed in the uppermost of the three holes the beam oscillates rapidly. When brought to rest a half-ounce weight placed on the end of the beam moves the beam out of its normal or horizontal position to a slight degree only. Place the pin in the lowest of the three holes, and the beam will not remain in the horizontal position at all, but will fall to one side or the other, showing a tendency to turn upside down, so as to bring the supporting pin in the upper edge of the beam as before. This is the condition of unstable equilibrium, or top-heaviness, while in the first test, in which the supporting pin is in the uppermost hole, the condition is one of stable equilibrium. By such tests we learn that there is a position of the supporting pivot which will give the best results, namely, great sensitiveness, without top-heaviness, and this position is almost at the exact centre of gravity of the beam, but a trifle above it. If the pivot be placed exactly at the centre of gravity, the beam will be in neither stable nor unstable equilibrium, but in what is called indifferent equilibrium, that is the beam will have neither a tendency to return to the horizontal position, nor to fall towards one side or the other, but will remain indifferently in any position it may happen to be placed. The centre of gravity must be just a little below the centre of support, so that as the centre of gravity always tends to take a position in a vertical line below the point of support, the beam will always tend to return to the horizontal. Placing the pin in the middle one of the three holes we notice that the beam oscillates very slowly, but always comes to rest in a horizontal position. A half-ounce weight placed on one end of the beam now deflects the beam to a much greater extent than when the pin was in the upper hole.

The round pin, however, is still objectionable as a pivot on account of its friction, which cannot be got rid of, as well as for other reasons of which time will not permit me to speak. Suffice it to say that the round pivot has been replaced by the knife edge in nearly all balances which have been used during the past 2,000 years at least. The balances used in ancient Rome, according to pictures which have come

down to us, show the use of the knife edge, and I have no doubt the balances shown in the sculptured pictures of the Egyptian pyramids also had knife edges. (The lecturer here showed on the screen a picture of an Egyptian balance, and also the picture showing the ceremony of weighing the Great Mogul of Tartary.)

Instead of suspending the scale pans by means of strings tied to the ends of the beam, as I have them in the model, it is usual to have knife edges firmly set into the ends of the beams, to which the pan supports are attached by means of swinging links. By means of the three strings which are attached to each end of the model beam, each passing through a hole and secured by a knot resting on one of the little steps at the end of the beam, we can show that it is a matter of great importance that the points from which the pans are suspended should have a certain position in reference to a horizontal line passing through the centre of gravity of the beam. With the pivot supporting the beam in its proper position, just a trifle above the centre of gravity of the beam, if we attach the pans to the strings attached to the outer and lower step, and put weights in them, the balance ceases to be sensitive. It oscillates rapidly, and is in too stable equilibrium. If we tie the pan supports to the innermost strings, which are suspended by the knot on the uppermost step, the balance again becomes top-heavy, or in unstable equilibrium. If we attach the pan supports to the middle string, supported by the knot on the middle step, the balance oscillates slowly, no matter whether lightly or heavily loaded, it is quite sensitive to small additions of weight to either pan, and it always returns to the horizontal position. Hence we learn that the end pivots of a suspended-pan balance of this general class, that is with round or knife-edge pivots, should always be on a horizontal line with the middle pivot, that is the edges of the three knife edges should be in the same plane.

I have here a very fine scale beam, with three knife edges, made by Mr. Troemner, of Philadelphia, from which may be learned the accuracy of workmanship necessary to

make a good knife-edge balance. The end pivots must be exactly equidistant from the centre pivot, the edges of the three pivots must lie in the same plane, the edges of each pivot must be exactly at right angles to the beam, both horizontally and vertically, which condition also assumes another essential requirement, that the edges must be parallel with each other. Those of my audience who are expert machinists will recognize the fact that to produce absolute exactness in all these conditions is a practical impossibility, and that to produce sufficient exactness to meet the requirements of a fine weighing machine is one of the most difficult tasks that can be performed by a mechanic. I have a number of knife-edge balances here, kindly loaned for this lecture by Messrs. Richlé Brothers, of this city, which show various modifications of construction.

(7) *The lever or steelyard balance.*

The even balance has one serious objection as a weighing machine for heavy weights, viz: the necessity of placing weights in one pan of the balance equal to the weight of the substance which is being weighed in the other. For weights up to ten or twenty pounds, this is no great inconvenience, but when we wish to weigh hundreds or thousands of pounds it becomes intolerable. This led to the adoption in very early times of the lever balance or, as it is sometimes called, the Roman steelyard. The picture shown on the screen of an ancient Roman balance bears very close resemblance to the steelyards used at the present day. The principle upon which this balance is based is that of the lever, namely, that a heavy weight suspended from the end of the short arm of the lever may be balanced by a smaller weight suspended from the end of the long arm, the weights being inversely proportional to the lengths of the arms, or the product of the heavier weight multiplied by the length of the short arm being equal to the product of the lighter weight multiplied by the length of the long arm. This may be illustrated by the model. If we place the supporting pivot in one of the holes near one end of the beam, so that the distance of the pin from one pan support is three times its distance from the other pan support, and add enough

weight in the pan on the short arm to balance the beam; then place three pounds in the pan on the short arm and it will be exactly balanced by one pound placed in the other pan (*Fig. 7*).

In practice the steelyard balance does not generally have a pan support, or weight holder attached to each end of the

FIG. 7.

balance, but only to the short end, while on the long arm there is a movable weight, which may be placed at any position required to balance the weight suspended from the short arm. This is shown in the ancient Roman picture. Marks and nicks are placed upon the beam to indicate the positions at which the movable weight should be placed to

FIG. 8.

counterbalance certain definite weights suspended from the short arm.

(8) *The compound lever balance.*

This is merely an extension of the principle of the steelyard, by using two or more steelyards linked together as shown in the sketch (*Fig. 8*). If a weight of 100 pounds is suspended from the short arm of a beam, whose arms are to each other as 10 to 1, a force of ten pounds applied at the

end of the long arm will balance it, but this force may be applied by means of a second lever. If this second lever also has a ratio of length of arms of 10 to 1, one pound applied at the end of the long arm of the second lever will balance 100 pounds at the short end of the first lever, the weights of the levers themselves of course being first counterbalanced so as to remain in balance when unloaded.

(9) *Combinations of two or more of the eight principles.*

A great variety of combinations of the various forms of balance already described are in actual use as weighing machines, but time will not permit of their being considered in detail.

The even balances and lever balances heretofore discussed all have the scale pans suspended from the beams, but for

FIG. 9.

many purposes suspended scale pans are inconvenient, and upright pans supported above the beam are desired.

Returning to our model, we rigidly attach two pans to the beam, by pinning each pan support to the ends of the beam by two pins (*Fig. 9*). We now place a pound weight on the centre of each pan, and the scale balances correctly. But if we move one of the weights a fraction of an inch outwards from the centre of the beam, the beam will be depressed on that side, an effect exactly the same as would follow from moving outwards the sliding weight on the lever balance. This method of attaching the pans is therefore incorrect, and the balance as thus constructed is a failure. In order to make a successful upright-pan balance, it is necessary to allow the pan supports freely to move on the end pivots of the beam, instead of rigidly attaching them to it, and at the same time to insure that they retain a vertical position, no

matter what deviation from the horizontal the beam may take in its oscillation. This is accomplished by adding to the structure a parallel motion, which we do on the model by placing a second beam, a duplicate of the first, underneath the latter, and pinning each pan support near its lower end to this lower beam, by one pin, and near its upper end by one pin to the upper beam, so that while both beams oscillate together the pan supports, slightly rotating on the pins, preserve a parallel and vertical position (*Fig. 10*). If we now place a pound weight on each pan, the scale will balance as before, but if we move either weight to any side or corner of the pan, it will still balance, eccentricity of loading having no effect to disturb the equilibrium. This is the principle of all upright-pan pivot balances. In ordinary knife-edge

FIG. 10.

upright-pan scales, such as the one shown on the table, an ordinary grocer's counter scale, the lower beam is hidden in the casting, close to the table, and here it takes the form of a slender rod, pivoted to the central support and to the pan supports by round pin pivots instead of knife edges. The lower beam is sometimes made in two parts, being cut in half in the middle of its length, each half being separately pivoted to the central support. Each half of the lower beam is then called a radius arm.

The same principle is adopted in upright-pan lever or steelyard balances, such as the small Fairbanks' letter scale here shown, only one-half of the lower beam, or one radius arm being needed in this case to keep the pan support vertical, since there is only one pan and one pan support used in this style of balance.

The modern platform scale, which was invented by Thaddeus Fairbanks more than fifty years ago, is a compound lever balance. Imagine four of these small steelyard scales placed at the four corners of a rectangle, their graduated beams all pointing towards the centre, where they are separately connected to four strings tied one to the end of each beam, the four strings tied together into one a short distance below, and carrying a pan for holding weights. A flat plate or board may now be placed on top of the four scales, its four corners resting one on each scale. We now have a crude representation of the main principle of the platform scale. I have here an actual platform scale, of the same general style as Fairbanks', which is made by Riehle Brothers, of this city. (The lecturer here described the construction of this scale.)

I have already stated that the round pivot for balance scales is so objectionable on account of its friction that it is almost never used in weighing machines which have any pretension to accuracy, and that for 2,000 years, or more, knife-edge pivots have been in use as the best kind of pivot. In this age of invention, when the things that were good enough for our ancestors are not good enough for us, many attempts have been made to find a better pivot than the knife edge, which shall be free from the objections of the friction which is a necessary defect of its action, even when new and finely finished, but especially of its want of permanency, the knife edge being extremely susceptible to injury from crushing, bruising, the action of the atmosphere, dust, etc. In quite recent times these attempts have met with success, and I now have the pleasure of exhibiting some forms of pivot balance in which knife edges are dispensed with.

The flexure pivot balance.—The first of these is the flexure pivot balance, which I will show by attaching a flexure pivot to the wooden model which has already done us such good service (*Fig. 11*). I have here two short pieces of flat spring steel, which are clamped tightly in two small vises or clamping pieces, so that about half an inch of each strip is left exposed between the faces of the

clamps. Attaching one clamp to our model beam, and one to the pedestal, you notice that the beam oscillates freely, by bending slightly the thin steel strips. I have copies of two English patents, one by J. M. Taurines, of Paris, dated December 18, 1861, and one by Frederick D. Artingstall, of Manchester, dated September 8, 1862, which show and describe balances made on flexure pivots very similar to the one here shown, but I am not aware whether Mr. Taurines or Mr. Artingstall ever made weighing machines for the market on this principle, although it would appear that practical weighing machines might be made under their patents, provided that sufficient mechanical skill were used in the details of construction. In 1872 and also in

FIG. 11.

1883 and 1884, Mr. A. H. Emery obtained patents on weighing machines, in which are combined the principles of the flexure-pivoted beam and the hydrostatic balance with flexible metallic diaphragms, which has already been referred to. In 1879 Mr. Emery completed, at a cost of over \$100,000, his famous testing machine on this combination of principles, which has since that date been in the service of the United States Government, at the Arsenal at Watertown, Mass. It is used for testing the strength of metals and other materials of construction, and has a capacity of 800,000 pounds for strains of either tension or compression, and can test specimens of any length up to thirty feet. A very complete description of this remarkable testing machine was published in a pamphlet entitled, "A New
WHOLE NO. VOL. CXXVI.—(THIRD SERIES, Vol. xcvi.)

System of Weighing Machinery," issued in 1884, by the Yale & Towne Manufacturing Company, of Stamford, Conn. From this pamphlet I have taken a few cuts which I will show on the screen.

The Yale & Towne Manufacturing Company, made a number of testing machines under the Emery patents, of capacities ranging from fifty to 150 tons, one of which is also shown on the screen. They also made a few platform scales of large capacity on the same principle, but I believe these have not yet been brought into regular use, nor have any small weighing scales for ordinary uses yet been put on the market on the flexure principle. Messrs. Wm. Sellers & Co., of this city, have recently secured the right to build testing machines under Mr. Emery's patents, but I am not aware how far they have progressed in their manufacture. *Figs. 12 to 16*, illustrate various portions of this machine.

The torsion balance.—We will now return to our model beam again, and attach another kind of a pivot, which consists of a flat band or wire tightly stretched around an iron frame or truss, so that two straight portions of the endless band are free to be twisted (*Fig. 17*). The beam is firmly clamped to one of these straight portions, at right angles to its length, and supported by it (*Fig. 18*). You observe that as the beam is oscillated, the wire or band twists, hence the name, the Torsion Balance. Professors Gauss and Weber, of the University of Gottingen experimented upon balances with a twisting wire for a pivot a good many years ago, but their experiments and those of some others were so unsuccessful, that the Encyclopædia Britannica states that if used for large weights the beam would have to be supported by knife edges and bearings, and that, therefore, the torsion balance has no *raison d'être*. One of the causes of the failure of the earlier experiments was the want of a correct way of stretching the wire. This was remedied in 1882 by the late Prof. Frederick Roeder, of Cincinnati. A greater trouble, however, and one which seemed to be fatal to the hopes of the torsion balance, was the fact that as soon as the torsion band or wire was given sufficient size to make it strong

enough to carry a desired weight, the elasticity of the wire became so great as to destroy the sensitiveness of the balance. Dr. Alfred Springer, Prof. Roeder's associate, discovered that the objectionable effect of the elasticity of the wire could be entirely overcome by making the beam "top-

FIG. 17.

heavy," that is by raising its centre of gravity above the point of support or axis of oscillation. He thus used the force of gravity to overcome the resistance of the wire to being twisted. On this model beam we can raise or lower the centre of gravity by raising or lowering the poise

FIG. 18.

weight which is carried on a screw bolt above the centre of the beam. On removing the poise weight we notice that the beam oscillates quite rapidly, say twice every second, while the weight of a penny placed on one end of the beam has little appreciable effect in depressing that end. If we

now raise the centre of gravity of the beam, by placing the poise ball on the screw, the oscillations decrease to ten or less per minute, and a penny placed on one end deflects the beam through a large angle. By raising the poise a little, we decrease the rate of oscillation still more, and the beam becomes exceedingly sensitive, while still remaining in stable equilibrium, that is, the beam always comes to rest in the horizontal position, but if we raise the poise too far the gravity of the poise overbalances the resistance of the elastic wire, the balance is top-heavy, or in unstable equilibrium, and the beam tips to one side or the other, but never remains horizontal.

This system has been thoroughly tested since 1882, and balances have been made on the principle so delicate that

FIG. 19.

they would indicate a weight as small as $\frac{1}{25,000}$ th part of a grain, and others so strong that they have carried a weight of 500 pounds on each pan, and yet when loaded with that weight would indicate an addition to either pan of two grains. Platform scales of 1,000 pounds capacity, and automatic grain scales, which are filled, weighed and dumped automatically, at the rate of 200 bushels per hour, have also been made on this principle. The torsion balance in its improved form is now being placed on the market by The Springer Torsion Balance Company, which has recently built a new factory especially for its manufacture in Jersey City. I have several samples of various styles on the table, which may be examined at the close of the lecture. The druggist's counter scale (*Fig. 19*), has a working capacity of twenty pounds on each pan and will indicate an addition of

half a grain. This scale has now been on the market nearly two years. The small letter scale, eight ounces capacity (*Fig. 20*), is only a few weeks old, and the prescription scale (*Fig. 21*), which is the most recent improvement on one which was introduced two years ago, may be called an advance copy of one which will be put on the market next month, this particular scale having been finished only to-day in order to be shown here to-night.

FIG. 20.

FIG. 21.

It contains a new style of sliding weight or rider beam for small weights (*Fig. 22*), a beam engraved on its upper edge for weights from one-eighth grain up to eight grains, and on its lower edge from one-half centigram to five decigrams, so that this balance shows an instantaneous comparison between the English and French weights within this range. The balance has a working capacity of eight ounces, and is sensitive to one-sixty-fourth grain. The torsion balance

FIG. 22.

principle is capable of a great variety of applications in weighing machines, and the Springer Torsion Balance Company expect to bring out a number of other forms during the present year.

(The lecturer closed by exhibiting and explaining samples of torsion and other balances, and answering questions concerning them.)