

A projected Image Reconstruction Algorithm For Electrical Impedance Tomography Using Time Difference Data

Ts. Munkh-Erdene*, E.J. Lee*, J.K. Seo* and S.W. Kim**

**Department of Computational Science and Engineering, Yonsei University, Seoul*

***Division of Liberal art, Hanbat National University, Daejeon, Korea*

ts.munkherdene, eunjunglee, seoj@yonsei.ac.kr, and sungwhankim@hanmail.net

Abstract-This work develops the projection image of conductivity from 3D to 2D. The new proposed method has four electrodes, which we call driving electrodes. There are used to inject current into the object. The other electrodes are the voltage sensing electrodes placed between the driving electrodes, which we called the voltage-sensing electrodes. Based on the derived voltage-current relation, we produce images of conductivity changes within a local region underneath the voltage-sensing probe. We describe the new image reconstruction algorithm and its numerical simulation results.

Keywords-Electrical, Impedance, Tomography, Projection, Time difference

I. Introduction

In conventional electrical impedance tomography (EIT), the electrodes attached around an imaging object. For instance: image of the human chest or image of abdomen and so on. The electrodes on boundary for injection of currents are designed also to measure induced voltage distributions (Adler *et.al* 2009), (Cheney *et.al* 1999), (Holder 2004), and (Brown *et.al* 1985). Sometimes, practically it is not continent and impossible to attach electrodes around a boundary of chosen region. Therefore it is important to focus only in interested region. About it related previous studies to use a planar array of electrodes (Boverman *et.al* 2009) and (Sadleir *et.al* 2009). They used some or all of the electrodes for current injections and voltage measurements and adopted conventional EIT image reconstruction methods. Rabbani *et.al* proposed focused impedance measurement (FIM) using two pairs of current injection electrodes and one pair of voltage-sensing electrodes (Rabbani *et.al* 1999). By adding two voltage measurements subject to two orthogonal current injections, they could enhance the sensitivity of the tetra-polar surface impedance measurement to the admittivity of the internal local region underneath the voltage-sensing electrodes. In this paper, we propose a new EIT image reconstruction algorithm for the electrode configuration of the PHI shown in figure 1.




Figure 1. Electrode configuration: human torso model and meshing

II. Mathematical Model

The time difference Dirichlet data is used to investigate admittivity change of human's organs such as insult to subsurface cortical blood vessels of stroke, lung monitoring and the human abdomen. In this work observation is the admittivity distribution in the human abdomen. Let the human body satisfies a three-dimensional domain Ω with its boundary $\partial\Omega$. The complex admittivity is depending of time t , and position $\mathbf{r} = (x, y, z)$ in Ω denoted by

$$\gamma(t, \mathbf{r}) = \sigma(t, \mathbf{r}) + i\omega\epsilon(t, \mathbf{r}),$$

where $\sigma(t, \mathbf{r})$ is conductivity, $\epsilon(t, \mathbf{r})$ is permittivity, and ω is angular frequency. We attach two pairs of driving electrodes \mathcal{E}_1^\pm and \mathcal{E}_2^\pm to the surface to inject current and a sensing probe occupying Γ to measure the resulting voltages. When we apply a sinusoidal voltage of $V_0 \sin \omega t$ using a chosen pair of electrodes \mathcal{E}_j^\pm , the resulting complex potential $w^j(t, \mathbf{r})$ satisfies the following mixed boundary value problem: $j = 1, 2$

$$\begin{cases} \nabla \cdot (\gamma(t, \mathbf{r})(\mathbf{r})\nabla w^j(t, \mathbf{r})) = 0 & \text{in } \Omega \\ w^j(t, \mathbf{r})|_{\mathcal{E}_j^-} = 0, \quad w^j(t, \mathbf{r})|_{\mathcal{E}_j^+} = V_0 \\ (\gamma(t, \mathbf{r})(\mathbf{r}))\frac{\partial w^j(t, \mathbf{r})}{\partial \mathbf{n}} = 0 & \text{on } \partial\Omega \setminus (\mathcal{E}_j^- \cup \mathcal{E}_j^+) \end{cases} \quad (1)$$

where \mathbf{n} is the unit outward normal vector to the boundary $\partial\Omega$. Through the probe Γ we measure the time varying boundary voltage

$$f^j(t, \cdot) := w^j(t, \cdot)|_\Gamma \quad \text{for } j = 1, 2. \quad (2)$$

The goal is to reconstruct time change of the complex admittivity $\frac{\partial}{\partial t}\gamma(t, \mathbf{r})$ underneath the probe Γ using time difference voltage data $\frac{\partial}{\partial t}f^j$ on Γ . We use the algorithm introduced in (Lee *et.al* (2011)). Figure 1 shows proposing electrode configuration for planar EIT.




Figure 2. a) Simple tank model with logan phantom anomalies b) Current and voltage distributions c) Meshing of moving object d) The trajectory of moving ball




Figure 3. The reconstructed images at different fixed times t_1, t_2, \dots, t_{10}

III. Numerical Computation

We observe simple tank with sizes $10\text{cm} \times 10\text{cm} \times 4\text{cm}$ and background conductivity set to 1s/m . In Figure 2, we placed 8 objects with conductivity 3s/m , which are static. One ball is moving in sinusoidal trajectory with conductivity 4s/m . The reconstruction images shown Figure 3 at different fixed times.

IV. Conclusion

While it has numerous biomedical applications as a non-invasive, portable and low-cost medial imaging modality with high temporal resolution, the spatial resolution is low. To improve the resolution, a new direct EIT image reconstruction algorithm for the electrode configuration of the PHI is proposed and analyzed in this paper. Through various numerical simulations, we showed reconstructed images of the internal admittivity distribution of the body using boundary voltage measurements.

Reference

- Ammari H, Kwon O, Seo J K and Woo E J, T-Scan electrical impedance imaging system for anomaly detection *SIAM J. Appl. Math.* **65**(1) 252–266(electronic), 2004
- Adler A, Arnold J, Bayford R, Borsic A, Brown B, Dixon P, Faes T, Frerichs I, Gagnon H, Garber Y, Grychtol B, Hahn G, Lionheart W, Malik A, Patterson R, Stocks J, Tizzard A, Weiler N, and Wolf G, GREIT: A unified approach to 2D linear EIT reconstruction of lung images, *Physiol. Meas.*, vol. 30, pp. S35S55, 2009.
- Boverman G, Kao T, Isaacson D, and Saulnier G J, An implementation of calderons method for 3-d limited-view eit. *IEEE Trans Medical Imaging*, 28(7):10731082, 2009.
- Brown B H, Barber D C, and Seagar A D, Applied potential tomography: possible clinical applications. *Clin. Phys. Physiol. Meas.*, 6(2):109121, 1985.
- Cheney M, Isaacson D, and Newell J C, Electrical impedance tomography. *SIAM Rev.*, 41(1):85 101 (electronic), 1999.
- Kim S, Lee J, Seo J K, Woo E J and Zribi H, Multifrequency trans-adimttance scanner: mathematical framework and feasibility *SIAM J. Appl. Math.* **69** 22–36, 2008
- Lee E, Seo J K, Woo E J and Zhang T, Mathematical framework for a new microscopic electrical impedance tomography (micro-EIT) system *Inverse Problems* **27**, 2011
- Oh T I, Koo W, Lee K H, Kim S M, Lee J, Kim S W, Seo J K and Woo E J, Validation of a multi-frequency electrical impedance tomography (mFEIT) system KHU Mark1: impedance spectroscopy and time-difference imaging *Physiol. Meas.* **29** 295–307, 2008
- Rabbani K S, Sarker M, Akond, H R and Akert, Focused impedance measurement(FIM): a new technique with improved zone localization, *Ann. New York Aca. of Sci.* **873**, 1999
- Schlappa J, Annese E and Griffiths H, Systematic errors in multi-frequency EIT *Physiol. Meas.* **21** 111–118, 2000
- Yerworth R J, Bayford R H, Brown B, Milnes P, Conway M and Holder D S, Electrical impedance tomography spectroscopy (EITS) for human head imaging *Physiol. Meas.* **24** 477–489, 2003