

A Gold Thin Film Based Compound Electrode Array Suitable for Electrical Impedance Tomography (EIT)

Tushar Kanti Bera* and J. Nagaraju

Department of Instrumentation and Applied Physics,
Indian Institute of Science Bangalore, Bangalore - 560012, India
Phone: +919448853308, Email: tkbera77@gmail.com

*Corresponding Author

Abstract:

Electrical Impedance Tomography (EIT) [1-2] has been researched extensively in medical diagnosis [3] as well as in other fields of science and technology. In medical EIT, electrode contact impedance [1], produced by the skin-electrode interface [1], plays a significant role in the boundary data accuracy which has a great impact on the reconstructed image quality. In EIT, the current is injected and surface potentials are measured on all the electrodes except the current injecting electrodes for contact impedance problem [4]. But for obtaining the greatest sensitivity to the resistivity changes in the domain, measurements on current electrodes are also done [5] though the contact impedance problem produces some error in the EIT systems. A compound electrode [6] array allows us to conduct the current injection and voltage measurement independently by providing two separate sets of electrode geometries (one for current injection and other for voltage measurement). In this paper, a gold thin film based compound electrode array is designed and developed for electrical impedance tomography. As the gold electrodes are found suitable for improved image reconstruction in EIT [7], a flexible compound electrode array (Fig.-1) is developed with gold thin film to eliminate the contact impedance problem in EIT. The thin film based flexible compound electrode is developed by depositing a gold thin film (2 μm) on a flexible FR4 [8] sheet (acting as the substrate) using electro-deposition process [9]. The compound electrode array is found suitable for electrical impedance tomography measurement without contact impedance problem. The compound electrode array (Fig.-2) is put inside a shallow polypropylene tank and the NaCl phantoms [10-12] are developed with nylon cylinders as the inhomogeneities. 1 mA 50 kHz constant sinusoidal current is injected to the phantom boundary using a constant current injector [13]. Boundary potentials are processed with the signal conditioners blocks and then acquired by a USB based high speed data acquisition system controlled by LabVIEW based user interface system. Boundary data are collected with compound electrode array and the results are compared with the data collected for a non compound electrode system. Data are also collected for the phantoms with different inhomogeneity positions and the potential profiles are studied to analyze the contact impedance effects on the boundary data. Resistivity images are reconstructed in Electrical Impedance Tomography and Diffuse Optical Tomography Reconstruction Software (EIDORS). The images obtained for different phantom configurations are studied with compound electrode array and the results are compared with the noncompound electrode array of identical dimensions. It is observed that the design and development of thin film compound electrode array are simple, flexible, biocompatible and low expensive. It is observed that the electronic connections between the wires and electrode connection pads are very easy to make with low contact impedance [7] by soldering of the wires with the gold connecting pads in the electrode array. The

electronic connections of the compound electrode array also reduced the contact impedance problem compared to the other systems especially using stainless steel electrodes [7]. Results also show that the compound electrode array is suitable for boundary data collection and impedance reconstruction avoiding contact impedance problem. As the electrode array is developed on a flexible substrate it would be suitably used for irregular domain geometry in medical imaging. Hence it is concluded that the proposed gold thin film based flexible compound electrode array can be suitably used for improved impedance reconstruction in medical EIT.

Keywords: Electrical Impedance Tomography (EIT), compound electrodes array, flexible electrode array, gold thin film, electrodeposition, resistivity imaging, EIDORS.




Figure 1: Schematic Design of the gold thinfilm based flexible electrode array (a) Noncompound electrode array, (b) compound electrode array.




Figure 2: Figure 1: The current electrode (outer larger rectangle) and the voltage electrode (inner smaller square) in the gold thinfilm based flexible electrode array, (b) the fabricated gold thinfilm based flexible electrode array containing eighteen electrodes (two electrodes are made extra for testing and evaluation purpose).

References:

- [1]. Webster JG, Electrical impedance tomography. Adam Hilger Series of Biomedical Engineering, Adam Hilger, New York, USA 1990.
- [2]. Holder DS, Electrical impedance tomography: methods, history and applications (Series in Medical Physics and Biomedical Engineering) Institute of Physics Publishing Ltd., 2005.
- [3]. Holder DS, Clinical and Physiological Applications of Electrical Impedance Tomography, Taylor & Francis; 1 edition (July 1, 1993).
- [4]. Bera TK and Nagaraju J, Studying The 2D Resistivity Reconstruction of Stainless Steel Electrode Phantoms Using Different Current Patterns of Electrical Impedance Tomography (EIT), Biomedical Engineering, Narosa Publishing House, In: Proceeding of the International Conference on Biomedical Engineering 2011 (ICBME-2011), Manipal, 2011, pp 163-169, ISBN 978-81-8487-195-1.
- [5]. Cheng KS, Simske SJ, Isaacson D, Newell JC, Gisser DG (1990): Errors due to measuring voltage on current-carrying electrodes in electric current computed tomography. IEEE Trans. Biomed. Eng. 37:(6) 60-5.
- [6]. P Hua, E J Woo, J G Webster, W J Tompkins, Using compound electrodes in electrical impedance tomography, IEEE Transactions on Biomedical Engineering (1993), Volume: 40, Issue: 1, Pages: 29-34
- [7]. Bera TK and Nagaraju J, Gold Electrode Sensors for Electrical Impedance Tomography (EIT) Studies, In: Proceedings of the IEEE Sensors Application Symposium 2011 (IEEE SAS 2011), 22nd-24th Feb'2011, USA, pp 24-28.
- [8]. Ataman C, Urey H, Compact Fourier transform spectrometers using FR4 platform, Sensors and Actuators A 151 (2009) 9–16
- [9]. Paunovic M and Schlesinger M, Fundamentals of Electrochemical Deposition, The Electrochemical Society Series, 2nd Ed., John Wiley & Sons, Inc.
- [10]. Holder DS, Hanquan Y and Rao A, Some practical biological phantoms for calibrating multifrequency electrical impedance tomography, Physiol. Meas. 17 (1996) A167–A177.
- [11]. Bera TK and Nagaraju J, Resistivity Imaging of A Reconfigurable Phantom With Circular Inhomogeneities in 2D-Electrical Impedance Tomography, Measurement, Volume 44, Issue 3, March 2011, Pages 518-526.
- [12]. Bera TK and Nagaraju J, A Chicken Tissue Phantom for Studying An Electrical Impedance Tomography (EIT) System Suitable for Clinical Imaging, Sensing and Imaging: An International Journal, Vol. 12, No. 3-4, 95-116.
- [13]. Bera TK and Nagaraju J, A Multifrequency Constant Current Source for Medical Electrical Impedance Tomography, In: Proceedings of the IEEE International Conference on Systems in Medicine and Biology 2010 (IEEE ICSMB 2010), 16th-18th Dec'2010, Kharagpur, India, pp 278-283. DOI: 10.1109/ICSMB.2010.5735387