New Instruments and Suggestions

THE USE OF RADIUM IN CANCER OF THE PROSTATE AND BLADDER

A. PRESENTATION OF NEW INSTRUMENTS AND NEW METHODS OF USE*

HUGH HAMPTON YOUNG, M.D., BALTIMORE

In 1913, Pasteau and DeGrais reported a few cases of cancer of the prostate and bladder in which truly striking results had been obtained by the use of radium introduced into the urethra through an ordinary catheter (Fig. 1). This led me,

Fig. 1.—Catheter method of Pasteau and DeGrais for applying radium in cancer of bladder and prostate.

Fig. 2.—Author's cystoscopic rongeur for obtaining sections of vesical tumors for microscopic study.

in 1914, to procure some radium and to try it in certain selected cases. It soon became evident that for accurate work much more exact methods and special instruments would be necessary, and in the machine shop of the Brady Institute I set about to design and construct instruments to meet the varying needs in vesical and prostatic cancer.

The accompanying illustrations show the different models so well that little description is necessary. The small, straight cystoscope which is used in my cystoscopic rongeur (Fig. 2)

Fig. 3.—Author's Cystoscopic Radium Instrument 1 carrying a tube of radium in fenestrated platinum capsule in beak; closed with obturator.

Fig. 4.—Instrument 1 with straight cystoscope introduced.

was used as a basis around which to construct the various radium instruments.

Radium Instrument 1 (Figs. 3 and 4) carries the radium in its beak. The slender glass tube containing the radium is enclosed in a platinum capsule which slips easily into the open tubular end of the beak of the instrument and is held there

by friction. The instrument is passed through the urethra with an obturator (Fig. 3) which is then withdrawn, the bladder washed out and the cystoscope introduced (Fig. 4), making it possible to locate the region to be treated (Fig. 5 a)

Fig. 5. (a) Observation of tumor through cystoscope. (b) Application of radium capsule against tumor.

and to apply the capsule containing radium directly on it (Fig. 5 b). When in the proper position the instrument is held fixed by the mechanical hand attached to the cystoscopic table shown in operation in Figure 6. This instrument and

Fig. 6.—"Mechanical hand" or clamp holding radium instrument in position desired.

accurate methods of use were demonstrated to several medical societies in 1915.

Radium Instrument 1 has proved satisfactory for applications to the neck of the bladder, trigon and certain portions

Fig. 7.—Author's Cystoscopic Radium Instrument 2 closed (radium in beak) with obturator.

Fig. 8.—Instrument 2 opened ready for use in bladder.

of the lateral and posterior walls. It is usually easy to see satisfactorily with the cystoscope and to place the radium directly on the lesion to be treated.

For tumors well out on the lateral walls of the bladder or on the upper part of the anterior wall this instrument is not

* From the James Buchanan Brady Urological Institute, Johns Hopkins Hospital.

satisfactory, as the end and not the length of the radium tube is presented to the tumor. On this account another instrument was necessary, and to meet this need I have devised a second instrument.

As shown in Figure 7, Radium Instrument 2 carries the capsule of radium within the beak, which is made up of two hollow halves which, when separated by an external operating mechanism, carry the radium out to a position parallel to the shaft and about 2.5 cm. distant from it, as shown in Figure 8, in which it is in position to be placed against tumors quite remote from the middle of the bladder, as shown in Figure 9. This instrument has proved satisfactory in many such cases.

Fig. 9.—Instrument 2 in use: (a) inspection of tumor; (b) application of radium.

For tumors on the posterior wall of the bladder, still another instrument was desirable in order to place the radium broadside against them, as both instruments (1 and 2) here presented only the end of the radium tube to the posterior wall, and a third instrument was therefore designed, as shown in Figure 10.

As seen here, the beak carrying the radium revolves about its center (Fig. 11), being operated by two wires in the shaft and a screw outside. Figures 12 and 13 show the radium carried crosswise the shaft in position to be applied, under cystoscopic direction, to tumors on the posterior wall.

With these three instruments it is possible to place the radium flat (lengthwise) against a tumor in any part of the

Fig. 10.—Author's Cystoscopic Radium Instrument 3 closed, with obturator.

Fig. 11.—Instrument 3: beak with radium revolved for application to posterior wall.

bladder, and to inspect it through the cystoscope from time to time, during the application. A capsule with a fenestra along one side is usually employed in vesical tumors so that the radium may be placed unscreened directly against the neoplasm, but screened from the rest of the bladder. By this means we have been able to secure all the action of both beta and gamma rays, and at the same time have avoided burning the bladder in surrounding regions, thus making much longer treatments quite possible.

The duration and frequency of the treatments vary according to the case, the tolerance of the individual and the size of the tumor. The instrument is held in the desired position

by the mechanical arm, described before, and the treatment continues an hour or more if possible. Intensive treatments are desirable, and in order to shorten the applications in sensitive cases several hundred milligrams of radium in a high

Fig. 12.—Instrument 3 in use against tumor on posterior wall of bladder.

percentage of purity are required to secure the best results. The treatments are made every two or three days, if possible, and a different site on the surface of the tumor attacked each time. We have seen many large tumors respond rapidly, but sometimes a few fulgurations may be helpful in causing

Fig. 13.—Sectional view, showing use of Instrument 3.

large tumor masses to slough away quickly, so that the base can be reached for direct radium applications (Fig. 14).

Later, with Dr. William Frontz, I shall report at length on the results obtained, some of which have been quite remarkable.

In cancer of the prostate, Instrument 1 (Fig. 3) has proved satisfactory, not only for intravesical but also for intra-

Fig. 14.—Application of radium to base of tumor destroyed by fulguration: (a) observation; (b) application.

urethral and rectal applications of radium. In the urethra and rectum the cystoscope is not necessary. I usually employ a nontubular instrument (Fig. 15); but a finger in the rectum tells definitely when the beak of the instrument carrying the

radium capsule is in the desired place, and the mechanical arm holds it there (Fig. 16).

In carrying out these treatments I make use of charts (Figs. 17, 18, 19, 20). As shown here, the size and extent of the cancers of the prostate, seminal vesicles and bladder are indicated in outline and the location of the treatments are given with the date and the number of milligram-hours affixed, thus showing everything graphically on one sheet. By this means the position of successive radium applications is varied from day to day so as to avoid overlapping and consequent burn-

Fig. 15.—Rectal radium carrier, applied under direction of index finger in rectum, and held in place by mechanical hand or clamp.

ing, particularly when normal urethral or rectal mucosa intervenes between radium and tumor, as in cases of prostatic cancer.

By thus applying the radium at a different place each time it has been possible to accomplish as many as twenty-five transrectal treatments (each of one hour's duration) in cases of cancer of the prostate and seminal vesicles without producing any rectal burn or even marked irritation. Heretofore transrectal radium treatments of cancer of the prostate have

Fig. 16.—Mechanical hand in use in treating cancer of seminal vesicles and prostate through rectum.

been considered inapplicable on account of the vulnerability of the rectal mucosa. The methods here described make it possible to eliminate these complications, and greatly increase the ability for intensive treatment. With the instrument here described the radium (screened by a platinum capsule) can be applied, with the direction of a gloved finger in the rectum, accurately to almost any desired point of the carcinomatous prostate or seminal vesicles, and held there for an hour or more. Not infrequently a region above the seminal vesicle along the lateral wall of the pelvis and

adjacent to the sacro-iliac synchondrosis is reached in order to pass well above the upper limits of the disease, as shown in Figure 20, and treat the lymphatics or glands of this region.

The urethral and vesical mucosae have proved very tolerant to transurethral and transvesical treatments of cancer of the prostate and seminal vesicles. Not infrequently from 200 to 300 milligram-hours at one seance have been borne by the

Fig. 17.—A case of cancer of the prostate and left seminal vesicle (right vesicle slightly involved). The radium treatments, each representing 100 milligram-hours, are shown, and the date of each is recorded.

urethra or bladder without burn or evident irritation. The deep urethra affords three nonoverlapping positions for transurethral radiation, and each of these may be again used without fear in two weeks or more. Four or five transvesical positions are available for treatment of the seminal vesicle,

Fig. 18.—Chart of transvesical and transurethral treatments in case shown in Figure 17. Each line represents 100 milligram-hours.

Fig. 19.—Chart of a case of tumor of bladder (involving left ureter) showing radium treatments given with Instruments 1 and 2.

and at the prostatic orifice the beak of the instrument may be turned in various directions against as many different enlarged lobes as may be present.

I began this radium work timidly and cautiously, as I was fearful of injury caused by burns from short urethral or

Fig. 20.—Chart of transrectal treatments in case of tumor of bladder (Figure 19). The higher applications are intended for possible glands along the pelvic wall.

Fig. 21.—Chart of treatments given tumor of anterior wall of bladder near prostate.

rectal treatments. The methods described above were evolved slowly, step by step, after we had demonstrated that by exact methods it was possible to place the radium where we wished and to change its location sufficiently often to avoid injury of normal intervening adjacent or remote structures. By these means far greater localized treatment, made directly

on the neoplasm or through an insignificant layer of intervening mucous membrane, has become possible. I have also used radium in prostatic hypertrophy cases in which the patient was too desperately sick to be operated on.

This article is not intended to speak of ultimate results, but we may safely say here that truly astonishing results have been obtained in some cases: disappearance of obstruction, shrinkage and great softening of certain cancers of the prostate, and extensive retrogressive changes in inoperable cancers of the bladder. Fairly large amounts of radium, in high concentration, are desirable in order to make the treatments, amounting to 200 or more milligram-hours, of as short duration as possible. This is particularly true of cancers of the bladder, in which the patient is often intolerant of long treatments.

These methods have opened up a new and fertile field of therapeutics in a class of urologic cases which have heretofore been almost beyond relief. While we do not as yet know all the possibilities of cure, we do know that much relief can be afforded.

A CONVENIENT OIL DROPPER

Laura Leonard Gilman, B.A., Madison, Wis.

Assistant Bacteriologist, State Laboratory of Hygiene

A dropper for immersion oil, which will be found to be an asset in a busy laboratory, can be made easily and inexpensively from an ordinary rubber ear and ulcer syringe and a bit of glass tubing. The glass tubing, which should be just large enough to pass through the orifice of the syringe, should be bent as shown in the accompanying drawing, and should be long enough to reach almost to the bottom of the bulb. The bent delivery end is to be drawn out so that its lumen is large enough to admit only a fine platinum needle.

This oil dropper puts accurate delivery of the oil and the size of the drop under the control of the user. It does not spill its contents when tipped over, nor does its exterior become covered with sticky oil drippings. It requires only one hand to operate. It delivers uniformly whether full or nearly empty, and if it is originally filled with good fresh oil, the lumen of the glass tube will not become caked with dried oil, even after months of service.

For those who use an oil immersion lens frequently, this device will save a great deal of inconvenience and delay.

A NEW INCISION FOR THE RADICAL RESECTION OF TUBERCULOUS GLANDS OF THE NECK

W. E. McILVAINE, M.D., HAMMONTON, N. J.

I have termed this the "T incision." It consists of a horizontal portion and a vertical portion, as its name indicates.

The horizontal portion practically coincides with Dowd's upper incision. It is begun about 1 inch below the lobe of the ear and runs forward toward the median line of the neck parallel with the ramus of the inferior maxillary bone and in one of the furrows of the skin. This is represented in the illustration by the line A-B.

From the middle of this incision begins the vertical portion, which runs downward and slightly backward toward the junction of the middle and outer thirds of the clavicle, a-b.

This incision is particularly advantageous in extensive tuberculous or malignant disease of the cervical glands.

Through the horizontal portion may be removed the superficial collar of glands or any deeper glands in the submaxillary triangle. It may be extended backward to the hair line to facilitate the removal of the retro-auricular glands, or prolonged forward to remove the submental glands.

The vertical portion of the incision permits the excision of the substernomastoid chain of glands. It crosses the

sternocleidomastoid muscle at about the middle, and both edges of the muscle can be easily freed and lifted out of the way. Through the lower end of the incision it is an easy matter to remove the supraclavicular glands.

If such an extensive excision as this were contemplated with the Dowd incisions, at least two separate incisions would be necessary, which would not offer so much freedom to the surgeon as would the incision just described. The anatomic incision does not give free access to the submaxillary triangle or the posterior triangle.

I close this wound with plain catgut subcuticular stitches. It is drained for from twenty-four to thirty-six hours. The wound heals rapidly and leaves only a linear cicatrix. The upper horizontal segment is not conspicuous because it lies in a furrow of the skin. The lower vertical portion is covered by the collar.

408 Bellevue Avenue.

Incision for radical resection of tuberculous glands of the neck: A-B, horizontal segment of T incision, a-b, vertical segment; the broken line represents the anatomic incision.

THE ANTIFORMIN SPUTUM CUP

M. J. FINE, M.D., NEWARK, N. J.

Chief Clinical Assistant, Newark Tuberculosis Bureau; Visiting Physician, Newark Tuberculosis Sanatorium at Verona, N. J.

All bacteriologists are familiar with the antiformin method of examination for tubercle bacilli. Antiformin is an alkaline antiseptic which dissolves the mucin and freezes the tubercle bacilli so that they can be sedimented.

For a more expeditious use of the antiformin process, I have devised a special sputum cup, which should prove of particular benefit to private laboratories and to general practitioners sufficiently interested in tuberculosis to desire to make their own examinations of suspected sputum. This cup and the antiformin solution are given to the patient with proper instructions how to admix the sputum with the antiformin. The essential principles operative are the insolubility of tubercle bacilli in the antiformin, and the filtering of the homogeneous mass through an exceedingly fine filter, on which as a residue the tubercle bacilli are found clumped together abundantly.

The cup consists of two metal receptacles, A and B, which may be screwed together tightly without any leakage of con-

The antiformin sputum cup: Fig. 1, the sputum cup with the receptacles, containers and mixture as sent to the patient; Fig. 2, the action of the antiformin and filtrative; Fig. 3, the containers in the sputum cup; Fig. 4, the receptacles and the screw thread for adjustment.

tents. The outward aspect of the cup is represented by Figure 1. Fitting snugly into the lower receptacle is a metal container, D, constructed on the same principle as that of an unspillable inkwell, made of tin and properly soldered, and in the upper receptacle is a similar container, C; this container is corked, and holds the antiformin solution. A fine