I. A comparison of Rowland's mercury thermometers with a Callendar-Griffiths' platinum thermometer; a comparison of the platinum thermometer with a Tonnelot thermometer standardized at the Bureau International; and a reduction of Rowland's value of the mechanical equivalent of heat to the Paris nitrogen scale

Charles W. Waidner Jr. & Francis Mallory

To cite this article: Charles W. Waidner Jr. & Francis Mallory (1899) I. A comparison of Rowland's mercury thermometers with a Callendar-Griffiths' platinum thermometer; a comparison of the platinum thermometer with a Tonnelot thermometer standardized at the Bureau International; and a reduction of Rowland's value of the mechanical equivalent of heat to the Paris nitrogen scale , Philosophical Magazine Series 5, 48:290, 1-45, DOI: [10.1080/14786449908621303](https://doi.org/10.1080/14786449908621303)

To link to this article: <http://dx.doi.org/10.1080/14786449908621303>

Published online: 08 May 2009.

Submit your article to this journal [↗](#)

Article views: 3

View related articles [↗](#)

THE
LONDON, EDINBURGH, AND DUBLIN
PHILOSOPHICAL MAGAZINE
AND
JOURNAL OF SCIENCE.

[FIFTH SERIES.]

JULY 1899.

- I. *A Comparison of Rowland's Mercury Thermometers with a Callendar-Griffiths' Platinum Thermometer; a Comparison of the Platinum Thermometer with a Tonnelot Thermometer standardized at the Bureau International; and a Reduction of Rowland's Value of the Mechanical Equivalent of Heat to the Paris Nitrogen Scale.* By CHARLES W. WAIDNER, Jr., and FRANCIS MALLORY*.

THE recent determinations of the mechanical equivalent of heat by electrical methods, while in fair agreement among themselves, seem to give results considerably higher than those obtained by experimenters using the more direct mechanical methods; these differences, which are probably greater than the errors of experiment, must be due to a difference in the standards of thermometry employed or to a still undiscovered error in the system of electric units. The object of the present investigation was to obtain a connexion between Rowland's standard of thermometry and that employed in the electrical determinations of the equivalent, in order to render the different determinations more easily comparable.

Joule compared his thermometer with Rowland's Baudin 6166 (*Proc. Amer. Acad.* xvi. p. 38, 1880), and in this way Joule's determinations of the mechanical equivalent have been reduced to the Rowland "air-scale." By a further elaborate comparison by Prof. Schuster (*Phil. Mag.* xxxix. 1895)

* Communicated by Prof. H. A. Rowland. An abstract of this investigation was read at the Toronto meeting of the British Association, and a preliminary account appeared in the Johns Hopkins University Circulars, June 1897, 1898.

of Joule's thermometers with a thermometer standardized at the Bureau International, an indirect connexion was obtained between Rowland's Air-scale and the Nitrogen-scale of the Bureau International. This comparison pointed to differences in the two scales as great as $0^{\circ} 05$ C. Inasmuch as the details of the method employed by Joule in his comparison are not known, Professor Ames suggested that another more direct comparison of Professor Rowland's thermometers was desirable.

Method of Comparison.

As we wished to use the results of this experiment to reduce Rowland's values of the equivalent to the scales used by experimenters employing the electrical methods, and thus render the results comparable, it was deemed best to make the comparisons of these thermometers under conditions as nearly as possible the same as those under which they were employed by him in his experiments on the mechanical equivalent. For this reason, a Callendar-Griffiths' platinum thermometer was selected for the comparison. This instrument offers many advantages for the standardization of calorimetric thermometers. The platinum thermometer was subsequently compared with a Tonnelot thermometer standardized at the Bureau International. A connexion was thus obtained between Rowland's air-scale and the Callendar-Griffiths' air-scale, as well as with the nitrogen-scale of the Bureau International.

The zero readings of the mercury thermometers were taken in the same way that Rowland used them, before each comparison. The thermometers were placed in a metallic vessel well wrapped with thick felt and filled with finely cracked ice of great purity steeped with distilled water. After they had remained in ice about one and one-half hours, the zero readings were taken by means of a micrometer-telescope. The mercury thermometer was then hastily transferred to a well stirred calorimeter with its bulb near that of the platinum thermometer. The temperature of the water in the calorimeter at the start was generally about 2° or 3° C. A series of five or six readings on the mercury thermometer was then taken by one observer with the aid of a micrometer telescope, while the other observer took simultaneous readings of the resistance of the platinum thermometer.

The temperature of the water in the calorimeter was then raised at a rate of about 10° to 15° per hour to the next temperature at which a comparison was to be made; this was accomplished by sending a suitable current through a coil wrapped on the outside of the calorimeter. When the desired temperature was reached, the current was cut off or so regulated that the temperature was very slowly rising, when

a further series of simultaneous observations were taken, and so on to the end of the scale.

The mercury thermometers were used in a vertical position, and all observations were taken while the meniscus was very slowly rising, usually at a rate of about $\cdot 002$ or $\cdot 003^\circ$ per minute, so that the slight uncertainties of a falling meniscus were avoided. At the point where the thermometer emerges from the calorimeter, it is surrounded by a small water-jacket in which a small thermometer and stirrer are placed. The temperature of the mercury column above the water-jacket is assumed to be that of the surrounding air as indicated by a thermometer hung within an inch or two of the stem. Thus the stem correction was applied in two parts by means of the formula $\Delta N = \cdot 000156 n (t - t')$, where ΔN is the correction to the observed stem-reading, t the observed temperature, t' the temperature of that portion of the stem emerging from the calorimeter, and N the number of stem-divisions at the temperature t' . The stem correction was applied to the observed stem-reading, and the corresponding temperature on Rowland's air-scale was then taken from his tables (Proc. Amer. Acad. xv. 1879, pp. 115 & 116); this temperature was then reduced by the rise of the zero since Rowland's experiments.

The corresponding temperature on the Callendar-Griffiths' air-scale was obtained from the observed resistance of the platinum thermometer; the way in which this was done will be explained more fully later.

*Manner in which Baudin Thermometers were used by
Professor Rowland.*

These thermometers were first calibrated by measuring the length of a short column of mercury in different portions of the stem, and in this way the relative volumes of different parts of the tube were found. Temperature on the mercury-in-glass thermometers was then defined as proportional to the apparent volume of mercury in glass, when the thermometer is vertical. It was then assumed that the apparent volume of mercury-in-glass could be expressed as a function of the second degree of the temperature on the air thermometer, of the following form:

$$T = C'V - t_0' - nT(40 - T) \{1 - n(40 + T)\}$$

using the 0° and 40° points as fixed by the air thermometer; where T is the temperature on the air thermometer,

V is the volume of the stem of the mercurial thermometer, as determined by the calibration, and measured from any arbitrary point,

C, t_0', m, n are constants to be determined.

From a series of direct comparisons of the mercurial thermometers with a standard air thermometer, a number of equations of the above form were obtained, from which the most probable values of these constants were deduced by the method of least squares.

From these formulæ tables were constructed giving the temperature on the scale of the standard air thermometer corresponding to each centimetre of the stem. These tables are found on pp. 115 & 116, Proc. Amer. Acad. xv. 1879.

These thermometers were always used with a constant zero determined before each series of observations, by immersing the thermometer in a mixture of finely cracked ice and distilled water for a period of one or two hours. As the position of the zero, owing to the imperfect elasticity of the glass, depends on the temperature to which the thermometer has been subjected recently, the thermometers were always kept at least one week at a temperature of about 20° C. before they were used in a comparison, or in a determination of the mechanical equivalent, in which interval the zero will have almost recovered from the depression due to a recent heating of the thermometer.

In the original comparisons of these thermometers with the air thermometer, no corrections were necessary for internal pressure due to the mercury column and meniscus, as the thermometers were compared and always used in a vertical position. A correction for external pressure was made, due to about 60 cms. of water in the comparison tank.

In the experiments on the mechanical equivalent, the thermometer was placed in the calorimeter with the top of the bulb about 5 cm. below the surface of the water, and the remainder of the stem projected out into the air. The stem correction was then applied in two parts by surrounding the stem, where it projected from the calorimeter, with a small water-jacket, the remainder of the stem being assumed to be at the temperature of the surrounding air.

Platinum Thermometry.

The measurement of temperature by means of the variation of resistance of a wire was first seriously proposed by Siemens, who submitted for trial to a committee of the British Association some platinum-resistance pyrometers (Brit. Asscc. Report 1874). The results of these experiments were unfavourable to the use of the platinum-resistance thermometer as a scientific instrument of precision. It was found that, for this type of pyrometer, the resistance of the platinum wire was not a constant for a fixed temperature, but depended greatly

on the previous history of the wire. The development of this type of thermometer was next taken up by Callendar, who, in an exhaustive series of researches (Phil. Trans. Roy. Soc., vol. clxxviii. A), showed that if the platinum wire is fairly pure to begin with, and if it has been well annealed, the resistance is always the same at the same temperature, even if the wire has been repeatedly exposed to quite high temperatures in the interval. In the light of these researches, it is plain that the failure of the Siemens' resistance pyrometer was simply due to faulty construction, which exposed the wire to the action of injurious vapours &c. from the clay on which it was wound and the iron cylinder in which it was enclosed.

The extensive researches of Dewar and Fleming (Phil. Mag. 1893) and Cailletet and Colardeau (*J. de Phys.* t. viii. 1888) at very low temperatures, and those of Heycock and Neville (Chem. Soc. Trans. 1895) at very high temperatures, taken with those of Callendar and Griffiths (Phil. Trans. Roy. Soc. vol. clxxviii. A, 1887; vol. clxxxii. A, 1891) at ordinary temperatures, establish beyond a doubt the remarkable accuracy and constancy of this instrument over a very wide range of temperature.

Temperature on the platinum scale is defined by the equation

$$pt = \frac{R - R_0}{R_1 - R_0} \times 100,$$

where pt denotes the *platinum temperature*, R the resistance of the platinum coil of the thermometer at that temperature, R_0 its resistance at 0°C. , and R_1 the resistance at 100°C. From the above definition it is seen that *platinum temperature* is independent of the unit of resistance employed.

Callendar and Griffiths, in an elaborate comparison of the platinum and air thermometers (Phil. Trans. Roy. Soc. vol. clxxxii. A, 1891), have shown that the relation between the two scales of temperature can be expressed by an equation of the form

$$t - pt = \delta \left\{ \left(\frac{t}{100} \right)^2 - \frac{t}{100} \right\},$$

where t is the temperature on the air-scale, pt the *platinum temperature* as defined above, and δ a constant depending on the specimen of wire used. This relation being a parabolic one, only three temperatures are necessary for the complete standardization of a platinum thermometer, *i. e.*, for the determination of δ . Two temperatures always used for this purpose are that of melting ice, 0°C. , and that of water boiling freely under a pressure of 760 mm. of mercury whose

density is that of 0°C. , sea-level latitude 45° , 100°C. For the third temperature the boiling-point of some substance, which is accurately known from a direct measurement with an air-thermometer, is used. The one most frequently used is the boiling-point of sulphur, $t=444^{\circ}\cdot53$, on account of the accuracy with which it is known. Callendar and Griffiths have shown that if platinum thermometers be standardized by these three temperatures (melting ice, steam sulphur vapour), the above formula will give temperatures on the air-scale to within $0^{\circ}\cdot01\text{C.}$ over the range 0°C. to 100°C. This conclusion is further confirmed by a direct comparison made by Dr. W. S. Day and the authors between a platinum thermometer and a Tonnelot thermometer standardized at the Bureau International, whose scale of temperature is based on measurements made with a constant-volume gas thermometer. The comparison referred to above, made by Callendar and Griffiths, is based on a constant-pressure air thermometer.

Description of Apparatus.

Mercury Thermometers.—The thermometers used by Rowland in his experiments on the mechanical equivalent were constructed by Baudin in 1876–77, and are numbered 6163, 6165, and 6166. Of these thermometers, Baudin 6163 is perhaps the most important, as it was used in eight of the fourteen determinations. Baudin 6166 was used in four determinations, and has besides an historical interest, as Joule compared his thermometers with it; Baudin 6165 was used in only one determination of the mechanical equivalent. Besides these, Kew 104, constructed by Welsh in 1853, was used in one determination. This thermometer was not available for this comparison, but inasmuch as it was only calibrated to $0^{\circ}\cdot5\text{F.}$, and was given but little weight in the final results, its effect on the values of the mechanical equivalent is entirely negligible.

The stems of the thermometers are graduated in millimetres.

Thermometer.	Range.	1°C. occupies about
Baudin 6163	-6°C. to 40°C.	9·0 mm.
„ 6165	-3°C. to 33°C.	11·7 „
„ 6166	-2°C. to 31°C.	12·9 „

Platinum Thermometer.—The platinum thermometer used in these experiments, and shown in fig. 1, was constructed by

the Cambridge Scientific Instrument Co., under the supervision of Mr. E. H. Griffiths. It consists of a coil of fine platinum wire wound on a mica frame whose edges are serrated to receive the windings. The thick leads running down to this coil are held apart by mica washers, and the whole thermometer is enclosed in a thin glass tube. The thick leads are connected with the terminals marked P P.

Fig. 1.

This thermometer was also provided with compensating leads, so that its indications are independent of the depth of stem-immersion, provided this immersion is sufficient to prevent conduction from the outside along the leads to the coil. The compensating leads C C are simply a loop (not connected with the platinum coil) running down the stem parallel to the main leads P P throughout their length, so that they are

exposed to the same temperature changes. These compensating leads are adjusted so that their resistance is equal to that of the main leads P P. Now, obviously, if these leads C C are placed in the other arm of the Wheatstone-bridge, as shown in fig. 2 below, they will exactly compensate for the change in resistance of the leads to the coil, and consequently the resistance we measure, and hence the temperature, is independent of the depth of the immersion of the thermometer in the liquid or vapour whose temperature is sought. A little consideration will show that it is not necessary that the resistance of the compensating leads be exactly equal to the leads to the coil in order to have compensation. It is only necessary that they be so nearly equal that the temperature changes of their difference produce no appreciable effect on the resistance.

Calorimeter.—The calorimeter, in which these comparisons were made (shown in fig. 1), was made of copper and provided with an efficient stirrer, which was driven by a small hot-air engine. The outside of the calorimeter was closely wrapped with a coil of single silk-covered german-silver wire, so that the temperature of the calorimeter and its contents could be raised at any desired rate by an electric current of suitable strength. The calorimeter was placed inside a large double-wall copper vessel (ext. diam. 36 cm., ht. 35 cm.), which surrounded it on all sides except the top. The space between these walls (2.5 cm. apart) could be filled with water at any desired temperature, and thus the temperature of the calorimeter could be maintained constant for some minutes or caused to vary at will. The entire arrangement was then surrounded by a large box (46 cm. on the side), and all intervening spaces packed loosely with feathers.

Resistance-Box.—The resistance-measurements of the platinum thermometer were made with a Callendar-Griffiths resistance-box (No. 7) especially designed for the measurement of platinum temperature. The construction and calibration of a resistance-box similar to the one used by us has been described by Mr. E. H. Griffiths in 'Nature,' Nov. 14th, 1895. A general idea of the construction of the box may be gathered from the following diagrammatic sketch.

A_1, A_2 are two equal coils which serve as the equal arms of a Wheatstone-bridge. The coils of the box, which were made of platinum-silver, have approximately the following nominal values in box-units (approximately, one box unit = 0.01 ohm):—

H	G	F	E	D	C	B	A	FI
5	10	20	40	80	160	320	640	100

The bridge-wire W W was made of platinum-silver, was 30 cm. long, and had a resistance of about 0.30 ohm. Contact was made with the galvanometer through a similar wire stretched parallel to the bridge-wire. The point of contact

Fig. 2.

was determined by a very carefully constructed contact-key carrying a vernier reading directly to $\frac{1}{50}$ mm., so that $\frac{1}{100}$ mm. was easily estimated (corresponding to about 0.00001 ohm). This ingenious contact-key, which was designed by Mr. Horace Darwin, renders any injury to the bridge-wire almost impossible. The scale and bridge-wire are connected in such a way that the readings are practically unaffected by very considerable changes in temperature. The leads to the coil of the platinum thermometer were connected to the terminals P P and the compensating leads to the terminals C C. The galvanometer is shown at G, and the battery at B.

When the intervals C to C and P to P were short-circuited by strips of copper, and all plugs inserted, the bridge was balanced when contact was made with the bridge-wire at some point very near the centre, O. The reading of the bridge-wire when the bridge was balanced under these conditions is called the zero correction to the bridge-wire. The platinum thermometer was then inserted, and its resistance at any temperature could be measured by suitable combinations of coils and bridge-wire. The corrections to reduce the

nominal values of the coils to mean box-units at some standard temperature, as determined by a previous calibration of the box, were then applied, as well as the corrections to reduce the bridge-wire readings to mean box-units. The resistance of the platinum thermometer was then known in terms of the mean box-unit. Any resistance could be measured with several different combinations of coils and bridge-wire, and thus the accuracy of the measurements was exposed to a severe test.

The top of the box was constructed of white marble, which was found to have a very high insulation resistance. The brass plugs were large and carefully ground, but at no point were they larger than the top of the hole into which they were inserted. This is a very important detail, as it insures the impossibility of wearing a shoulder on the plug, in which case no amount of pressure will give good contact.

This box was also provided with small switches by which resistances could be thrown into the battery-circuit and the galvanometer shunted, which proved a great convenience in finding an approximate balance.

The bridge was always used with a key by which the circuits could be closed and opened in such a way that the effects of thermoelectric currents were eliminated.

In these experiments the resistance-box was supported in a large double-wall copper tank. The inner compartment of this tank, in which the resistance-coils are placed, is filled with a highly insulating neutral oil, which several careful tests showed to be free from acid or alkali. The outer compartment was filled with water which was maintained at a nearly constant temperature by a suitable thermal regulator. The top of the box was enclosed in a glass case, thus preventing sudden changes in the temperature of the top of the box and the bridge-wire and the accumulation of dust.

We take this opportunity to express our deep sense of obligation to the authorities of the University of Chicago, who, through the kindness of Professor W. S. Stratton of the Ryerson Physical Laboratory, placed this valuable apparatus at our disposal for this investigation.

Galvanometer.—The galvanometer was of the four-coil Thomson type, having what is practically a double metallic case so that the disturbing effects of air-currents are reduced to a minimum. The dimensions of these coils were 19 mm. external diameter, 4 mm. internal diameter, 8 mm. deep, the distance between centres of coils when mounted being 25 mm. Each coil was wound in three sections (Nos. 36, 33, and 30 B. & S. wire being used) approximately according to

the theoretical curves of best winding with a total of 530 turns. The coils were covered with gold-leaf, and in this way connected with the surrounding metallic case to avoid electrostatic disturbances. The resistance of each coil was about 13 ohms. The sensibility of this galvanometer with all the coils in parallel ($\text{Res.} = 3.2$ ohms) when tested with a Wiess suspension system (two vertical magnets 27 mm. long, 1.5 mm. apart, made of small magnetized sewing-needles and mounted on a thin strip of mica; total weight about 45 milligrammes) was $C = 6 \times 10^{-10}$ ampere; *i. e.*, the current required to produce 1 mm. deflexion on a scale 1 metre distant when the period of a complete swing was 10 seconds. When tested with a very light suspension system weighing about 5 milligrammes (built of 10 or 12 magnets about 1 mm. long each, made of fine watch-spring, tempered and strongly magnetized, total weight about 6 milligrammes), the sensibility under the same conditions was $C = 2.5 \times 10^{-10}$. All the magnets used in the suspension systems were repeatedly magnetized and boiled for several hours, so that magnets of great permanency were obtained. Throughout the present investigation the Wiess suspension system was used on account of its greater freedom from external disturbances.

Barometer and Standard Metre.—The barometer was of the Regnault standard type with a tube about 1 m. in height and 25 mm. internal diameter. The lower meniscus was defined by a steel screw of known length. The barometric height was measured on the standard metre bar placed beside the mercury column by means of a cathetometer. The barometric heights as measured by the scale on the cathetometer column, merely as a check, were always found to be in fair agreement with the values found by the standard metre. This barometer was compared with a Fortin standard barometer, and the agreement (to $\frac{1}{10}$ mm.) was as close as we could read on the latter instrument.

The standard Bartel and Diedrich's metre was divided into millimetres, with the rulings on a strip of silver, imbedded in a bronze bar of square section about 21 mm. on the side. The corrections to the divisions of this metre were known from previous comparisons with a standard steel metre which had been compared with the Coast Survey and other standards by Prof. Rogers.

Standardization of Resistance-Box.

The resistance-box, mounted as already described, was twice standardized with the greatest care. While we cannot give within the limits of the present paper the details of the

method used or the mass of figures involved, it is sufficient to state that the method was in general similar to the one described by Mr. E. H. Griffiths in 'Nature,' Nov. 14th, 1895. The figures given below are interesting as showing the degree of accuracy attainable. It will be seen that the two calibrations gave almost identical results.

TABLE I.

Coils.	Nominal values in Box Units. [approx. 1 B.U. = .01 ohm.]	Corrections (in mean B.U.) to reduce nominal values to mean box-units at 20° C.		Mean.
		1st Cal.	2nd Cal.	
H	5	-.0009	-.0007	-.001
G	10	+.0042	+.0036	+.004
F	20	+.0312	+.0310	+.031
E	40	+.0803	+.0799	+.080
D	80	+.0284	+.0277	+.028
C	160	-.0415	-.0430	-.042
B	320	-.0383	-.0362	-.037
A	640	-.0637	-.0635	-.063
FI	100			+.037

Calibration of Bridge-wire.—This consists in finding the resistance of each centimetre of the bridge-wire in terms of the mean box-unit at some standard temperature. In resistance-box No. 7 the scale of the bridge-wire, extending in each direction from the centre, is divided into centimetres, and very approximately 1 cm. of the bridge-wire has a resistance equal to 1 mean box-unit. The final results of the calibration are therefore given in the table below as a correction which is added to the bridge-wire reading (in cms.) to reduce this to its equivalent resistance in mean box-units. The process is exactly similar to the calibration of a graduated glass tube, where the length of a short column of mercury is measured in different parts of the tube. A gauge-coil of manganin wire whose resistance was equal to about 2 cms. of the bridge-wire was balanced against different portions of the bridge-wire. Different parts of the bridge-wire (approx. 5 cms. long) were next balanced against the H (5) coil of the resistance-box, and a connexion thus obtained between the bridge-wire and box-unit.

TABLE II.—Correction to reduce the Bridge-wire Reading to the Mean Box Unit at 20° C.

Bridge-wire Reading.	Calibration Correction.	Bridge-wire Reading.	Calibration Correction.
+0.5	-.004	-0.5	+.004
+1.0	-.007	-1.0	+.007
+1.5	-.011	-1.5	+.011
+2.0	-.015	-2.0	+.014
+2.5	-.018	-2.5	+.018
+3.0	-.022	-3.0	+.022
+3.5	-.025	-3.5	+.026
+4.0	-.028	-4.0	+.029
+4.5	-.031	-4.5	+.033
+5.0	-.034	-5.0	+.036
+5.5	-.037	-5.5	+.040
+6.0	-.041	-6.0	+.043
+6.5	-.044	-6.5	+.046
+7.0	-.047	-7.0	+.049
+7.5	-.051	-7.5	+.051
+8.0	-.054	-8.0	+.054
+8.5	-.057	-8.5	+.057
+9.0	-.060	-9.0	+.060
+9.5	-.063	-9.5	+.064

The bridge-wire vernier could be easily read to the nearest $\frac{1}{100}$ mm., and this with considerable confidence. Observations of any constant resistance taken with the same combinations of coils and bridge-wire rarely differed by as much as 0.00002 ohm; when several different combinations were used to measure a constant resistance, as will be seen on referring to the measurement of R_0 (the resistance of the platinum thermometer at 0° C.), the several measurements rarely differed by as much as 0.00005 ohm. This is a severe test of the accuracy of the standardization. Even these slight discrepancies are not entirely due to errors in calibration, but could, we believe, be still further diminished by stirring the water in the outer tank surrounding the oil-bath in which the coils are immersed. Such great accuracy in the standardization of a resistance-box is, of course, easily obtained so long as we are only concerned with the relative values of the coils, which is all that we care about in the measurement of platinum temperature. But if the absolute values of the coils were required, a similar degree of accuracy would involve a vast amount of most painstaking work, mainly on account of the unscientific construction of most standard coils, which renders the accurate determination of their temperature impossible, and of their having undergone considerable changes in resistance with time.

For the platinum thermometer used throughout this investigation a change in temperature of $0^{\circ}\cdot001$ C. corresponds to a change in resistance of about $0\cdot00001$ ohm. Owing to the great sensibility of the galvanometer, changes in resistance far more minute than the vernier would indicate (*i. e.* $0\cdot00001$ ohm) were easily detected. Advantage was taken of the high sensibility of the galvanometer to reduce the current used for measuring the resistance of the coil of the platinum thermometer to $0\cdot002$ or $0\cdot003$ of an ampere, thus almost entirely eliminating any heating up of the coil by the measuring current, which was only kept on for a few seconds at most. Perhaps an equally satisfactory method of reducing to a minimum the uncertain effects of heating of the platinum coil by the measuring current would be to so regulate the current through the coil that the energy used up in heating the coil is always the same at all temperatures.

CONSTANTS OF PLATINUM THERMOMETER.

Determination of R_0 (the resistance of the platinum thermometer at 0° C.).—The ice used in these determinations was that manufactured by the Diamond Ice Co., of Baltimore, and was very clear and pure. The water used in this ice is first filtered; it is then frozen in large forms (long and *high*, but of shallow depth) from the sides towards the centre. When a thin sheet of water remains at the centre, the freezing is stopped and this water drawn off, thus insuring ice of great purity provided it is kept from contact with the freezing-mixture. Conductivity tests made on water obtained from this ice show a high specific resistance.

The thermometer was inserted in a mixture of this ice (pounded very fine) and distilled water. In the first and second determinations of R_0 a double-wall vessel was used, the inner vessel being nickel-plated. Ice and water were placed in both compartments. The outer vessel in all the zero determinations of the platinum, as well as for the mercurial thermometers, was wrapped with heavy boiler-felt to a thickness of about 4 cms.

The second and third determinations (Tables V. and VI.) were made in the same mixture of ice and distilled water, the only difference being that in the third the inner vessel was removed. The higher resistance obtained in this case (corresponding to about $0^{\circ}\cdot002$ C.) strongly suggests that, in a room whose temperature is 20° C. above that of the thermometer, the effect of radiation cannot be neglected. We were at first of the opinion that the removal of the inner

vessel would allow a more free circulation of the slightly warmed water coming from the top and sides, and thus the difference could be accounted for; but considering the distance of the thermometer-coil from the top and sides, and the compactness of the ice, this does not seem possible.

The following (Table III.) is a preliminary determination of R_0 , made merely to test the working of the apparatus. In this experiment ordinary ice and tap-water were used. The thermal regulator was not yet adjusted and the temperature of the coils was varying quite rapidly, yet the value of R_0 is almost identical with that obtained subsequently.

TABLE III.
Platinum Thermometer in Ice.

Time. Mar. 10, 1897.	Coils.	Bridge-Wire.	Temperature of Coils.	R_0 .
4.10	C, D, G, H = 255 Coil cor. -011	B.W. Reading +3.461 Cal. cor. ... -025 Zero cor. ... -011	20.80 Temp. Cor. +053	258.467
4.20	C, D, F ... = 260 Coil cor. +017	B.W. Reading -1.608 Cal. cor. ... +012 Zero cor. ... -011	20.90 Temp. Cor. +061	258.471
4.23	C, D, F, H = 265 Coil cor. +016	B.W. Reading -6.643 Cal. cor. ... +017 Zero cor. ... -011	20.90 Temp. Cor. +062	258.471
Mean				258.470

Determination of R_1 .—The resistance at 100° C., R_1 , was determined in an hypsometer in which the thermometer was screened on all sides (by polished metal screens) from the effects of radiation. The hypsometer was provided with a small water manometer to indicate the excess of pressure within. To determine the atmospheric pressure, the barometer already described was used in connexion with the standard B and D scale. The temperature of the barometric column was taken by means of two Bender and Hobein thermometers of normal glass, graduated to $\frac{1}{10}^{\circ}$ C., whose indications were reduced to the air-scale through a previous comparison with the platinum thermometer and a Tonnelot thermometer standardized at the Bureau International. The entire barometric column was wrapped to a thickness of several centimetres with asbestos paper and cotton to ensure

TABLE IV.

Platinum Thermometer in Ice.

Time.	Coils.	Bridge-Wire.	Temperature of Coils.	R ₀ .
12.20	C, D, F, = 260 Coil cor. ... +·017	B.W. Readg. -1·556 Cal. cor. +·011 Zero cor. -·016	20°·29 Temp. cor. +·020	258·476
12.27	C, D, G, H ... = 255 Coil cor. ... -·011	B.W. Readg. +3·506 Cal. cor. -·025 Zero cor. -·016	20·29 Temp. cor. +·019	258·473
12.32	C, D, F, = 260 Coil cor. ... +·017	B.W. Readg. -1·558 Cal. cor. +·011 Zero cor. -·016	20·30 Temp. cor. +·020	258·474
12.40	C, D, G, H ... = 255 Coil cor. ... -·011	B.W. Readg. +3·503 Cal. cor. -·025 Zero cor. -·016	20·33 Temp. cor. +·022	258·473
12.50	C, D, F, H ... = 265 Coil cor. ... +·016	B.W. Readg. -6·600 Cal. cor. +·046 Zero cor. -·016	20·34 Temp. cor. +·023	258·470
12.55	C, D, F = 260 Coil cor. ... +·017	B.W. Readg. -1·563 Cal. cor. +·011 Zero cor. -·016	20·35 Temp. cor. +·024	258·473
1.05	FI, C = 260 Coil cor. ... -·005	B.W. Readg. -1·544 Cal. cor. +·011 Zero cor. -·016	20·37 Temp. cor. +·025	258·471
1.15	FI, D, E, F, G, H = 255 Coil cor. ... +·179	B.W. Readg. +3·306 Cal. cor. -·024 Zero cor. -·016	20·38 Temp. cor. +·025	258·470
1.20	C, D, F = 260 Coil cor. ... +·017	B.W. Readg. -1·566 Cal. cor. +·011 Zero cor. -·016	20·39 Temp. cor. +·026	258·472
Mean				258·472

Note.—In this determination of R₀ a double-wall vessel was used, with ice in both inner and outer compartments. Outer vessel wrapped with heavy boiler-felt.

uniformity of temperature. The thermometers were inserted between the wrappings touching the glass of the barometer. The temperature of the standard scale was taken by another similar thermometer. A series of measurements of the barometric height were made by one observer while the other

TABLE V.

Platinum Thermometer in Ice.

Time. May 6.	Coils.	Bridge-Wire.	Temperature of Coils.	R ₀ .
3.45	C, D, F, H... = 265 Coil cor. ... +·016	B.W. Readg. -6·589 Cal. cor. +·047 Zero cor. -·003	19·94 Temp. cor. -·004	258·467
3.55	C, D, G, H... = 255 Coil cor. ... -·011	B.W. Readg. +3·515 Cal. cor. -·025 Zero cor. -·003	19·90 Temp. cor. -·006	258·470
4.00	C, D, F, H... = 265 Coil cor. ... +·016	B.W. Readg. -6·588 Cal. cor. +·047 Zero cor. -·003	19·90 Temp. cor. -·007	258·468
4.05	C, D, F = 260 Coil cor. ... +·017	B.W. Readg. -1·550 Cal. cor. +·011 Zero cor. -·003	19·91 Temp. cor. -·005	258·470
4.12	FI, C = 260 Coil cor. ... -·005	B.W. Readg. -1·527 Cal. cor. +·011 Zero cor. -·003	19·92 Temp. cor. -·005	258·471
4.20	FI, D, E, F, G, H = 255 Coil cor. ... +·179	B.W. Readg. +3·323 Cal. cor. -·024 Zero cor. -·003	19·93 Temp. cor. -·005	258·470
4.28	C, D, F = 260 Coil cor. ... +·017	B.W. Readg. -1·550 Cal. cor. +·011 Zero cor. -·003	19·94 Temp. cor. -·004	258·471
4.40	C, D, F = 260 Coil cor. ... +·017	B.W. Readg. -1·552 Cal. cor. +·011 Zero cor. -·003	19·90 Temp. cor. -·006	258·467
Mean				258·469

In this determination a double-wall vessel wrapped in felt was again used.

observer measured the resistance of the platinum thermometer in steam, with different combinations of coils and bridge-wire. A time-chart of the barometric pressure served to give the pressure at the instant when the resistance was taken. Usually the observations of the barometric height were taken simultaneously, as nearly as possible, with the resistance of the platinum thermometer, for the variations of the atmospheric pressure seem to take place suddenly, and not uniformly as

Phil. Mag. S. 5. Vol. 48. No. 290. July 1899. C

TABLE VI.

Platinum Thermometer in Ice.

Time. May 6.	Coils.	Bridge-Wire.	Temperature of Coils.	R ₀ .
5.00	C, D, F, H... = 265 Coil cor. ... +·016	B.W. Readg. -6·589 Cal. cor. +·047 Zero cor. -·003	19·97 Temp. cor.--·002	258·469
5.10	C, D, F = 260 Coil cor. ... +·017	B.W. Readg. -1·554 Cal. cor. +·011 Zero cor. -·003	19·97 Temp. cor.--·002	258·469
5.15	C, D, G, H... = 255 Coil cor. ... -·011	B.W. Readg. +3·509 Cal. cor. -·025 Zero cor. -·003	19·97 Temp. cor.--·002	258·468
5.20	FI, C = 260 Coil cor. ... -·005	B.W. Readg. -1·527 Cal. cor. +·011 Zero cor. -·003	19·98 Temp. cor.--·001	258·475
5.24	FI, D, E, F, G, H = 255 Coil cor. ... +·179	B.W. Readg. +3·320 Cal. cor. -·024 Zero cor. -·003	19·98 Temp. cor.--·001	258·471
5.29	FI, C = 260 Coil cor. ... -·005	B.W. Readg. -1·527 Cal. cor. +·011 Zero cor. -·003	19·99 Temp. cor.--·001	258·475
5.35	C, D, F = 260 Coil cor. ... +·016	B.W. Readg. -1·550 Cal. cor. +·011 Zero cor. -·003	20·00 Temp. cor. ·000	258·473
Mean				258·471

This determination of R₀ was made immediately after the preceding, the inner nickel-plated calorimeter being first removed. The same freezing-mixture was used, and the contacts of the platinum thermometer had not been touched.

indicated by a time-chart. Moreover, the platinum thermometer seemed to respond to the changes more rapidly than the barometer, which appeared to have an appreciable lag.

The observed barometric height, corrected for temperature and errors of graduation of the standard metre-bar, was reduced to 0° C., latitude 45°, sea-level. From the reduced barometric height the temperature of the steam was obtained by interpolation in Broch's tables of the pressure of aqueous vapour, recalculated from Regnault's experiments (*Trav. et Mém. du Bur. Int. des Poids et Mes.* t. i.).

From the observed resistance of the platinum thermometer and the corresponding temperature of the steam, the resistance R_1 at 100°C. was deduced as follows:—

For example, from the first observation in the following table (VII.) we have

$$\begin{cases} \text{Resistance } R = 358.053 \text{ mean box-units (Box No. 7).} \\ \text{Red. barometric ht.} = 755.19 \text{ mm.} \end{cases}$$

Corresponding temperature, t , of steam, obtained from Broch's tables $= 99.822$.

The correction, dt/dp , to be applied to the air temperature, t , of the boiling-point of water to reduce to the boiling-point at 760 mm. of mercury at 0°C. , latitude 45° , sea-level, is therefore 0.0368 per mm. of mercury.

To find the corresponding correction to be applied to the platinum temperature we have

$$t - pt = \delta \left\{ \left(\frac{t}{100} \right)^2 - \frac{t}{100} \right\},$$

differentiating

$$\frac{d.pt}{dp} = \frac{d.pt}{dt} \frac{dt}{dp} = .0368 \left\{ 1 - \delta \frac{2t - 100}{10000} \right\}.$$

We may assume an approximate value of $\delta = 1.50$ (from a previous knowledge of the constants of platinum thermometers) as sufficiently accurate for the purpose of this reduction.

If it happens that the values of δ assumed differ greatly from the final value obtained from the standardization, it may be necessary to substitute this new value of δ ; we have then

$$\frac{d.pt}{dp} = 0.0362 \text{ per mm. of mercury.}$$

From the equation defining platinum temperature,

$$pt = \frac{R - R_0}{R_1 - R_0} \times 100,$$

we have

$$R_1 = R_0 + \frac{100}{pt} (R - R_0).$$

Reducing $t = 99.822$ to the corresponding temperature on the platinum scale, we have $pt = 99.826$.

Hence

$$\begin{aligned} R_1 &= 258.471 + 100 \frac{(358.053 - 258.471)}{99.826} \\ &= 358.227 \text{ mean box-units.} \end{aligned}$$

TABLE VII.—Platinum Thermometer in Steam.

Time.	Coils.	Bridge-Wire.	Temp. of Coils.	Barom. pressure reduced to 0° C., sea-level, lat. 45°.	Results at obs. pres.	R ₁₀₀ .
May 4. 2.37	B, E = 360 Coil cor. ... +.043	B.W. Readg. -1.976 Cal. cor. +.014 Zero cor. -.003	19° 73 -025	755.19	358.053	358.227
2.46	B, F, G, H ... = 355 Coil cor. ... -.003	B.W. Readg. +3.103 Cal. cor. -.023 Zero cor. -.003	19° 75 -023		358.052	358.226
3.08	FI, C, D, F ... = 360 Coil cor. ... +.054	B.W. Readg. -1.993 Cal. cor. +.014 Zero cor. -.003	19° 79 -020		358.052	358.226
3.20	B, E = 360 Coil cor. ... +.043	B.W. Readg. -1.982 Cal. cor. +.014 Zero cor. -.003	19° 80 -019		358.053	358.227
3.34	FI, C, G, D, H = 355 Coil cor. ... +.026	B.W. Readg. +3.080 Cal. cor. -.022 Zero cor. -.003	19° 80 -019		755.32 358.062	358.232
May 7. 12.17	B, E = 360 Coil cor. ... +.043	B.W. Readg. -1.864 Cal. cor. +.014 Zero cor. -.003	20° 00 000	758.95	358.190	358.228
12.25	B, F, G, H ... = 355 Coil cor. ... -.003	B.W. Readg. +3.220 Cal. cor. -.023 Zero cor. -.003	20° 01 +001		358.192	358.230
12.34	B, E, H = 365 Coil cor. ... +.042	B.W. Readg. -6.900 Cal. cor. +.049 Zero cor. -.003	20° 02 +002		358.192	358.230
May 7. 3.55	B, E = 360 Coil cor. ... +.043	B.W. Readg. -1.876 Cal. cor. +.014 Zero cor. -.003	20° 35 +033	759.36	358.211	358.235
4.05	B, F, G, H ... = 355 Coil cor. ... -.003	B.W. Readg. +3.214 Cal. cor. -.022 Zero cor. -.003	20° 36 +033		358.219	358.236
4.15	B, E = 360 Coil cor. ... +.043	B.W. Readg. -1.870 Cal. cor. +.014 Zero cor. -.003	20° 37 +035	759.54	358.219	358.236
4.18	B, E = 360 Coil cor. ... +.043	B.W. Readg. -1.870 Cal. cor. +.014 Zero cor. -.003	20° 37 +035		358.219	358.236
Mean						358.231

∴ R₁=358.231 mean box-units (Box No. 7).

From the preceding observations we have :—

$R_0 = 258\cdot471$ (mean box-units, at 20° C., of box No. 7),

$R_1 = 358\cdot231$ „ „ „ „

This gives

$$\frac{R_1}{R_0} = 1\cdot385962.$$

Mr. E. H. Griffiths was kind enough to standardize this thermometer with great care before sending it, and the values found by him (as a mean of many observations with different combinations of coils and bridge-wire of Box No. 6) were :—

January 21, 1896.

$R_1 = 358\cdot075$ mean box-units Box 6,

$R_0 = 258\cdot366$ „ „ „

$Pt_{sulphur} = 422\cdot54$ (barometer at $771\cdot35$ mm.):

hence

$$\delta = 1\cdot490.$$

“In this determination the average temperature of the box was about $15^\circ\cdot4$, but it was rapidly varying. Moreover, the temperature correction was very large, the box being right at 20° C.”

January 23, 1896.

$R_1 = 258\cdot362$,

$R_0 = 358\cdot078$,

$R_s = 679\cdot510$ (barometer at $769\cdot20$).

Hence

$$\frac{R_1}{R_0} = 1\cdot38596,$$

$$\delta = 1\cdot491.$$

The remarkable agreement of our value of $\frac{R_1}{R_0}$ with that found by Griffiths renders it unnecessary for us to determine the resistance of the thermometer in boiling sulphur (the third temperature usually employed), and we therefore accepted the value of δ , $1\cdot491$, given by Griffiths, as correct.

[It is interesting to note that for the standardization given under date of January 21, the value of

$$pt = -\frac{R_0}{R_1 - R_0} \times 100 = -259^\circ\cdot12,$$

which corresponds to zero resistance of the platinum thermometer, gives for the absolute zero [by using $\delta = 1.490$ in $t - pt = \delta\{(t/100)^2 - t/100\}] - 274^{\circ}.44$. Similarly the standardization of January 23 gives the absolute zero $-274^{\circ}.41$.]

STANDARDIZATION OF PLATINUM THERMOMETER.

The details of the elaborate comparisons by Mr. Griffiths, on which the constant, δ , of this thermometer is based, are given below (Tables VIII., IX., X.). The resistance-box (No. 6) used in this standardization was similar to the one described in 'Nature,' Nov. 14, 1895.

"For all purposes of comparison between the results obtained here and in Baltimore, we only require the values of $\frac{R_1}{R_0}$ and δ , the magnitude of the unit used being of no consequence. I, however, give the observations in full, to show the probable order of accuracy."

January 21, 1896.

"On the above date a preliminary series of observations was made in ice, steam, and sulphur-vapour. The observations, however, were hurried, the resistance-box was not contained in its tank, and its temperature was about $15^{\circ}.50$ C. Thus the temperature correction was large. I did not regard the work as anything but preliminary, and will therefore only give results (the resistances are corrected for temperature of box, &c.) :—

$$R' = 358.480 \text{ when barometer}^* = 771.24,$$

$$R_0 = 258.366,$$

$$R_s = 679.680 \text{ when barometer} = 771.35.$$

Hence we get

$$R_1 = 358.075$$

$$R_0 = 258.366$$

$$R_1 = 358.075$$

$$\frac{R_1}{R_0} = 1.38592$$

$$\delta = 1.490$$

* The barometer is in each case corrected for temperature, scale errors, and for value of g to sea-level, latitude 45° .

TABLE VIII.

Platinum Thermometer in Ice.

Time. Jan. 23.	Coils.	Bridge-Wire Reading.	Temperature of Coils.	R ₀ . (Box No. 6.)
5.19	C, D, F, H ... = 265 Coil cor. ... -·102	B.W. Readg. -6·602 Cal. cor. +·109 Zero cor. -·078	⁰ 20·50 Cor. +·034	258·361
5.21	„	B.W. Readg. -6·600 Cal. cor. +·109 Zero cor. -·078	„	258·363
5.24	„	B.W. Readg. -6·601 Cal. cor. +·109 Zero cor. -·078	„	258·362
5.27	C, D, F ... = 260 Coil cor. ... -·094	B.W. Readg. -1·526 Cal. cor. +·026 Zero cor. -·078	„	258·362
5.28	„	B.W. Readg. -1·524 Cal. cor. +·026 Zero cor. -·078	„	258·364
5.30	C, D, G, H ... = 255 Coil cor. ... -·078	B.W. Readg. +3·546 Cal. cor. -·060 Zero cor. -·078	„	258·364
5.32	„	B.W. Readg. +3·545 Cal. cor. -·060 Zero cor. -·078	„	258·363
5.35	C, D, G = 250 Coil cor. ... -·079	B.W. Readg. +8·626 Cal. cor. -·144 Zero cor. -·078	„	258·359
5.37	„	B.W. Readg. +8·627 Cal. cor. -·144 Zero cor. -·078	„	258·360
Mean				258·362

Determination of R₁, R₀, and FI.

“ We have

$$R' = 358·416 \text{ when bar.} = 769·39, \therefore \text{b.p.} = 100^{\circ}344,$$

$$R = 258·362$$

$$R' - R_0 = 100·054 \text{ for dif. in temp.} = 100^{\circ}344.$$

TABLE IX.
Platinum Thermometer in Steam.

Time. Jan. 23.	Coils.	Bridge-Wire Reading.	Temperature of Coils.	Baro- meter (corr.).	R'. (Box No. 6.)
4.36	C, D, F, H, FI = 365 Coil cor. ... -·179	B.W. Readg. -6·466 Cal. cor. +·106 Zero cor. -·078	20·32 Cor. +·030	mm. 769·40	358·413
4.43	"	B.W. Readg. -6·469 Cal. cor. +·106 Zero cor. -·078	20·38 Cor. +·035	"	358·415
4.46	C, D, F, FI = 360 Coil cor. ... -·171	B.W. Readg. -1·394 Cal. cor. +·023 Zero cor. -·078	20·40 Cor. +·037	"	358·417
4.51	"	B.W. Readg. -1·397 Cal. cor. +·023 Zero cor. -·078	20·42 Cor. +·039	769·45	358·416
4.55	C, D, G, H, FI = 355 Coil cor. ... -·164	B.W. Readg. +3·674 Cal. cor. -·061 Zero cor. -·078	20·45 Cor. +·042	"	358·413
4.59	"	B.W. Readg. +3·676 Cal. cor. -·061 Zero cor. -·078	20·47 Cor. +·044	769·40	358·417
5.6	C, D, G, FI = 350 Coil cor. ... -·156	B.W. Readg. +8·752 Cal. cor. -·146 Zero cor. -·078	20·49 Cor. +·046	769·35	358·418
5.9	"	B.W. Readg. +8·750 Cal. cor. -·146 Zero cor. -·078	20·49 Cor. +·046	769·30	358·416
Mean				769·39	358·416

R' = 358·416 mean box-units (Box No. 6) when the barometer is
769·39 mm.

" Hence mean $\frac{\delta R}{\delta t} = 0·99711$.

" Now $\frac{\delta pt}{\delta t}$ at 100° = 0·985.

$$\therefore \delta R' \text{ at } 100^\circ = ·9971 \times ·985 = ·0·9821,$$

$$\therefore \delta R' \text{ for } 0^\circ·344 = ·9821 \times ·344 = 0·338,$$

$$\therefore R_1 = 358·416 - 0·338 = 358·078.$$

TABLE X.

Platinum Thermometer in Sulphur Vapour.

In sulphur vapour all the precautions must be taken which are mentioned on pp. 144-147, Phil. Trans. Roy. Soc., A, 1891, vol. clxxxii.

Time. Jan. 23.	Coils.	Bridge-Wire Reading.	Temperature of Coils.	Baro- meter (corr.).	R*. (Box No 6.)
6.15	A, E = 680 Coil cor. ... +.193	B.W. Readg. -·675 Cal. cor. +·009 Zero cor. -·078	$\overset{\circ}{20}\cdot36$ Cor. +·064	mm. 769·20	679·513
6.18	„	B.W. Readg. -·671 Cal. cor. +·009 Zero cor. -·078	20·35 Cor. +·062	„	679·515
6.21	„	B.W. Readg. -·672 Cal. cor. +·009 Zero cor. -·078	„	„	679·514
6.25	A, E, H = 685 Coil cor. ... +.185	B.W. Readg. -5·754 Cal. cor. +·095 Zero cor. -·078	„	769·15	679·510
6.26	„	B.W. Readg. -5·755 Cal. cor. +·095 Zero cor. -·078	„	769·20	679·509
6.30	A, F, G, H ... = 675 Coil cor. ... +·068	B.W. Readg. +4·527 Cal. cor. -·076 Zero cor. -·078	„	„	679·503
6.32	„	B.W. Readg. +4·528 Cal. cor. -·076 Zero cor. -·078	„	769·25	679·504
Mean				769·20	679·510

R*=679·510 mean box-units (Box No. 6) when barometer=769·20 mm.

“Hence $R_1 = 358\cdot078,$
 $R_0 = 258\cdot362,$
 $FI = 99\cdot716,$
 $R_1/R_0 = 1\cdot38596.$

Determination of δ .

" $R_s = 679.510$ when barometer = 769.20 mm.

$$P_t_s = \frac{679.510 - 258.362}{99.716} = 422^{\circ}.35.$$

" Now the boiling-point of sulphur at 769.20 mm.

$$= 444^{\circ}.53 + 0^{\circ}.082 \times 9.20 = 445^{\circ}.28.$$

$$\therefore 445^{\circ}.28 - 422^{\circ}.35 = \delta \{ (4.453)^2 - 4.453 \},$$

$$\therefore \delta = 1.491.$$

" *Final Results* :—

$$R_1 = 358.078$$

$$R_0 = 258.362$$

$$\hline FI = 99.716$$

$$\frac{R_1}{R_0} = 1.38596$$

$$\delta = 1.491."$$

RESULTS OF COMPARISONS.

In order to show in detail the method of comparison, we have selected at random the results of a single comparison of Baudin 6166 with the platinum thermometer (Table XI.).

In Columns II., III., and IV. are shown respectively the coils used, the bridge-wire reading, and temperature of coils, for the measurement of the resistance of the platinum thermometer. The corresponding stem-reading on the mercurial thermometer, taken at the same instant, is shown in column VIII. The temperature of the air near the portion of the stem of the Baudin thermometer projecting from the calorimeter, and of the water in the small water-jacket around the stem, where it emerges from the calorimeter, is shown in columns IX. and X.

The resistance of the platinum thermometer is deduced from the observed quantities, shown in columns II., III., and IV., as follows :—To the nominal value of the coils is applied the "coil-correction" to reduce to mean box-units; these are taken from Table I., and are designated "coil cor." in column II.; a further correction for the temperature of the coils must be applied to reduce to "the mean box-unit of Box No. 7 at 20° C."; these corrections are applied in column IV. To the bridge-wire reading is applied the "calibration correction" to reduce the observed reading to mean box-units; to this must be further applied the "zero correction," which takes into account the amount by which the zero of the vernier differs from the zero of the scale when the bridge is balanced with the intervals C_1C_2 and P_1P_2 short-

circuited and all plugs replaced with care ; these corrections are applied in column III.

By the application of these corrections, taken with proper signs, we get R , the resistance of the platinum thermometer, shown in column V.

The platinum temperature, shown in column VI., is deduced from the observed value of R by substitution in the formula defining platinum temperature, remembering that $R_0=258.471$ and $R_1=358.231$ "mean box-units at $20^\circ \text{C}.$ "

From the platinum temperature the corresponding temperature on the air-scale, shown in column VII., is deduced by the formula

$$t-pt=1.491 \left\{ \left(\frac{t}{100} \right)^2 - \frac{t}{100} \right\}.$$

In order to facilitate these reductions, a table was constructed giving the value of $(t-pt)$ for every degree pt on the platinum scale.

Six determinations of the zero of Baudin 6166 gave the following results:—

	mm.
	24.159
	24.147
	24.151
	24.148
	24.151
	24.154
Mean . . .	24.15

Rowland's zero (table xx. p. 116, Proc. Amer. Acad. of Arts and Sciences, xv. 1879) was 20.43 .

Temperatures by Baudin 6166 are reduced as follows:—To the mean stem-reading is applied the stem-correction which is divided into two parts, one portion of the stem extending from 28 mm. to 99 mm. being assumed to be at the temperature of the water in the surrounding jacket, the remainder of the stem at the temperature of the surrounding air. This correction is applied in column XI.; from the corrected stem-reading the corresponding temperature on Rowland's air thermometer is obtained from the results of his comparisons of these thermometers with the air thermometer given in his tables xviii., xix., and xx. The temperatures thus obtained must be further corrected by the rise of the zero since Rowland's comparisons : this is shown in column XII., which gives the mean temperature on the Rowland air-scale (obtained through 6166), corresponding to the mean temperature on the Callendar-Griffiths air-scale given in column VII.

TABLE XI.
Comparison of Baudin 6166 with Platinum Thermometer.

Time, May 28,	Coils.	Bridge-Wire.	Temperature of Coils.	Platinum Resist- ance. R.	Plat. Temp. $R - R_0 \times 100$ $p = \frac{R - R_0}{R_0}$	Pt. Temp. red. to Air-Scale. $t - pt$ $= t \left\{ \frac{t}{100} - \frac{t^2}{100} \right\}$	Stem Reading on 6166, mm.	Temp. of Air near Stem.	Temp. of Water in Jacket.	Corrected Mean Stem Reading.	Corresponding Temp. on Rowland's Air-Scale.
1.39	C, D, F = 260 Coil cor. ... + 017	B.W. Readg. + 3737 Cal. cor. ... - 026 Zero cor. ... - 003	19° 64 Temp. cor. - 024	263-701	5° 243	5° 170	90-21				
1.43	FI, C = 260 Coil cor. ... - 005	B.W. Readg. + 3784 Cal. cor. ... - 027 Zero cor. ... - 003	19° 64 Temp. cor. - 024	263-725	5° 267	5° 193	90-49	20° 22	17° 3		
1.46	C, D, F, H = 265 Coil cor. ... + 016	B.W. Readg. - 1248 Cal. cor. ... + 009 Zero cor. ... - 003	19° 64 Temp. cor. - 025	263-749	5° 291	5° 217	90-74	20° 22	17° 6	91-20 Stem cor. - 12	
1.54	FI, C, H ... = 265 Coil cor. ... - 006	B.W. Readg. - 1182 Cal. cor. ... + 008 Zero cor. ... - 003	19° 64 Temp. cor. - 025	263-792	5° 334	5° 260	91-43	20° 22	17° 4	91-08	5° 522 Cor. for Rise of - 28. Zero.
2.00	C, D, F, G... = 270 Coil cor. ... + 021	B.W. Readg. - 6221 Cal. cor. ... + 014 Zero cor. ... - 003	19° 65 Temp. cor. - 025	263-816	5° 338	5° 283	91-83	20° 22			5° 522 Cor. for Rise of - 28. Zero.
2.08	C, D, F = 260 Coil cor. ... + 017	B.W. Readg. + 3905 Cal. cor. ... - 027 Zero cor. ... - 003	19° 67 Temp. cor. - 022	263-870	5° 412	5° 337	92-48	20° 22			5° 522 Cor. for Rise of - 28. Zero.
2.40	C, D, F, G... = 270 Coil cor. ... + 021	B.W. Readg. - 1350 Cal. cor. ... + 009 Zero cor. ... - 003	19° 69 Temp. cor. - 022	263-855	10° 209	10° 075	153-22				
2.44	FI, C, G ... = 270 Coil cor. ... - 001	B.W. Readg. - 1316 Cal. cor. ... + 009 Zero cor. ... - 003	19° 69 Temp. cor. - 022	268-667	10° 221	10° 086	153-31	20° 80	17° 92		
2.50	FI, C, H ... = 265 Coil cor. ... - 006	B.W. Readg. + 3730 Cal. cor. ... - 026 Zero cor. ... - 003	19° 69 Temp. cor. - 021	268-674	10° 228	10° 093	153-43	20° 70	18° 02	153-46 Stem cor. - 09	
2.56	C, D, F, G... = 270 Coil cor. ... + 021	B.W. Readg. - 1333 Cal. cor. ... + 009 Zero cor. ... - 003	19° 71 Temp. cor. - 020	268-674	10° 228	10° 093	153-54	20° 75	18° 20	153-28	10° 361 Cor. for Rise of - 28. Zero.
3.02	C, D, F, H = 265 Coil cor. ... + 016	B.W. Readg. + 3711 Cal. cor. ... - 026 Zero cor. ... - 003	19° 72 Temp. cor. - 019	268-679	10° 233	10° 098	153-59	20° 75			
3.08	C, D, F, G, H = 275 Coil cor. ... + 020	B.W. Readg. - 6360 Cal. cor. ... + 045 Zero cor. ... - 003	19° 73 Temp. cor. - 019	268-683	10° 237	10° 102	153-66	20° 75			10° 069

[illegible]

COMPARISON OF PLATINUM THERMOMETER WITH
TONNELOT 11801.

This comparison was made by Dr. W. S. Day and the authors, in order to bring the results of the different standards of thermometry into comparison, and thus, if possible, to account for the differences in the values of the capacity for heat of water obtained by different observers.

The comparison was carried out in a tank especially designed by Dr. Day for the comparison of mercurial thermometers in a horizontal position. This tank consists of a long rectangular copper box 98 cm. long, 21 cm. wide, and 23 cm. deep, with a movable cover, holding a piece of plate-glass, through which the stem-readings of the thermometers could be observed by means of a micrometer-telescope which slides along parallel ways supported from the cover of the tank. Within this copper tank was another small brass box, 10.5 cm. long, 8 cm. wide, and 5 cm. deep, in which were placed side by side the bulbs of the thermometers to be compared. This box shielded the bulbs from the effects of convection-currents; and in this way the thermometers were kept at a constant temperature for a considerable time, for it required an interval of some minutes for a very appreciable change in the temperature of the large mass of surrounding water to produce a minute change in the temperature of the enclosed water owing to the absence of convection-currents. This brass box was provided with two movable lids on the top and two on the bottom, which could be opened and closed from the outside of the tank. The large tank was provided with three paddle-wheels, operated from the outside, which produced efficient stirring. The entire tank was then placed inside a wooden box, the interspace between the walls of the tank and box being loosely filled with cotton-wool. The greater portion of the lid, with the exception of the plate-glass, was covered with a layer of thick felt. The tank and its contents could be raised to any desired temperature by allowing steam to flow through a copper tube coiled along the bottom of the tank.

As the platinum thermometer is peculiarly adapted to show minute changes in temperature, the efficiency of this tank was exposed to a severe test in these comparisons. It was thus ascertained that if the lids of the inner brass box were opened, the entire contents of the tank thoroughly stirred, and the lids then tightly closed, the temperature would remain constant to within $0^{\circ}001$ for at least several minutes, for a difference of 20° C. between the temperature of the room and the water in the tank.

Constants of Tonnelot 11801.

Tonnelot's 11800, 11801, and 11811 were made of French "verre-dur," with transparent stems divided into tenths of a degree, about December 1895. They were subsequently sent to the Bureau International, where a most exhaustive study of them was made by M. Guillaume, and comparisons made at several different temperatures with their standards. These standards have been compared by M. Chappuis with the standard nitrogen and hydrogen thermometer (*Trav. et Mém. du Bur. Int. des Poids et Mesures*, tom. vi. 1888), so that the scale of these thermometers can at once be reduced to the nitrogen or hydrogen scale.

We can only briefly outline the methods used at the Bureau International for the standardization of thermometers; for to do full justice to their beautiful and painstaking researches on thermometry would not lie within the limits of the present communication.

The thermometers are first roughly examined for uniformity of bore and graduation. If these are not sufficiently accurate, the thermometer is rejected. If these are found satisfactory, a *calibration* is next made by measuring the lengths of suitable columns of mercury in different portions of the scale. This determines the correction that must be applied to certain "principal points" of the scale, to reduce the scale-reading to what the reading would have been if the bore had been perfectly uniform. In these thermometers, whose range includes 0° and 100° , the corrections are usually so calculated that the calibration corrections for 0° and 100° are zero.

The next constants determined are the external and internal pressure-coefficients, usually denoted by β_e and β_i . The *external pressure-coefficient* is determined by subjecting the thermometer to known variations in pressure and observing the corresponding change in the stem-reading. The external pressure-coefficient, β_e , is then the change in degrees produced by a change in the external pressure of 1 mm. of mercury. The *internal pressure-coefficient*, β_i , is deduced from β_e by adding $0^{\circ}.0000154$, a quantity depending on the difference of compressibility of mercury and "verre-dur."

The *fundamental interval*, i. e., the number of scale-divisions between 0° and 100° , is then determined by observing the "fixed points" (in steam and ice) of the thermometer. This gives the value of the degree or the scale-division (if the graduation is arbitrary) on the centigrade scale.

From these known constants the temperature on the centigrade scale of this particular thermometer can be de-

terminated. But a scale of temperature defined in this way would vary with each different specimen of glass used, and would not be exactly the same even for the same kind of glass. The mean scale of a number of French hard-glass thermometers kept at the Bureau International, which have been compared directly with the gas-thermometers (H, N, and CO₂), defines temperature on the French hard-glass ("verre-dur") scale. The slight deviations of the hard-glass scale of any other Tonnelot thermometer which is compared with these standards is observed, and these corrections serve to reduce the indications of this thermometer to the mean French hard-glass scale.

Method of Comparison.

The Tonnelot thermometer was adjusted with its stem in a horizontal position in the comparison-tank, and with its bulb near the centre of the inner brass box. The micrometer-telescope was then adjusted by means of a level, so that its axis of collimation was perpendicular to the stem of the thermometer. When these adjustments had been made, no certain difference could be detected between observations made with the divisions in front of the stem and those made with divisions back of the stem. The glass stem of the platinum thermometer passed from the outside through the sides of the wooden box and comparison-tank into the brass box, where its coil was almost in contact with the bulb of the mercury thermometer. Before taking an observation the top and bottom of this box were opened, and the entire contents of the tank thoroughly stirred; the lids were then tightly closed, and a series of observations taken. A single series included a setting on the division below the meniscus, on the meniscus, and on the division above the meniscus, repeated three times, but in inverse order. No certain variation of temperature was ever detected during the time required to take a series of observations. The box containing the bulbs was then opened, the water again thoroughly stirred, the box closed, and a similar series of observations taken. If the temperature of the water in the tank was below that of the room, after each stirring there was a slight rise in temperature: if the temperature of the water in the tank was above that of the room, there would be a slight decrease after each stirring: steam was therefore passed through the heating-coil to make sure of a rising meniscus. Usually four such series of observations were taken at each temperature. Immediately after the com-

parisons at one temperature were completed, the Tonnelot was removed from the tank and its zero quickly determined (in a mixture of pure ice and distilled water) before any recovery of the zero had taken place. In the zero determinations the thermometer was adjusted vertical and the axis of collimation of the micrometer-telescope horizontal. Occasional barometer readings were taken to reduce the indications of the Tonnelot thermometer to standard pressure.

While one observer was taking observations on the Tonnelot thermometer, another observer was taking simultaneous observations of the resistance of the platinum thermometer.

RESULTS OF COMPARISON.

In order to show in detail the methods used, &c., we give below the reduction of a single series of observations at one temperature.

Observations :—

Platinum Thermometer { Coils FI, C, F.
Bridge-wire reading = -1.015 .
Temperature of coils $20^{\circ}31$.

Micrometer Readings.

	(1)	(2)	(3)
20°4	15.587	15.579	15.586
Meniscus ...	15.109	15.102	15.100
20°5	14.995	14.995	14.989
Tonnelot Thermometer No. 11801. {	Barometer (brass scale)... 762.0 mm.		
	Temperature 19°6		
	Zero determination :—		
0°0	12.610	12.612	12.605
Meniscus ...	12.428	12.430	12.426

Reductions :—

		Zero.
Stem-reading of Tonnelot 11801 ...	20.482	+0.031
Calibration correction	+0.004	0.000
Correction for external pressure due to 57 mm. of water + excess of atmospheric pressure. }	-0.001	-0.001
Internal pressure correction : thermometer horizontal in comparison tank, vertical in zero tank. }	0.000	+0.002
Zero correction	-0.038	+0.038
	20.447	
Correction to fundamental interval...	0.000	

Temperature on scale of 11801 $20^{\circ}\cdot447$
 Correction M, to reduce to mean }
 French hard-glass scale. } $-0\cdot006$

Temperature on mean hard-glass } $20^{\circ}\cdot441$
 scale. }
 Correction to hydrogen scale..... $-0\cdot086$
 „ „ nitrogen „ $-0\cdot076$

Temperature on hydrogen scale ... $20^{\circ}\cdot355$
 „ „ nitrogen „ ... $20\cdot365$

Coils FI, C, F = 280
 Coil corrections (Table I.)..... $+0\cdot026$
 Temperature correction $+0\cdot023$
 Bridge-wire reading $-1\cdot015$
 Calibration correction (Table II.). $+0\cdot007$
 Zero correction of bridge-wire ... $-0\cdot006$

$279\cdot035$ mean box-units.

$$\therefore pt = \frac{279\cdot035 - 258\cdot471}{358\cdot231 - 258\cdot471} \times 100$$

$$= 20^{\circ}\cdot613 ;$$

$$\therefore t = 20^{\circ}\cdot372 \text{ (air-scale).}$$

The results of the two independent series of comparisons of the Platinum Thermometer with Tonnelot 11801 are given in the following table (XII.).

TABLE XII.

Comparison of Platinum Thermometer and Tonnelot 11801.

Tonnelot 11801. Nitrogen- Scale.	Pt. Therm. Air-Scale.	Δ [Cor. to reduce Pt.-Air Scale to Paris Nitrogen-Scale.]	Tonnelot 11801. Nitrogen- Scale.	Pt. Therm. Air-Scale.	Δ [Cor. to reduce Pt.-Air Scale to Paris Nitrogen-Scale.]
$7^{\circ}\cdot804$	$7^{\circ}\cdot800$	$+0\cdot04$	$10^{\circ}\cdot197$	$10^{\circ}\cdot186$	$+0\cdot11$
12-941	12-941	$\cdot000$	20-519	20-523	$-0\cdot04$
18-332	18-327	$+0\cdot05$	23-400	23-391	$+0\cdot09$
23-106	23-114	$-0\cdot08$	25-568	25-571	$-0\cdot03$
33-703	33-704	$-0\cdot01$	30-004	30-005	$-0\cdot01$
39-999	40-001	$-0\cdot02$	43-817	43-820	$-0\cdot03$

These results are also plotted in the form of a curve in fig. 5.

Fig. 4.

The results of these comparisons can best be shown by means of the accompanying curves.

Fig. 4 shows the result of each of the independent comparisons of Professor Rowland's Baudin thermometers with the platinum-resistance thermometer. Abscissæ represent temperature on the centigrade scale, and ordinates the corresponding corrections that must be added to Rowland's air-scale, as determined by the Baudin thermometers, to reduce to the Callendar-Griffiths' air-scale. The almost constant difference between the curves of May 18th and May 24th suggests at once a constant error affecting the entire series of one or the other comparisons. This difference is, however, not of any great importance; for in the determination of the mechanical equivalent we are practically only concerned with temperature ranges over the interval 5° to 35° , and these are practically identical on either curve. After looking over the results, we have attributed these differences to a slight error in the value of the temperature-coefficient of the coils of the box, which, on account of the high temperature of the room, could not be kept at 20° in the comparisons of May 24th and May 21st. This conclusion is further strengthened by the fact that the comparison of May 28th, when the box was again near 20° , practically coincides with that of April 10th.

A *résumé* of all the comparisons between the Platinum and Baudin thermometers is also given in the table opposite.

Fig. 5 gives the final correction curves for each of the Baudin thermometers. The ordinates of curves I. (mean of the individual comparisons shown in fig. 4) give the corrections that must be added to Rowland's air-scale, as determined from the Baudin thermometers, to reduce to the Callendar-Griffiths' air-scale. To pass from the air-scale to the absolute scale of temperature use was made of Rowland's table xvii. (p. 114, Proc. Am. Acad. xv. 1879). In this way curves II. were obtained which give the corrections to reduce Rowland's absolute scale to the Callendar-Griffiths' air-scale.

The results of the comparison of the platinum-resistance thermometer with Tonnelot 11801, which had been standardized at the Bureau International, are shown in curve A, in which ordinates represent the corrections that must be added to the Callendar-Griffiths' air-scale to reduce to the Paris nitrogen-scale as given by Tonnelot 11801. The close agreement of these two scales is a strong confirmation of the accuracy of the platinum-air interpolation formula. Indeed such close agreement must be partly fortuitous, as we have certainly no right to expect so close an agreement, considering the difficulties of gas-thermometry.

TABLE XIII.

Date.	No. of Observations.	Temp. on Callendar-Griffiths Air-Scale.	Temp. on Rowland's Air-Scale.	Difference Δ .
Comparison of Platinum Thermometer with Baudin 6163.				
May 18th.	5	5.261	5.249	+0.012
	6	9.937	9.922	+0.015
	6	15.340	15.325	+0.015
	6	20.066	20.052	+0.014
	4	24.881	24.861	+0.020
	5	34.916	34.921	-0.005
	4	40.051	40.052	-0.001
Comparison of Platinum Thermometer with Baudin 6163.				
May 24th.	6	9.997	9.998	-0.001
	6	20.055	20.056	-0.001
	6	25.228	25.222	+0.006
	6	34.899	34.909	-0.010
	6	40.283	40.305	-0.022
Comparison of Platinum Thermometer with Baudin 6166.				
Apr. 10th.	5	10.873	10.844	+0.029
	5	19.650	19.611	+0.039
	5	30.136	30.114	+0.022
Comparison of Platinum Thermometer with Baudin 6166.				
May 21st.	5	5.131	5.127	+0.004
	6	9.869	9.852	+0.017
	6	14.899	14.871	+0.028
	7	19.951	19.923	+0.028
	6	24.987	24.965	+0.022
	6	30.226	30.212	+0.015
Comparison of Platinum Thermometer with Baudin 6166.				
May 28th.	6	5.242	5.227	+0.015
	6	10.091	10.066	+0.025
	6	19.978	19.937	+0.041
	4	30.191	30.163	+0.028
Comparison of Platinum Thermometer with Baudin 6165.				
Apr. 14th.	5	10.129	10.104	+0.025
	5	20.274	20.247	+0.027
	5	29.735	29.709	+0.026

Fig. 5.

Curves III., whose ordinates must be added to Rowland's absolute scale to reduce this to the Paris nitrogen-scale, were then obtained by combining curves II. and curve A.

TABLE XIV.
Corrections to Rowland's Baudin Thermometers.

Temp. ° C.	Baudin 6163.		Baudin 6165.		Baudin 6166.	
	Cor. to Paris Nitrogen- Scale.	Cor. to Paris Hydrogen- Scale.	Cor. to Paris Nitrogen- Scale.	Cor. to Paris Hydrogen- Scale.	Cor. to Paris Nitrogen- Scale.	Cor. to Paris Hydrogen- Scale.
1°	+·002	+·002	+·006	+·005	+·004	·003
2	·004	·003	·010	·009	·007	·005
3	·006	·005	·015	·013	·010	·008
4	·009	·007	·019	·016	·013	·010
5	·011	·008	·022	·019	·016	·013
6	·013	·009	·025	·022	·019	·015
7	·014	·010	·027	·024	·022	·017
8	·015	·011	·030	·026	·025	·019
9	·015	·011	·032	·028	·027	·021
10	·016	·010	·034	·028	·030	·024
11	·015	·009	·035	·028	·032	·025
12	·015	·009	·036	·028	·034	·027
13	·015	·007	·036	·028	·037	·029
14	·015	·007	·036	·027	·038	·031
15	·015	·007	·036	·027	·040	·032
16	·015	·006	·035	·026	·041	·033
17	·015	·006	·035	·026	·042	·034
18	·014	·005	·035	·025	·043	·034
19	·014	·005	·035	·025	·043	·034
20	·014	·004	·034	·024	·043	·033
21	·015	·004	·033	·023	·042	·031
22	·016	·004	·033	·023	·041	·029
23	·017	·005	·033	·022	·039	·028
24	·017	·006	·033	·021	·037	·026
25	·017	·006	·032	·021	·035	·024
26	·016	·005	·032	·021	·034	·022
27	·015	·003	·032	·021	·032	021
28	·013	·002	·032	·021	·030	·019
29	·011	·000	·032	·021	·028	·017
30	·009	—·002	·032	·021	·027	·016
31	·007	—·004				
32	·005	—·005				
33	·003	—·007				
34	·002	—·009				
35	·000	—·011				
36	—·002	—·013				
37	—·003	—·015				
38	—·004	—·016				
39	—·005	—·016				
40	—·005	—·016				
41	—·006	—·017				

Curves IV. give the corrections that must be added to Rowland's absolute scale to reduce this to the Paris hydrogen-scale; these curves are obtained from curves III. by making use of the relation of the hydrogen and nitrogen scales of temperature as determined by the experiments of M. Chappuis (Guillaume, 'Thermométrie de Précision,' p. 258).

The corrections to reduce the readings of the Baudin thermometers (when referred to Rowland's absolute scale) to the Paris nitrogen and hydrogen scales of temperature, as obtained from curves III. and IV. fig. 5, are given in Table XIV.

RECALCULATION OF MECHANICAL EQUIVALENT.

Reduction to the Paris Nitrogen-Scale.—Rowland's values of the mechanical equivalent, as expressed in terms of the rise of temperature of water, are deduced from an equation of the following form :—

$$J_{T^{\circ}} = \frac{W}{(T + 5^{\circ}) - (T^{\circ} - 5^{\circ})},$$

where W is equal to the energy in ergs required to raise the temperature of the water from $T^{\circ} - 5^{\circ}$ to $T^{\circ} + 5^{\circ}$.

Hence, if C_{20} and C_{10} are the corrections that have to be applied to the temperature indicated by the Baudin thermometers to reduce these to the Paris nitrogen-scale, we shall have for the value of $J_{15^{\circ}}$ reduced to the nitrogen-scale :—

$$\begin{aligned} J_{15^{\circ}}^N &= \frac{W}{(20^{\circ} + C_{20}) - (10^{\circ} + C_{10})} = \frac{W}{10} \frac{1}{1 + \frac{C_{20} - C_{10}}{10}} \\ &= J_{15^{\circ}} \frac{1}{1 + \frac{C_{20} - C_{10}}{10}} = J_{15^{\circ}} \left(1 + \frac{C_{10} - C_{20}}{10} \right) \text{ approximately} \\ &= J_{15^{\circ}} K_{15^{\circ}}, \text{ where } K_{15^{\circ}} = \left(1 + \frac{C_{10} - C_{20}}{10} \right). \end{aligned}$$

The values of K (the factor to reduce Rowland's values of the equivalent to the Paris nitrogen-scale) were computed, from the corrections found by the preceding comparisons, for each degree centigrade for each of the Baudin thermometers.

The results of Rowland's experiments on the mechanical equivalent of heat are summarized on pp. 192–196, Proc. Am. Acad. xv. 1879.

The values of J at any one temperature were averaged for

Comparisons of Mercury and Platinum Thermometers. 41

each thermometer separately, and each mean was then reduced to the nitrogen-scale by the proper correction factor K . The general mean value of J at that temperature was then obtained by taking the mean of the values given by each thermometer, giving to each thermometer a weight equal to the number of experiments performed with that thermometer at the given temperature. These values of J , corrected as indicated above, were then plotted, and the values obtained from the resulting smooth curve were taken as the final value of J on the nitrogen-scale. For the sake of comparison, the original as well as the corrected values of J are given in the following table.

TABLE XV.

Temp.	Rowland's values of J (absolute scale) on the C.G.S. System.	Rowland's values of J reduced to Paris Nitrogen-Scale on the C.G.S. System.	Temp.	Rowland's values of J (absolute scale) on the C.G.S. System.	Rowland's values of J reduced to Paris Nitrogen-Scale on the C.G.S. System.
7°	4.207×10^7	4.200×10^7	22°	4.176×10^7	4.178×10^7
8°	4.204	4.198	23°	4.175	4.177
9°	4.202	4.196	24°	4.174	4.177
10°	4.200	4.195	25°	4.173	4.176
11°	4.198	4.193	26°	4.172	4.175
12°	4.196	4.192	27°	4.171	4.175
13°	4.194	4.190	28°	4.171	4.174
14°	4.192	4.189	29°	4.170	4.174
15°	4.189	4.187	30°	4.171	4.175
16°	4.187	4.186	31°	4.171	4.175
17°	4.185	4.184	32°	4.171	4.175
18°	4.183	4.183	33°	4.172	4.175
19°	4.181	4.182	34°	4.172	4.176
20°	4.179	4.181	35°	4.173	4.177
21°	4.177	4.179	36°	4.173	4.178

These results are also plotted in fig. 6 (p. 42). For the sake of comparison the electrical determinations of the mechanical equivalent by Griffiths and by Schuster and Gannon are also shown on the same figure.

CONCLUSIONS.

These comparisons would seem to show that Baudin 6163, when its indications are reduced to the absolute scale by means of Rowland's tables, reads lower than the Paris nitrogen-scale over the range 0° to 35°, these differences amounting to 0°·016 at 10°, 0°·014 at 20°, and 0°·009 at 30°.

Fig. 6.

The corrections to reduce to the Paris hydrogen-scale and the Callendar-Griffiths' air-scale are respectively $0^{\circ}\cdot010$ and $0^{\circ}\cdot010$ at 10° , $0^{\circ}\cdot004$ and $0^{\circ}\cdot012$ at 20° , and $-0^{\circ}\cdot002$ and $0^{\circ}\cdot009$ at 30° . Baudin 6166, when its indications are reduced to the absolute scale, reads lower than the Paris nitrogen-scale throughout the range 0° to 31° ; the corrections are $0^{\circ}\cdot030$ at 10° , $0^{\circ}\cdot043$ at 20° , and $0^{\circ}\cdot027$ at 30° . The corrections to reduce to the Paris hydrogen-scale and the Callendar-Griffiths' air-scale are respectively $0^{\circ}\cdot024$ and $0^{\circ}\cdot025$ at 10° , $0^{\circ}\cdot033$ and $0^{\circ}\cdot041$ at 20° , and $0^{\circ}\cdot016$ and $0^{\circ}\cdot027$ at 30° . The indications of Baudin 6165, when reduced to the absolute scale, are about $0^{\circ}\cdot035$ too low throughout the range 10° to 30° .

As will be seen from the curves for the "capacity for heat of water" shown in fig. 6, the changes in Rowland's values are small, amounting to a decrease of about 1 part in 850 at 10° C., the value at about 18° remaining unchanged, while those at 20° and 25° are respectively increased by about 1 in 2100 and 1 in 1400. The variation of the specific heat of water with temperature between 15° and 25° , when Rowland's values are referred to the Paris nitrogen-scale, is practically identical with that given by Griffiths' curve. This suggests at once that an explanation of the differences between the mechanical and electrical determinations of the mechanical equivalent must be sought in the energy-measurements.

In Rowland's experiments an error in the energy-measurements may be due either to an error in the diameter of the torsion-wheel of his calorimeter or to the system of weights employed. Inasmuch as the diameter of the torsion-wheel was measured many times by comparison with two standard metre-bars, each of which had been compared with the Coast Survey and other standards, the possibility of an error sufficient to account for the observed differences between the mechanical and electrical determinations must be sought elsewhere. Turning to the question of the weights used in these experiments, we see that it is not necessary that they be correct absolutely with the standards, if they are only relatively correct, as the formula for the mechanical equivalent contains a weight in both numerator and denominator. On the other hand, if they were not correct relatively, we should hardly expect to find the almost constant difference between the determinations of Griffiths and those of Rowland (reduced to nitrogen-scale) throughout the range 15° to 25° . The evidence accumulated thus far would, we believe, suggest, as a possible explanation of these differences, a still undiscovered

error in the system of electric units employed. This, indeed, in the light of the enormous and painstaking work which serves as the basis of these units, is not probable, but, considering the difficulties encountered in the standards of electromotive force and current, it is not altogether impossible. The enormous experimental evidence which served as the basis of the international ohm, together with the recent confirmation of the accuracy of this unit, renders it almost certain that the difference need not be looked for in this direction.

At the Toronto meeting of the British Association, before which an abstract of this paper was read, the Committee on Electric Standards received an appropriation for the redetermination of the electrochemical equivalent of silver and the absolute electromotive force of the Clark cell, and their results are awaited with much interest. In the meanwhile the results of a determination of these quantities, made by Dr. Kahle at the Reichsanstalt, have been published (*Zeitschr. f. Instk.* xviii. pp. 229, 267, 1898; *Wied. Ann.* lix. p. 532, 1896; *Wied. Ann.* lxvii. p. 1, 1899). His final value of the E.M.F. of the German H standard form of Clark cell was

$$E_{15^{\circ}} = 1.4325 \text{ volts.}$$

By making use of a previous comparison between the Cambridge standard and the German H standard (B. A. Report, 1892), the Cambridge standard, when reduced in accordance with Kahle's value, becomes

$$\text{Cambridge}_{15^{\circ}} = 1.4329 \text{ volts.}$$

The value of the Cambridge standard (upon which is based Griffiths' values of the capacity for heat of water) as determined by Glazebrook and Skinner (*Phil. Trans. A*, 1892) was

$$\text{Cambridge}_{15^{\circ}} = 1.4342 \text{ volts.}$$

As has already been shown by Dr. F. A. Wolff (Johns Hopkins Univ. Circular, June 1898), when the values of the E.M.F. of the Clark cell, as found by Kahle, are applied to Griffiths' values of the capacity for heat of water, as well as those of Schuster and Gannon, they are brought into very fair agreement with those of Rowland as corrected by the results of our comparisons (differing by about 1 part in 1400). This is shown by the following table:—

		Old.	Corrected.
At 15°	{ Rowland (air-scale) ...	4·189 × 10 ⁷	4·187 × 10 ⁷
	{ Griffiths (N. scale) ...	4·198 „	4·190 „
At 20°	{ Rowland	4·179 „	4·181 „
	{ Griffiths	4·192 „	4·184 „
	{ Schuster and Gannon		
	{ (N. scale) at 19°·1...	4·1905 „	4·185 „
At 25°	{ Rowland	4·173 „	4·176 „
	{ Griffiths	4·187 „	4·179 „

Dr. Guthe has kindly informed me of the result of a redetermination of the electrochemical equivalent of silver, which was carried out during the past year by Professor Patterson* and himself at the University of Michigan. Their final value of the electrochemical equivalent of silver is 0·0011193 gramme per ampere per second for a “used” solution. This is very near the value found by Kahle for a used solution. For a “fresh” solution Kahle found 0·0011182.

A recalculation of the E.M.F. of the Cambridge standard Clark cell as determined by Glazebrook and Skinner, if we use the value of the equivalent found by Guthe and Patterson, is

$$\text{Cambridge}_{15^{\circ}} = 1\cdot4327 \text{ volts,}$$

which is in very close agreement with the value as given by Kahle. If Griffiths’ values are corrected in accordance with this determination, the agreement with Rowland’s corrected values is very close, they being greater by about 1 part in 2300 throughout the range 15° to 25°.

In conclusion the authors wish to express their sincere appreciation of Professor Rowland’s kindness in allowing the use of his thermometers for these comparisons, and their many obligations to Professor Rowland and Professor Ames for their frequent advice and assistance throughout the course of this work; and also to Dr. W. S. Day for so freely placing his apparatus at our disposal, and his many acts of kindness.

Johns Hopkins University,
Baltimore.

* The results of this investigation have since been published in the *Phys. Rev.*, Dec. 1898, p. 257.