DISCUSSION.

Mr. N. T. M. Wilsmore pointed out that the difficulty found by Dr. Lowry in using toluene in connection with the original handy form of the Ostwald mercury valve could be got over by inserting a "safety-tube" arrangement in the stem. One could then fill the upper part of the regulator with water, so that the toluene did not reach the tap.* If, however, one filled the bulb with a liquid heavier than water, such as chloroform, the apparatus could retain the simpler form, water being used as above to transmit the motion to the mercury.

An exceedingly efficient form of thermostat regulator had been devised by Dolezalek.† In this the expanding liquid was contained in a long copper spiral; and the apparatus had consequently a smaller temperature lag than one having a "bulb" of like dimensions made of glass. He (the speaker) had obtained good results from such a spiral connected to an Ostwald mercury valve. No "creeping" took place if the joint between copper and glass was properly made. Dolezalek had also recommended a more vigorous stirring of the bath than that shown by Dr. Lowry.

Dr. T. M. Lowry (*in reply*) said that the difficulty arising from the leakage of toluene could be very easily overcome by using clean toluene and clean mercury. This was much simpler than the introduction of an additional layer of water, and was so efficient that the change in the temperature of regulation did not amount to more than a few thousandths of a degree per month.

The Dolezalek regulator was undoubtedly more sensitive than the glass spirals which had been exhibited at the meeting. It was, however, no easy matter to make a good joint between the copper and the glass, and in practice a glass spiral costing only three or four shillings was sufficiently sensitive for all experimental work in which a variation of $\pm 0.01^\circ \text{C.}$ was allowable. The amount of stirring required depended on the temperature of the bath, on its size, and on the efficiency of the heating insulation of the walls of the bath. In the apparatus shown it was not possible to detect any temperature gradient even with a thermometer graduated to 0.01°C. when the bath was working at 20° .

* See also Dony-Hénault, *Z. Elektroch.* **11**, 3 (1905).

† *Z. Physik. Ch.* **26**, 321 (1898).