Dolff,²¹ Bohland,²² and Sabrezas and Frezal²³ indicate that ingestion of tannic acid and tannin causes a decrease in uric acid excretion; those of Weiss²⁴ and Ulrici²⁵ fail to show any such results. This phase of the problem could be settled by an experiment similar to the second one of this series, a detannated coffee being used in place of a decaffeinated product.

Finally, it would doubtless prove interesting to perform similar series of experiments with dogs and rabbits in order to study the effect of caffeine ingestion on the excretion of uric acid and of allantoin in species in which uric acid is not the prominent normal end-product of purin metabolism.²⁶

SUMMARY

Slight modifications of Benedict's method for determining uric acid make possible more accurate determinations than have hitherto been obtained.

The addition of a strong coffee infusion to a purin-free diet causes a marked increase in the excretion of uric acid.

The addition of Kaffee Hag—a decaffeinated coffee product—to a purin-free diet does not cause any increase in the excretion of uric acid. If, however, caffeine is added to the Kaffee Hag, the excretion of uric acid is decidedly increased, as in the case of coffee.

The effect of adding tea to a purin-free diet is similar to that obtained by adding coffee to the same diet.

The increase in excretion of uric acid after adding coffee, tea or caffeine to a purin-free diet seems to be proportional to the quantity of caffeine ingested.

The increase in the amount of uric acid excreted under these conditions is equal to the quantity of uric acid which would be obtained by the demethylation and subsequent oxidation of from 10 to 15 per cent. of the ingested caffeine.

AN AUTOMATIC METHOD OF PROLONGED SERIAL BLOOD-PRESSURE REGISTRATION IN MAN*

BERNARD FANTUS, M.D.

Professor of Pharmacology and Therapeutics, University of Illinois, College of Medicine

CHICAGO

Work on this apparatus was commenced in 1909, when the desirability of a method for the continuous observation of human blood pressure was strongly borne in on my mind, as I investigated at the Charité in Berlin in association with Professor R. Stachelin the effect of muscle work on human blood pressure by means of the Uskoff sphygmotonometer. It became probable that this effect was, in general, a curve with a primary rise above and a secondary fall below the previous level; but we could not prove this: for, in spite of taking an observation every five minutes, one could not

feel certain that one had caught the maximum and the minimum of the effect.

The work of elaborating such an apparatus has since then been continued abortively until, in the summers of 1915 and 1916, Dr. Otis M. Cope, of the department of physiology of the University of Michigan, joined me in a final effort, which we believe has succeeded. The result of this joint work, carried on at the laboratory of physiology of the University of Michigan, was the elaboration of the principles that underlie the construction of an apparatus for the purpose in mind, and

the development of a valve, chiefly the design of Dr. Cope, which, however, was not entirely satisfactory. Under the impression that a somewhat different method of attack on the problem of construction of the valve might possibly lead to a more satisfactory result, I have continued work on this apparatus at my laboratory with the cooperation of the mechanic, Edward L. Feeney, who deserves thanks for patient help and useful suggestions in the elaboration of the rotary valve.

THE APPARATUS

The essential portion of the apparatus is a rotary valve, which takes care of the rhythmic inflation and deflation of the sphygmomanometer cuff, while it simultaneously opens the tambour space of an Erlanger arrangement during the inflation and deflation, and closes it during the time of recording.

* Presented before the Chicago Society of Internal Medicine, March 26, 1917.

* From the Pharmacologic Laboratory of the University of Illinois, College of Medicine.

Fig. 1.—Rotary valve, assembled in center, the working faces of the valve on each side. The numbers 1, 2 and 3 have been placed on corresponding parts of the valve in all portions of the illustrations. The longitudinal sections a and b, in the left hand upper corner, show the manner in which air-tight closure of the valve faces is obtained.

21. Dolff, F.: Inaug. Diss., Bonn, 1898.

22. Bohland, K.: München. med. Wchnschr., 1899, 46, 505.

23. Sabrezas and Frezal: Jour. de physiol. et path. gén., 1899, 1, 221.

24. Weiss, J.: Ztschr. f. physiol. Chem., 1899, 27, 216.

25. Ulrici, I.: Arch. exper. f. Path. u. Pharmacol., 1901, 46, 321.

26. In addition to the references already given, the following will be found of interest:

Brugach, T., and Schittenhelm, A.: Der Nucleinstoffwechsel und Seine Störungen, 1910.

Jones, W.: Nucleic Acids, 1914.

Krüger, M., and Schmidt, J.: Arch. f. exper. Path. u. Pharmacol., 1901, 45, 259; Ztschr. f. physiol. Chem., 1905, 45, 1.

Levison, C.: Ztschr. f. klin. Med., 1894, 26, 293.

McCrudden, F.: Uric Acid, 1905.

The Work Treatment.—Those of us who have any imagination cannot fail to realize the difference in atmosphere and morale in hospitals where the patients have nothing to do but smoke, play cards, or be entertained, from that found in those where for part of the day they have regular, useful and productive work.—Col. Robert Jones, C.B.

There are, then, in the apparatus, two air spaces that do not open into each other. I shall call them

Fig. 2.—Rotary valve connected with the recording apparatus. The valves *a* and *b* introduce a regulable fine leak in the tambour space and cuff space, respectively. The cam *c* operates the starting and stopping mechanism of the kymograph.

“cuff space” and “tambour space,” respectively (Fig. 1). The cuff space communicates freely with a mercury manometer and an elastic rubber bulb. The manometer records, on smoked paper, the height of the pressure existing in the cuff space at a given moment. The rubber bulb transmits to the tambour space the pulse oscillations occurring during the gradual deflation of the cuff. To inflate the cuff, a constant source of compressed air at a pressure of over 300 mm. of mercury is required. In the air intake pipe, there should be provided a valve that readily permits slight changes to be made in the air admitted to the cuff space. This valve, once turned on to the

proper degree, remains open throughout the period of use of the apparatus. By means of the rotary valve, which revolves at the rate of about once a minute, a momentary connection is made between the tube carrying the compressed air and the cuff space. Thus inflation is accomplished. Then, for two thirds of the circumference of the valve, the cuff space is closed, excepting for a small regulable leak (Fig. 2 *a*), which permits gradual deflation to take place. Since, during the period of gradual deflation, the tambour (to be presently described) records the tracing to be interpreted, I shall speak of this as the “recording time.” As soon as the recording has been accomplished, it is desirable to have the air escape rapidly from the cuff space, so as to permit a period of rest for unimpeded return circulation in the arm of the person under examination. We have found that one third of a minute is sufficient for this purpose. Hence, after the rotary valve has traveled for two thirds of its circumference, an opening is encountered by means of which the cuff space communicates with the surrounding air, permitting the pressure to fall instantly.

The cuff space is in relation with the tambour space by means of a glass bulb into which the rubber bulb of the cuff space projects, all the connections being air-tight. Thus the oscillations of the rubber bulb, imparted to it by the pulse, are magnified and recorded by means of the tambour. To eliminate, as much as possible, the recording of the large pressure oscillations that occur during inflation and rapid deflation of the cuff space and might damage the tambour, it is necessary, as has been shown by Erlanger, that the tambour space be freely open during these stages. This is accomplished by holes placed in that side of the rotary valve which controls the tambour space. These holes are spaced in such a manner as to act synchronously with those controlling the cuff space. The holes, in relation with the tambour space, must

Fig. 3.—Photograph of apparatus, showing valve and driving as well as recording mechanisms used. A much simpler driving mechanism would, of course, be entirely adequate. A pneumograph and its recording tambour is shown in addition to the parts given in the diagrammatic Figure 2.

be a little larger than those in relation with the cuff space, whereby the large and undesirable oscillations can be more effectually eliminated, than if they are of

Fig. 4.—Determination of systolic and of diastolic pressure: The Erlanger criterion for systolic pressure, namely, the first divergence of the two limbs of the tambour record (S in Curve B) is well shown. The diastolic criterion chosen is the last large oscillation of the tambour lever (D in Curve B). This is more difficult to decide on, owing to the disturbing influence of respiration, as is well shown in this tracing by the unusually deep inspiration at the end of the tracing, which obviously produced the increase in the size of the tambour oscillation following it and which would have given rise to difficulty of interpretation of the record, had the respiration record not been taken simultaneously. In the tracings, A is the manometer curve, B, the tambour record, and C the respiration record; S indicates the level of systolic pressure, and D the level of diastolic pressure.

exactly the same size. As Erlanger has shown, it is also essential that a fine leak exist in the tambour space. Cope has found that, by using an adjustable leak (Fig. 2 b), one may obtain an optimum record of the oscillations of the bulb.

In order to economize drum space, it is desirable to have a mechanism that automatically starts and stops the drum. This may be done by means of a cam (Fig. 2 c), attached to the shaft of the rotary valve, which operates the starting mechanism of the drum.

A long paper kymograph is required, as well as signal magnets for the time and other recording. A second tambour for a pneumograph tracing (C in Figs. 4 and 5) is also desirable, as respiration affects the character of the pulse oscillations, as is shown by Figure 4.

We have commenced work on the problem of an apparatus that will furnish its own compressed air, as well as, if possible, do away with the necessity of applying fresh loops of paper, when one takes a very prolonged record.

INTERPRETATION OF THE RECORD

The records (Figs. 4 and 5) furnished by this apparatus consist of a succession of so-called "continuous" Erlanger¹ tracings (B), placed above a manometer curve (A) that was obtained synchronously with the pulse tracing. The latter is used to determine the points of systolic and of diastolic pressure (Fig. 4). For determining systolic pressure, we have found the Erlanger² criterion, namely, the separation of the lower end of the limbs of the tambour record, much more useful than the von Recklinghausen³ criterion—that is, the sudden increase in size of the oscillations—for, at times, the increase in size is not at all sudden, but quite gradual.

The decision regarding the point at which the diastolic pressure is to be placed seems somewhat more difficult. Erlanger's¹ criterion for diastolic pressure—the maximum amplitude of tambour oscillations—has been shown by MacWilliam and Melvin⁴ not to be correct. They suggest using the last of the large oscillations for this purpose. Cope and I have found that the MacWilliam and Melvin criterion coincides quite well with the fourth auscultatory phase obtained by means of the Korotkoff method, as shown by records obtained with our apparatus, when a signal magnet (whose tracing is placed in the accompanying figures above the time record) is used to record the Korotkoff sounds, as has been done in some of the curves shown. Unfortunately, at times, the oscillations decrease gradually rather than abruptly, so as to leave one in doubt as to just where one should decide the change has taken place. This difficulty can be overcome, to a certain extent, by permitting the escape of air during the recording period to be sufficiently rapid for good definition of differences. Nevertheless, the influence of the respiratory pressure changes occasionally also adds to the difficulty of arriving at a judgment regarding the point to be chosen.

Making a simultaneous record of the Korotkoff sounds by means of the signal magnet is quite helpful

Fig. 5.—Portions of a record from a normal person, showing the practically uniform maintenance of pulse pressure that seems characteristic of normal tracings.

in the interpretation of the record. Nevertheless, I have found that frequently, when doubt exists as to

1. Erlanger, Joseph: *Johns Hopkins Hosp. Rep.*, 1904, **12**, 53.
2. Erlanger, Joseph: *Am. Jour. Physiol.*, 1908, **21**, 24.
3. Von Recklinghausen: *Arch. f. exper. Path.*, 1906, **55**, 375.
4. MacWilliam and Melvin: *Heart*, 1914, **5**, 153.

where one should place the diastolic pressure, because of interference of the respiratory oscillations, the same doubt existed in deciding on the end of the fourth phase of Korotkoff. For, at times, one or two of the muffled sounds of the fourth phase are followed by one or two sounds that are again quite loud. The tracing records this faithfully.

By projecting the point of systolic and of diastolic change in the pulse tracing (*B*) down on the manometer curve (*A*), and by connecting the corresponding points on succeeding manometer curves, a good picture of the blood-pressure changes from minute to minute can be obtained. These changes are comparatively slight under normal conditions, though waves, probably corresponding to the Traube-Herring waves, are easily discernible. By means of this apparatus, I have obtained satisfactory records of the influence on blood pressure of emotion, of attention as well as of muscular movement, and of the effects of a general electric light bath, of amyl nitrite inhalation, and of nitroglycerin administration.

CONCLUSION

An apparatus having been devised whereby it is possible to obtain graphic records of human blood pressure for considerable periods of time, it will probably be found useful not only in psychologic, physiologic and pharmacologic research, but also in clinical medicine, in diagnosis as well as in treatment. For instance, it might be of diagnostic as well as of therapeutic importance to differentiate between those cases of high blood pressure that respond, in a typical manner, to nitrite and those that do not. In critical operative cases, the continuous recording of blood pressure during the anesthesia might be of help to the surgeon.

Blood Changes in Trinitrotoluene Poisoning.—Capt. Matthew J. Stewart of the Royal Army Medical Corps has recently summed up his observations in fourteen clinical cases of trinitrotoluene poisoning with reference to the blood as follows: In the early stages blood changes are extremely slight, but pronounced deviations occur later. A polymorphonuclear leukopenia, often of extreme degree, with no corresponding failure of the lymphatic function, was observed in nine out of the fourteen cases. An anemia of "pernicious" type, but lacking certain notable characteristics of idiopathic anemia, especially poikilocytosis and punctate basophilia, occurred in three cases. The tentative conclusion arrived at is that the occurrence of either of these changes in any marked degree in a case of trinitrotoluene poisoning is to be regarded as a serious prognostic indication, but this is a point on which further observation should throw much light. An absolute but variable lymphocytosis was found in nine cases. It did not appear to bear any relationship to either of the foregoing conditions. A pronounced and progressive neutrophil leukocytosis occurred in one case which ended fatally. Qualitative changes in the red cells were present in a marked degree in one case only, and punctate basophilia was never observed. —*Proceedings of the Royal Society of Medicine*, March, 1917.

A NEW MASTIC TEST FOR THE SPINAL FLUID*

JAMES A. CUTTING, A.B., M.D.

Assistant Physician, Agnews State Hospital

AGNEW, CALIF.

Recently much work has been done in order to find a serologic test which will be positive in all cases of syphilis of the nervous system. Collins¹ estimates that 15 per cent. of paretics and 70 per cent. of cerebrospinal syphilitics fail to give a positive Wassermann reaction in the spinal fluid. While these figures are perhaps higher than the average, the fact still remains that a definite percentage do fail to give a positive Wassermann. Collins¹ also estimates that 97 per cent. of these cases give a positive colloidal gold reaction. Any one, however, who has prepared the gold solution and has carried out the test cannot but be impressed with the many difficulties it entails.

While working with the colloidal gold solution, Emanuel² endeavored to find a substitute which could be more easily prepared and in which the reaction could be carried out with less possibility of error. For this purpose it seemed that an emulsion of mastic was best adapted. Neisser and Friedman, among others, had already demonstrated the similarity in action on albumin of this substance and colloidal gold.

In brief, Emanuel's procedure was as follows: Five test tubes were used containing, respectively, 0.25, 0.125, 0.062 and 0.031 c.c. of spinal fluid mixed with 1 c.c. of a 1.25 per cent. salt solution. The fifth tube contained no spinal fluid, and was used as a control. Lastly, 1 c.c. of an emulsion of mastic was added to each tube. In this manner he tested out the spinal fluids of thirty

Fig. 1.—Mastic reactions: *A*, negative; *B*, positive; in both *A* and *B*, potassium carbonate has been used in making up the salt solution; *C* and *D*, the same cases in the absence of the potassium carbonate. (Note the confusion in *C*, in which Tubes 4 and 5 show a precipitate.)

psychiatric patients. In paresis, tabes and cerebrospinal syphilis there occurred a precipitation of the mastic in all the tubes, whereas in the normal fluid no flaking occurred except in the control. The mastic solution had to be made up fresh each time and quickly used.

In order to try out the reaction, in the last nine months I have made mastic tests on the spinal fluids of 200 different patients at the Agnews State Hospital. In each case I have made a cell count, in 100 cases the butyric acid test, and in thirty cases the colloidal gold reaction. In every case a Wassermann examination has been made.

The mastic reaction I found to be a very delicate one. If the salt solution and the emulsion of mastic were made with distilled water and promptly used, a rapid precipitation of the mastic occurred in the control tube which contained no spinal fluid. If, how-

* Read before the Agnews Psychiatric Society, Feb. 9, 1917.

1. Collins, Joseph: Pathognomonic Alterations of the Cerebrospinal Fluid in Syphilis of the Nervous System, *Am. Jour. Med. Sc.*, 1916, 151, 222.

2. Emanuel, G.: Eine neue Reaktion zur Untersuchung des Liquor cerebrospinalis, *Berl. klin. Wchnschr.*, 1915, 52, 781-804; abstr., *The Journal A. M. A.*, Sept. 4, 1915, p. 908.