Clinical Notes, Suggestions, and New Instruments

A NEW METHOD FOR THE STUDY OF BACTERIA*

W. D. FROST, PH.D., DR.P.H., MADISON, WIS.

Professor of Agricultural Bacteriology, University of Wisconsin.

The rapid identification of bacteria is a matter of the greatest importance. Certain organisms have characteristic morphology or staining reactions, but in most cases it is necessary to make a culture of the organism before it can be identified, a process that requires hours or even days.

Modifications of a method that I have already described¹ furnish a means of studying the morphology of the cells and the colonial structure of young cultures of bacteria, many of which show very interesting and, I believe, distinctive characters in a surprisingly short time, that is, in from three to five hours.

The medium in its simplest form is a 50-50 mixture of nutrient agar and sterile milk.

The apparatus is of such a simple nature that it will be found in all bacteriologic laboratories. It includes pipets, microscopic glass slides, platinum needles, an incubator, common stains, etc. The pieces of apparatus that could be called special are a "moist chamber cabinet," a warm plate for drying the films, and forceps for handling microscopic slides; but makeshifts for these are easily found in the common equipment of all laboratories.

Fig. 1.—Method of making "little plates": A, milk and nutrient agar, sterilized separately and mixed in equal quantities; B, medium, melted and ready for use; C, flaming the slide; D, inoculating the medium and spreading the drops; E, slide ready for incubation in the moist chamber, F.

METHOD OF MAKING CULTURES

A clean glass slide is sterilized in the direct flame and placed on the table. When it has cooled down there is placed

* A part of the expense connected with this and further work of the same kind, now in progress, was met by an unknown business friend of Madison, who made his contribution through Prof. C. S. Slichter, chairman of a university committee on research.

1. Frost, W. D.: A Rapid Method of Counting Living Bacteria in Milk and Other Richly Seeded Materials, J. A. M. A. 66: 889-890 (March 18) 1916.

on it, from a pipet, two separate, medium-sized drops of the culture medium, which has been previously melted and cooled to about 50 C. Into one of the drops of the culture medium, the material containing the bacteria to be cultivated is introduced by means of a platinum loop and thoroughly mixed, while at the same time the drop is spread over the slide in a thin film. When the first drop is spread, the loop is carried over to the second drop, which is thus inoculated with a

Fig. 2.—Colonies, four hours old, of organisms grown on the "little plates"; G, *Micrococcus pyogenes* var. *aureus*; H, *Streptococcus pyogenes* (?), from a throat culture; I, *Bacillus coli*; J, *Bacillus subtilis*. The drawings were made by Miss Margaret Brown under the camera lucida, and give the relative size as well as the shape and arrangement of the individuals in the colony.

small fraction of the same material. This drop is spread in the same way as the first. As soon as this is done, the slide is put in a moist chamber and placed in the incubator. In this way two little plates are made on the same slide from each sample to be examined, containing quite different amounts of the inoculum.

It is necessary to work rapidly in making these "little plates" in order to prevent the agar from hardening before it is spread. It is also necessary to use aseptic technic at every step, except that I have not found it necessary to take any special precautions against air infection during the preparation of the plates, or to keep the edges and underside of the glass slides sterile. The method is illustrated in Figure 1.

INCUBATION

Immediately after being made, the little plates are put in a moist chamber where they are kept until the colonies grow to sufficient size to be examined. A few slides, as a makeshift, may be put in Petri dishes, the bottoms of which are covered with wet filter paper, on glass rods which keep them from getting wet. It is much more convenient, however, to have a "moist chamber cabinet" and when a large number of slides is used, it is quite necessary. Such a cabinet, especially designed for this work, is represented in Figure 1 F.

Whatever the form of the moist chamber, it is put in an incubator and kept at the temperature best suited to the needs of the organisms under study. The colonies usually attain sufficient size in the course of three or four hours to give the characteristic arrangement of the cells. If the colonies are allowed to grow much larger, the individuals become so massed that only those at the periphery are distinctly discernible, so that there is usually no advantage in large colonies.

When the "little plates" are removed from the incubator they are taken from the moist chamber and dried. The drying is an important step and is apparently best done rapidly

by placing the slides on a warm plate kept at a temperature just below 100 C.

STAINING

When the slides are dry, they are ready for staining. They are put first in a 10 per cent. solution of acetic acid in 95 per cent. alcohol for one minute or more. They are then drained and put in Loeffler's methylene blue diluted 1:4 with distilled water. The slides are allowed to stain from two to three minutes and are then washed free from the stain in a jar of tap water. They are dried and are then ready for microscopic examination.

With practice it is soon possible to make these preparations in such a way that the colonies are deeply stained with the individual cells well differentiated and the background clear. The best results are usually obtained after the stain has been used for some time. For this reason the stain is used, in jars, over and over, rather than poured over the preparations fresh each time. So far, a diluted Loeffler's stain has been used exclusively; but it seems quite possible that other formulas can be developed and that perhaps other stains than methylene blue can be used to advantage. For example, it may be possible to use a red stain, such as fuchsin or safranin, which would be very desirable, especially in photomicrographic work. It is also quite possible, indeed, very probable, that stains can be used to bring out the metachromatic granules in the diphtheria and other germs. The double spore and the Gram stains have been successfully used.

AN IMPROVED SPINAL PUNCTURE NEEDLE AND PRESSURE APPARATUS

ABRAHAM LEVINSON, B.S., M.D., CHICAGO

It is now generally recognized that spinal puncture is one of the most useful diagnostic procedures, and that the pressure of the cerebrospinal fluid is a useful guide in the diagnosis of many diseases of the nervous system. Many needles have been used for doing a spinal puncture since Quincke first described the procedure. In using most of the needles, however, one is unable to tell whether or not he has reached the subdural canal until he has withdrawn the stylet entirely. If the canal has been reached, the fluid comes out with a greater or lesser force on removal of the stylet, and by the time the manometer is introduced into the needle, as for instance in the Quincke apparatus (Fig. 1), too much fluid has been lost to give a correct pressure reading; for it has been found that the pressure is greatly lessened, even if only a few drops of cerebrospinal fluid have been removed.

In order to prevent the loss of fluid in taking the cerebrospinal fluid pressure, and in order to make it easier to determine whether or not one has reached the subdural canal, I have devised the apparatus and method herein described.

The apparatus (Fig. 2) consists of a specially devised needle and a glass manometer.

The needle is the ordinary medium sized spinal puncture needle with a short piece of glass mounting and a stopcock. The glass tubing (A)

Fig. 1.—Quincke apparatus for measuring the cerebrospinal fluid pressure.

enables one to see whether the spinal canal has been reached with the first puncture. One need only withdraw the stylet beyond the glass tubing, and if the spinal canal has been reached, the fluid withdrawn will be seen immediately through the glass. If no fluid makes its appearance, the stylet can be inserted again and another attempt made. The fear of break-

ing the glass on boiling the needle for the purpose of sterilization is no objection. I have boiled the needles repeatedly, without ever breaking the glass.

The needle has a three-way stopcock (B), the distal end of which is about three-fourths inch long. This length is required in order to allow space from the stopcock to the end of the needle so as not to lose any of the fluid, as is the case when the valve of the stopcock is near the end of the needle. The stylet of the needle is long enough to reach from the distal to the proximal end of the needle.

The manometer (D) consists of a long glass tube of 1 mm. bore which is inserted into the projecting end of the needle (C). The glass tubing is 700 mm. long, as this was the highest pressure attained in any case studied. For convenience, it is divided into two portions, with a metal cap at the top of the first portion so that the other one can be screwed on. This makes the manometer convenient to carry in any small instrument case. Thick glass tubing is used for it so that it does not break easily in handling. A metal tape is attached to the glass to measure the pressure.

The needle (with the piece of glass tubing at the end, the stopcock and its prolongation) with the stylet inside is introduced into the spinal canal. After a snap has been felt, so that one is reasonably certain the subdural canal is reached, the stylet should be removed to a point beyond the glass attachment. If fluid is seen in the glass tubing, the stopcock (B) is turned at right angles so that the handle points upward. This shuts off the needle and prevents the fluid from coming out. The long glass tube (D) is then attached and the stopcock turned so that the handle points opposite the manometer. The fluid will then rush into the tube. When the fluid has mounted to the highest point, the pressure is read off in millimeters of water. If no fluid is seen coming out when the stylet is removed beyond the glass tubing, the stylet should be reintroduced until

the operator feels he is in the canal. It should then be removed beyond the glass, and the measurement taken. The handle is now turned parallel to the needle and

the fluid collected in test tubes. The whole needle, including the glass attachment, stopcock and prolongation, is not any heavier than an ordinary medium-sized spinal puncture needle.

The apparatus is made by Becton, Dickinson & Co., Rutherford, N. J.

30 North Michigan Avenue.

A SIMPLE ALL GLASS ATOMIZER

H. S. NEWCOMER, M.D. (PHILADELPHIA) AND W. S. McELLROY, M.D. (PITTSBURGH), PHILADELPHIA

Lieutenant and Lieutenant, Junior Grade, respectively, U. S. N. R. F.

The accompanying illustration gives the construction and approximate dimensions of an all glass atomizer which may be readily constructed from ordinary glass tubing and at a cost of only a few cents. It consists of an outer piece which is an 8 mm. glass T tube. One end of the T is drawn out to a nozzle. Another nozzle, slightly smaller, is drawn from a piece of straight 6 mm. glass tubing. The latter nozzle is inserted within the first in such a way that the tip of the inner nozzle is 2 mm. behind the tip of the outer nozzle and concentric with it. The opening of the outer must be a little larger than that of the inner one so that the air stream from the inner nozzle will have an unobstructed path. The inner nozzle should be nearly in contact with the walls of the outer nozzle. Both the nozzles must be in longitudinal section slightly concave outward, and neither should be at all constricted at the tip by flaming.