

Mass Spectrum SmartFormula Report

Analysis Info


Analysis Name E:\Data2\Rodrigo_AM\ROD-3-097000002.d
Method Tune_low_pos.m
Sample Name ROD-3-097
Comment Acetonitrile

Acquisition Date 2018-05-14 12:43:37

Operator Carin Larsson
Instrument / Ser# micrOTOF 125

Acquisition Parameter

| | | | | | |
|-------------|------------|----------------------|----------|------------------|-----------|
| Source Type | ESI | Ion Polarity | Positive | Set Nebulizer | 0.3 Bar |
| Focus | Not active | | | Set Dry Heater | 180 °C |
| Scan Begin | 50 m/z | Set Capillary | 4500 V | Set Dry Gas | 4.0 l/min |
| Scan End | 1000 m/z | Set End Plate Offset | -500 V | Set Divert Valve | Source |



| Meas. m/z | # | Formula | Score | m/z | err [mDa] | err [ppm] | mSigma | rdb | e ⁻ | Conf | N-Rule |
|-----------|---|---|--------|----------|-----------|-----------|--------|------|----------------|------|--------|
| 521.1918 | 1 | C ₂₉ H ₃₈ NaO ₂ Se | 100.00 | 521.1931 | 1.1 | 2.1 | 73.4 | 10.5 | even | | ok |