

Display Report

Analysis Info

Analysis Name D:\Data\Gosia\2024\J. Piskorz\24.02.12\JP-AM-20B_24.02.12_3.d
Method Tune_neg_Mid.m
Sample Name JP-AM-20B Instrument impact HD 1819696.00156
Comment

Acquisition Parameter

Source Type	ESI	Ion Polarity	Negative	Set Nebulizer	0.3 Bar
Focus	Active	Set Capillary	3900 V	Set Dry Heater	200 °C
Scan Begin	500 m/z	Set End Plate Offset	-500 V	Set Dry Gas	4.0 l/min
Scan End	1500 m/z	Set Charging Voltage	2000 V	Set Divert Valve	Source
		Set Corona	0 nA	Set APCI Heater	0 °C

