

CLG Microarray Test Results

Cell Line ID: P (WIBR3 E38, Wild type) p20 **Lab #:** CLG-45176 **Date received:** 12/09/21 **Date Reported:** 12/30/21

Contact Person: Hanqin Li **PI:** Dirk Hockemeyer **Institute:** University of California Berkeley

Test Code: aCGH 100 **Email:** hanqinli@berkeley.edu **PO #:** NA

Mailing Address: 188 Li Ka Shing Center, 1951 Oxford Street, Berkeley, CA 94720

Sample Type: Human PSC Frozen Cells	dsDNA Concentration: 406 ng/μl	Total dsDNA: 5.0 μg
Sex: Female	260/280 (1.7-1.9): 1.86	260/230 (≥1.0): 2.37
Array Type: Agilent 60K Standard aCGH		
Array ID Number: 252192453871_1_3	Human Assembly: hg38	Reference DNA: Agilent Euro Female

Quality Control

A sufficient amount of high quality genomic DNA, as determined by UV spec. (Nanodrop), fluorometer (Qubit) and Agarose Gel analysis, was extracted from cell line P (WIBR3 E38, Wild type) p20 and passed our internal quality standards for aCGH labeling.

aCGH Probes (PASS/FAIL): Pass

Experimental Deviations: None

Results:

Microarray analysis did not detect any amplifications or deletions.