



NANOCAVITY AND MICROCAVITY BASED PHOTONIC CRYSTAL BIOSENSOR FOR DENGUE DETECTION

Purvi Soni, Vijaylaxmi Kalyani

Purvi Soni, Electronics and Communication Department, Government Women's Engineering College, Ajmer, India

purvi_soni@outlook.com

Vijaylaxmi Kalyani, Electronics and Communication Department, Government Women's Engineering College, Ajmer, India

vijaylaxmikalyani@yahoo.com

Abstract— Photonic crystal biosensors is a new, advance and accurate measurement technology for bio sensing application. In this paper, a two dimensional (2D) photonic crystal nanocavity and microcavity sensor is designed for the detection of dengue using human blood plasma and platelets. The optical properties of plasma and platelets are observed and refractive index of normal plasma, infected plasma, normal platelets and infected platelets are used. The change in output transmission spectrum and Q factor has been observed with the shift in wavelength, thus by observing these changes dengue is detected. The designed sensor can predict results accurately and in short time, hence can be used for medical application. By designing micro and Nano cavity in same structure of biosensor, comparison is also done and results of Q factor and transmission are also observed. The band gap range of structure is 1360 nm to 2190 nm and 1550 nm continuous modulated wave is used in this design. The finite difference time domain (FDTD) method is used to analyze the result and Plane wave expansion (PWE) method is being used for band gap calculation.

Keywords— Nanocavity, Microcavity, Refractive index, Photonic Crystal, Finite Difference Time Domain (FDTD).

I. INTRODUCTION

With the rapid advancement in science and technology in today's world, especially in field of medicine, in different fields of medicine such as micro/nanotechnology, wireless technology, wireless communication, information technology, and biomedical sciences during past 10 to 15 years. Various wide range of models and designs are built for a wide range of biosensors [1]. Recently, biosensors are used in health care monitoring to detect diseases such as diabetics, dengue, early detection of cancer and various other diseases biomarkers. This provides the way of improving quality in health care system [2]. Biosensors provides monitoring of patients, athletes, premature infants, children, psychiatric patients, people who need long term care, they are also effective in prevention, timely diagnosis, control and treatment of diseases[3]. A biosensor mainly comprises three parts these are Bioreactor System, Transducer and an Output System [4]. These are designed to react only with a particular substance and result of this comes in form of a message that can be used by any electronic device. These are considered as receptor, communication systems based sensor

can display, simulate, treat, or substitute human biophysics performance. Biosensors are recently increasing attention in the field of health care monitoring, in order to solve the health related issues rapidly, economically, accurately and precisely for users, and also provides real time performance of athletes, or disease status of patient[5]. The important part in biosensors are the transducers that selectively and sensitively, translate biosignals into detectable signals in detectable form. Biosensors fulfill the requirement of healthcare system in terms of miniaturization, flexibility, conformability and starchability [6]. There are several wireless biosensors proposed in past 10 to 15 years [7]-[13].

1.1 Photonic crystal

Photonic crystals composed of the periodic arrangement of materials having different dielectric constants, which allows them to control and modulate the motion of photons [14]. These materials reflect specific wavelengths of light because of the photonic band gap, which allows them to display different structural colors. PC structures are common in nature, such as in opals and butterfly wings, presenting as periodic nanostructures and bright structural colors [15]. The structural color of PCs is obtained from their geometric structure, which is capable of controlling and manipulating the diffraction of light by the periodically structures. After various research [16], a large number of PC-related projects was proposed. Over the past several years, PCs have been broadly used in many applications, including sensing [17], biodetection [18]. One of the basic properties of PC is the photonic band gap (PBG), and the propagation of light within the frequency range of PBG [19]. Nevertheless, the periodicity of this structure will be broken when some defects are introduced in PC, which makes it possible for PC to give strong electromagnetic field confinement, small mode volume, and low extinction loss [20]. On the other hand, by adjusting the structural parameters of PC or infiltrating suitable materials in the air holes of PC, the propagation of light can be altered. Therefore, many PC based devices have been used in the applications of light flow control, such as filters [21,22], electro-optical modulators [23,24], switches [25,26], and delay devices[27]. Specially, PC based sensors are much more popular due to their promising characteristics like ultra-compact size, high measurement sensitivity, flexibility in structural



design, and more suitable for monolithic integration [28-30].

1.2 Dengue

Dengue is the most rapidly spreading, acute febrile mosquito-borne viral disease caused by dengue virus. Over the past 50 years, the incidence of dengue has increased by approx. thirty times, spreading across increasing number of countries as well as from the rural to urban regions. This disease has become endemic in most tropical and subtropical regions. An estimated number of about 50 million cases of dengue fever occur each year with 500,000 cases of dengue hemorrhagic fever and 22,000 deaths. Presently, about 2.5 billion people live in over 100 dengue endemic countries [31]. The presence of dengue has inflicted significant health, social and economic burden on these endemic areas [32]. There are several factors that contribute to the occurrence of the disease, mainly the increases in international travel and human population, besides global climate change [33]. Dengue virus belongs to the flavivirus genus of the flaviviridae family. The dengue virus is transmitted to vertebrates by infected *Aedes aegypti* and *Aedes albopictus* mosquitoes. The dengue virus is divided into four antigenically related but distinct Current and nano-diagnostic tools for dengue infection 807 serotypes; the Dengue 1, Dengue 2, Dengue 3 and Dengue 4. The mature dengue virus is spherical in shape and ~50 nm in size. Dengue fever is usually characterized by a sudden onset of fever which last for 2-7 days, together with a variety of signs and symptoms such as high fever, headache, nausea, vomiting, joint pain, fatigue and rashes. The entire stage of illness can be broadly divided into three phases; febrile, critical and recovery phase [31]. During the critical phase, some individuals may progress to severe plasma leakage in cases with dengue hemorrhagic fever (DHF) or dengue shock syndrome (DSS), severe hemorrhage and severe organ involvement. The warning signs for severe dengue include persistent vomiting, mucosal bleeding, lethargy, abdominal pain or tenderness, clinical fluid accumulation, restlessness, liver enlargement of more than 2 cm and an increase in hematocrit count with a rapid decrease in platelet count [31]. Shock ensues when a critical volume of the plasma is lost through leakage. Prolonged shock may lead to coagulopathy and multi-organ failure. Careful monitoring of the blood counts, assessment of hemodynamic status and recognition of warning signs allow early treatment intervention. Appropriate fluid management and supportive care are vital for the recovery of patient. Laboratory diagnosis methods relies on genome and antigen detection as well as serological methods have been widely applied routinely. However, several disadvantages presently limit these methods. For example, antibodies may take approx. 5 days to appear after an infection [31], thus antibody detection can contribute toward epidemic research but is less effective for early diagnosis compared to methods that detect virus or viral proteins directly. Alternative

diagnostic systems that can detect dengue virus or related biomarkers during the early phase of infection with comparable sensitivity, selectivity, diagnosis time as current methods can be operated by non-experts would be non-preferable. Basic test for dengue detection includes complete blood count (CBC), which provides the information of platelets count, WBC, RBC and many other blood components. In this paper, we used blood plasma and blood platelets refractive index as the detection tool for the determination of dengue. Samples from healthy person and infected person from dengue are taken for detection. The Refractive index of normal plasma, infected plasma, normal platelets and infected platelets is 1.35, 1.33, 1.39, and 1.35. All design and simulation are performed in the FDTD tool.

II. DESIGNING OF BIOSENSOR

In 1st design, the biosensor has four Nano cavities. The light propagated inside the structure is continuous wave which has wavelength of 1550 nm is used. The observation points are used on the output port to detect the input wave. In this layout, the structure uses a hexagonal lattice shape with Si rods and air in background wafer. The silicon material used in designing of biosensor have Refractive Index of 3.45 and refractive index of air is 1. Design 1 have radius of 150 nm with lattice constant $a = 600$ nm. We used four types of refractive index in the four Nano cavity. Transmission spectrum is changed according to refractive index and this changed is sense by bio sensor, four Nano cavities are created by reducing a radius from 150nm to 120nm. Fig 1 shows a 2-D photonic crystal liner wave guide nanocavity based bio sensor.




Fig 1: Nano cavity based biosensor layout

In design 2, the biosensor has four micro cavities. The light propagated inside the structure is continuous wave which has wavelength of 1550 nm is used. The observation points are used on the output port to detect the input wave. In this layout, the structure uses a hexagonal lattice shape with Si rods and air in background wafer. The silicon material used in designing of biosensor have Refractive Index of 3.45 and refractive index of air is 1. Design 2 have cavity radius of



Table 1. Design Parameter and its value in Biosensor		
S. No.	Name of Parameter	Value
1	Radius of Si(rod)	150nm
2	Lattice constant	600nm
3	Refractive index of Si	3.45
4	Refractive index of Wafer(air)	1
5	Refractive index of normal Plasma	1.35
6	Refractive index of normal Platelets	1.33
7	Refractive index of injected Platelets	1.39
8	PGB range	1360-2191
9	Polarization	TE

122 μm with lattice constant $a = 600 \mu\text{m}$. In this design, we used four types of refractive index in the four microcavity. Transmission spectrum is changed according to refractive index and this changed is sense by bio sensor four micro cavities are created by reducing a radius from 150nm to 122 μm . Fig 2 shows a 2-D photonic crystal liner wave guide microcavity based bio sensor.




Fig 2: Microcavity based biosensor layout

Fig 3 shows the band diagram of the sensor. The band gap structure is depend on three parameter, refractive index of material, lattice constant, ratio of radius to lattice constant (r/a) [34]. The entire structure consists a one band gap. The band gap range is 1360 nm and 2191 nm. The band gap is calculated by plane wave expansion (PWE) method.




Fig 3: TE band diagram square lattice with defects




Fig 4: Refractive Index of Structure

III. SIMULATION AND RESULTS

The performance of sensor is observed by 2 D finite difference time domain (FDTD) method. An input wavelength of 1550 nm is used at the input port. Fig 5 shows the output transmission spectrum of nanocavity based biosensor.




Fig 5: Output transmission spectrum of nanocavity based biosensor

Table 2. Analysis of design 1 wavelength according to Refractive Index			
S.No.	Refractive Index		Normalized Transmission
1	1.35	Normal plasma	15%
2	1.33	Infected plasma	24%
3	1.39	Normal platelets	29%
4	1.35	Infected platelets	28%

Fig 6 shows the output transmission spectrum of micro cavity based biosensor.




Fig 6: Output transmission of microcavity based biosensor.

**Table 3. Analysis of design 2 wavelength according to Refractive Index**

S.No.	Refractive Index		Normalized Transmission
1	1.35	Normal plasma	30%
2	1.33	Infected plasma	15%
3	1.39	Normal platelets	36%
4	1.35	Infected platelets	15%

Above tables shows the transmission of normal plasma infected plasma, normal platelets and infected platelets and with their respective refractive index in different cavities. The Nano cavity and microcavities are filled according their refractive index and transmission results measured. The design of bio sensor is highly sufficient for sensing various dengue pathogens. This sensor is providing a good accuracy and better transmission and suitable for real time application. As the sensitivity of any biosensor is defined by its Q factor, below tables shows the Q factor of two biosensors and thus result of this Q factor helps us to decide whether nanocavity or micro cavity biosensor is better.

Table 4. Analysis of Quality factor of Nanocavity Sensor

S.No.	Refractive Index	Wavelength ()	Quality Factor
1	1.35 (Normal Plasma)	1.552	154
2	1.33 (Infected Plasma)	1.548	145
3	1.39 (Normal Platelets)	1.550	165
4	1.35 (Infected Platelets)	1.550	145

Table 5. Analysis of Quality factor of Microcavity biosensor

S.No.	Refractive Index	Wavelength ()	Quality Factor
1	1.35 (Normal Plasma)	1.550	153
2	1.33 (Infected Plasma)	1.550	111
3	1.39 (Normal Platelets)	1.549	141
4	1.35 (Infected Platelets)	1.551	128

Quality factor is defined as the ratio of resonant wavelength (λ_0) to the full width at half maximum ($\Delta\lambda$) of the resonator Lorentzian response i.e. $Q = \lambda_0 / \Delta\lambda$ [35]. It is represented by 'Q'.

$$Q = \lambda_0 / \Delta\lambda = \omega_0 / \Delta\omega$$

Where λ_0 is the resonance wavelength and $\Delta\lambda$ is the full width at half maximum (FWHM) response.

IV. CONCLUSION

In this paper, transmission shift biosensor detects dengue in blood plasma and blood platelets using the two dimensional photonic crystal. The designed sensor is very small in size, compact, low cost, less losses and operates at very high speed. Simply by varying the refractive index of material transmission can be obtained. The designed biosensors are based two dimensional four nano cavities and four micro cavities with hexagonal shaped lattice. By observing the results obtained from Q factor we can say that nano cavity structure have better sensitivity as compared to microcavity structure. As nanocavity structure of biosensor reduces the overall size of biosensor, thus reducing the complexity and overload. For both structures 1550 nm wavelength is used at input port. The purpose of this structure is to detect dengue in blood plasma and platelets. The designing and simulation are performed in Finite Difference Time Domain (FDTD).

REFERENCES

- [1] "Flow Sensors and wearable in Biomedical Engineering", from: <http://vista.ir/article/258815>.
- [2] Ajami S, Bagheri-Tadi T. Health information technology and quality of care. J Inform Tech Soft Engg 2013.
- [3] Ajami S, Arzani-Birgani A. The use of telemedicine to treat prisoners. J Inform Tech Soft Engg 2013.
- [4] Ajami S, Teimouri F. Features and application of wearable biosensors in medical care. J Res Med Sci 2015.
- [5] D. Son, J. Lee, S. Qiao, R. Ghaffari, J. Kim, J. E. Lee, C. Song, S. J. Kim, D. J. Lee, S. W. Jun, Nat. Nanotechnol, 2014.
- [6] X. Chen, Small Methods 2017, Y. Qian, X. Zhang, L. Xie, D. Qi, B. K. Chandran, X. Chen, W. Huang, Adv. Mater. 2016, D. Qi, Z. Liu, W. R. Leow, X. Chen, MRS Bull. 2017.
- [7] Ting Wang, Hui Yang, Dianpeng Qi, Zhiyuan Liu, Pingqiang Cai, Han Zhang, and Xiaodong Chen "Mechano-Based Transductive Sensing for Wearable Healthcare", Small micro nano 2018, 1702933.
- [8] Amir mehmood, Adnan Nadeem, Kashif Rizwan, Nadeem Mehmmood, Ahmad Waqas, " A study of wearable Bio-sensor technologies and application in health care". Vol 1, Jan-June 2017.
- [9] Bhargava Teja Nukala, Taro Nakano, Amanda Rodriguez, Jerry Tsay, Jerry Lopez, Tam Q. Nguyen, Steven Zupancic and Donald Y. C. Lie, " Real-Time Classification of Patients with Balance Disorders vs. Normal Subjects Using a Low-Cost Small Wireless Wearable Gait Sensor", Biosensors 2016, bios6040058.
- [10] Wei Gao, Hnin Y. Y. Nyein, Ziba Shahpar, Li-Chia Tai, Eric Wu, Mallika Bariya, Hiroki Ota, Hossain M. Fahad, Kevin Chen and Ali Javey, " Wearable Sweat Biosensors", IEEE 2016, IEDM 61.
- [11] Fu-Kang Wang , Mu-Cyun Tang, Sheng-Chao Su and Tzyy-Sheng Horng, " Wrist Pulse Rate Monitor Using Self-Injection-Locked Radar Technology", Biosensors 2016, bios6040054.
- [12] Giovanni Diraco * ID , Alessandro Leone and Pietro Siciliano, " A Radar-Based Smart Sensor for



- Unobtrusive Elderly Monitoring in Ambient Assisted Living Applications", *Biosensors* 2017, bios7040055.
- [13] Pratibha Sharma, Kimberly Newman, Carlin S. Long, A. J. Gasiewski and Frank Barnes, "Use of Wavelet Transform to Detect Compensated and Decompensated Stages in the Congestive Heart Failure Patient", *Biosensors* 2017, bios7030040.
- [14] Huadong Chen¹, Rong Lou², Yanxiao Chen³, Lili Chen¹, Jingya Lu¹ and Qianqian Dong¹, "Photonic crystal materials and their application in biomedicine", *Drug Delivery*, 24:1, 775-780.
- [15] Zhang Y, Fu Q, Ge J. (2015b). Photonic sensing of organic solvents through geometric study of dynamic reflection spectrum. *Nat Commun* 6:7510.
- [16] Wang H, Tong K, Gu C, et al. (2013). Two dimensional photonic crystal biosensors. *Adv Mater Res* 683:419-24.
- [17] Aly AH, Aghajamali A, Elsayed HA, et al. (2016). Analysis of cutoff frequency in a one-dimensional superconductor-metamaterial photonic crystal. *Physica C Supercond Appl* 528:5-8.
- [18] Alba M, Delalat B, Formentin P, et al. (2015). Silica nanopills for targeted anticancer drug delivery. *Small* 11:4626-31.
- [19] J. D. Joannopoulos, S. G. Johnson, J. N. Winn, et al. *Photonic crystals: molding the flow of light*. 2nd ed., Princeton, 2008.
- [20] M. Notomi. Strong light confinement with periodicity. *Proceedings of the IEEE*, 2011, 99(10): 1768-1779.
- [21] M. Arjmand, R. Talebzadeh. Optical filter based on photonic crystal resonant cavity. *Optoelectronics and Advanced Materials-Rapid Communications*, 2015, 9(1-2): 32-35.
- [22] X. Ge, Y. Shi, S. He. Ultra-compact channel drop filter based on photonic crystal nanobeam cavities utilizing a resonant tunneling effect. *Optics Letters*, 2014, 39(24): 6973-6976.
- [23] J. M. Brosi, C. Koos, L. C. Andreani, et al. High-speed low-voltage electro-optic modulator with a polymer-infiltrated silicon photonic crystal waveguide. *Optics Express*, 2008, 16(6): 4177-4191.
- [24] Y. Gao, R. J. Shiue, X. Gan, et al. High-speed electro-optic modulator integrated with graphene-boron nitride heterostructure and photonic crystal nanocavity. *Nano Letters*, 2015, 15(3): 2001-2005.
- [25] K. Fasihi. High-contrast all-optical controllable switching and routing in nonlinear photonic crystals. *Journal of Lightwave Technology*, 2014, 32(18): 3126-3131.
- [26] K. Cui, Q. Zhao, X. Feng, et al. Thermo-optic switch based on transmission-dip shifting in a double-slot photonic crystal waveguide. *Applied Physics Letters*, 2012, 100(20): 201102(1-4).
- [27] C. Y. Lin, H. Subbaraman, A. Hosseini, et al. Silicon nanomembrane based photonic crystal waveguide array for wavelength-tunable true-time-delay lines. *Applied Physics Letters*, 2012, 101(5): 051101(1-4).
- [28] Y. Zhao, Y. Zhang, Q. Wang. Research advances of photonic crystal gas and liquid sensors. *Sensors and Actuators B-Chemical*, 2011, 160(1): 1288-1297.
- [29] R. V. Nair, R. Vijaya. Photonic crystal sensors: An overview. *Progress in Quantum Electronics*, 2010, 34(3): 89-134. [16] C. Fenzl, T. Hirsch, O. S. Wolfbeis. Photonic crystals for chemical sensing and biosensing. *Angewandte Chemie-International Edition*, 2014, 53(13): 3318-3335.
- [30] W.H.O: Dengue: guidelines for diagnosis, treatment, prevention and control - New edition. W.H.O: Geneva (2009).
- [31] D. J. Gubler: Epidemic dengue/dengue hemorrhagic fever as a public health, social and economic problem in the 21st century. *Trends Microbiol* 10(2), 100-103 (2002).
- [32] K. W. Koh, Y. Kita, C. S.Tang, L. W. Ang, K. Y. Wong, L. James, K. T. Goh and L. C. Ng: The 2005 Dengue Epidemic in Singapore: Epidemiology, Prevention and Control. *Annals of the Academy of Medicine, Singapore* 37(7) (2008).
- [33] Shivam Upadhyay & Vijay Laxmi Kalyani, "Designing of Ring -Resonator based Photonic Pressure Sensor", *Imperial Journal of Interdisciplinary Research (IJIR)*, pp 999-1003.
- [34] A.Lavin, R.Casquel Efficient design and optimization of bio-photonic sensing cells (BICELLS) for label free biosensing, INNBIOD (REF:IPT-2011 -1429-010000) under the Spanish Ministry of Economy and Competitiveness.