

Biocompatible materials for cardiovascular stents



Trinity College Dublin
Coláiste na Tríonóide, Baile Átha Cliath
The University of Dublin

Yann Blake

17334432

4Bio5: Biomechanics module supervised by Dr. David Hoey
Department of Mechanical and Manufacturing Engineering
Trinity College Dublin

December 2020

Abstract

One of the most common interventions performed today is intravascular stenting. A stent is a small tubular device placed into a blood vessel to hold it open. This review covers various aspects of stents from how it is placed into the vessel (e.g. the coronary artery) to the biocompatible materials used (e.g. metal). Mechanical considerations are also being discussed as well as comparison between self-expanding stents and drug eluting stents versus resorbable stents. Other latest research are also covered such as a 'smart stent monitor' based on sensors providing real-time feedback.

Contents

1	Introduction	2
1.1	The variety of stents	2
1.2	Cardiovascular diseases and the need for Stents	2
2	Arteries and Stents Mechanics	3
2.1	From Arteries studies to Stents	3
2.2	Mechanical considerations for Stent design	4
3	Influence of biocompatible materials on stent performances	4
4	Perspectives of improvement	5

1 Introduction

Biomaterials are increasingly used in numerous medical applications such as the treatment of cardiovascular diseases. These are natural or synthetic materials suitable to be introduced into living tissues and designed to interface with specific biological systems. Biomaterials' desired effects includes treatment, augmentation or replacement of biological functions. Artificial heart medical devices and joints implants are some concrete applications of that.[1, 2] Another important application where biomaterials are necessary is intravascular stenting. The Global Vascular Stent Market is indeed expected to grow significantly in the upcoming years with the increasing demand for minimally invasive surgeries, the rise of patients with cardiovascular diseases and overall the aging of the population.[3]

1.1 The variety of stents

Stents are small medical tubes inserted in the body to maintain open an obstructed hollow organ or passageway (more accurately a lumen - which is the cavity within a tube) such as a blood vessel.[4] These are woven (interlaced at right angles), knitted (interlocking loops) or braided (several parts woven together) cylindrical mesh structure made of biocompatible materials usually metals or plastics such as stainless steel, nitinol or chrome-cobalt alloy.[5] Once placed in the blood vessel, often an artery, the stents are expected to restore the flow of the fluid -in this case- blood. Traditional stents are expected to stay in place and last over the long-term. Other types of stent however may have different expected duration: dissolving or biodegradable stents.[6]

Stents have also other applications than extending obstructed blood vessels: these have also been tested for microvascular anastomosis surgery. Microvascular anastomosis is the formation of a connection between two normally distinct very small blood vessels performed under a surgical microscope. It was observed that using a dissolvable stent during this process achieved faster repair of the blood vessels and less dilation at the point of surgery.[7] In addition to these fully dissolvable stents, other stents have been designed to only allow a part of it to dissolve. An expandable stent having a dissolvable por-

tion may have an application in arterial intersections where an aneurysm is formed (abnormal bubble of blood formed adjacent to the blood vessel where the walls had a weak spot). In this case, the stent would be expected to close the aneurysm sac, at the neck point - in the axis of the blood vessel walls. On the other side, the dissolvable portion would be in contact with an activating agent to allow the intersecting blood vessel to flow normally.[8] Dissolvable stents are also used in other lumen such as in the urinary tract[9]

Different stent types based on new materials, methods or solutions now exist. The latest innovations focus on coated, resorbable, and drug-eluting stents. The desired improvements include the release of biological active agents able to control adhesion to the lumen walls, cell differentiation, or vessel tissue development. These new stents also provide new physical-chemical properties and degradation rate. Despite the extensive progress in the research for new stents which will be discussed in this review, no ideal stent yet exists.[10]

1.2 Cardiovascular diseases and the need for Stents

As it was mentioned, stents are often used to open some obstructed lumen, especially blood vessels. Let us then look into reasons why these blood vessels may be obstructed.

The heart is the key organ which keeps blood continually circulating throughout the body. Arteries are the blood vessels which supply oxygen-rich blood to the entire organism. Arteries have much higher pressure than vein blood vessels. The arteries supplying blood to the heart are referred to as coronary arteries. Sometimes, coronary arteries become narrow due to plaque deposits or other phenomena. The plaque deposits consist of an accumulation of cells, fats, excess of cholesterol, other lipids, calcium, cellular debris and other substances. The narrowing of these vessels which is called atherosclerosis is precisely what the stents try to compensate. Atherosclerosis not only occurs in coronary arteries but in many other sites of complex blood vessel geometry: abdominal aorta, iliacs (near the abdomens), femorals (in the legs), popliteals (knees), carotids (neck), and cerebrals (brain).[11, 12] Atherosclerosis is a common disorder of the arteries and an important cause

of Cardiovascular diseases (i.e. heart failure, stroke, high blood pressure and peripheral artery disease - similar to coronary artery disease but excluding the heart and the brain).[13]

The incidence of Atherosclerosis varies from 23% to 74% in various studies,[14, 15, 16, 17, 18] and it is reported to be the responsible for more than 25% of deaths on the Indian subcontinent.[18] These figures clearly demonstrate the importance in understanding, extending research and treating this condition.

Atherosclerosis is detected via electrocardiogram, ultrasound, computerized tomography (CT) scan or magnetic resonance angiography (MRA). Once the point of return is passed and prevention strategies would not work anymore, treatments become necessary. Angioplasty is the most common treatment which consists of opening up the obstructed vessel using a small flexible plastic tube or catheter with a 'balloon' at the end. The balloon is inflated which widens the vessel (Figure 1). The plastic tube is often inserted from femoral artery where a small cut allows it to be inserted. Stent implantation usually occurs at the same time.

2 Arteries and Stents Mechanics

2.1 From Arteries studies to Stents

In order to understand the mechanics of Arteries it is essential to understand their structure. Arteries like all blood vessels are made of three distinct layers also called tunicae. The first, tunica intima lines the inside walls and is very thin. It is a semipermeable membrane with various roles such as pressure regulation facilitator. When it comes to mechanical modelling, it is often neglected because of its small thickness. The second, tunica media is made of smooth muscle cells embedded in an extracellular plexus of elastin and collagen and an aqueous ground substance. The third and outermost layer is the tunica adventitia which consists of a dense network of collagen fibers with scattered fibroblasts, elastin and nerves. It appears to serve as a protective layer which prevents rupture of the vessel when an increase in pressure occurs, since fibers gradually straighten in these conditions.[3]

On the perspective of mechanical behaviors of blood vessels, many observations have been made through various research. The presence of resid-

ual stresses both in axial and circumferential directions was reported. Residual stresses appear to be related to the remodeling of the blood vessel wall which occurs when stress changes.[19] Luminal part (the innermost) is reported to be under compression while the outermost under tension which benefits to growth and remodeling of the vessels. Other characteristics include: anisotropy under load-free configuration (due to different tissue properties); incompressibility (due to the high water content); viscoelastic response of the tissues; hysteresis (dependence on history) under cyclic loads; stress relaxation after sustained deformation; pseudoelastic (behaving differently in loading and unloading). It is also observed that aged arteries are stiffer than younger arteries due to fracture of the elastic laminae (membrane) caused with fatigue in the tunica media - and due to cross linking among collagen fibers in the tunica adventitia during remodeling process.[3]

Beyond the mechanics of arteries mentioned, there are a number of deformations these vessels undergo when subjected to mechanical forces - be it from internal blood flow, contiguous tissue tethering or implanted devices such as stents. Musculoskeletal motions are also widely studied through magnetic resonance angiography imaging to show how these strongly impact arterial deformations: twisting,

shortening and bending angles are observed. These deformations are more significant in older patients due to loss of arterial elasticity. In addition to in vivo




Figure 1: Process of Stent implantation with angioplasty [3]

observations many biomechanical models on twisting, kinking or tortuosity deformations using finite elements methods are used to theoretically predict parameters such as critical loads. Similarly, 3D-model algorithms showed that fracture mechanism on the stent may occur in situations of compression or movement of the blood vessels in supine and fetal positions (of the patient) where maximum hip and knee flexion happen.[20, 21] Many others papers report on modeling algorithms among which the study of physical forces on stented and non-stented femoropopliteal arteries (commonly subject to Atherosclerosis) in patients with peripheral arterial disease based on different leg positions.[22, 23] All these findings provide an essential ground to understand the mechanical behaviors of stents and thus the requirements when designing them.

2.2 Mechanical considerations for Stent design

Mechanical behaviour of balloon-expandable stents (also called expanding stents) have been studied by Dumoulin et al.[24] The metallic P308 Palmaz stent with diamond-shaped cells (once expanded) and made of stainless-steel was tested through several models. One important challenge to deal with when designing these stents is the recoil phenomena.[25] Stent recoil refers to the percentage by which the diameter of the stent decreases between its expanded diameter (from the angioplasty) to its relaxed diameter (either following angioplasty or over the long-term).[26] That is an aspect where stent design firms often focus on with additional connectors on the more proximal end of the stent for more axial strength; or with wider peaks which focus strain to reduce recoil as well as other connector (to the lumen walls) design innovation.[27] Pattern of cells and thickness appeared to also provide less recoil and foreshortening and better flexibility.[28] From Dumoulin et al.'s models it was highlighted that not only expansion and intrinsic recoil is correlated with the stent shape but also its mechanical properties and the external stresses in that location - which leads to most stents being designed 10 to 15 percent larger than the targeted blood vessel. Their crushing modelling showed that buckling is a critical issue in stent design. Comparison between coated stents (with self-adhesive foil to fa-

cilitate embedding to lumen walls) and bare stents showed no change in stiffness. Palmaz stents have the advantage of high radial strength which is recommended for highly resistant plaque obstructions (severe Atherosclerosis) - but its plastic nature makes it less suitable for pulsating or compressing blood vessel sites, with higher permanent collapse risks. Alternatively, these locations would require purely elastic stents allowing reversible deformations and thus avoiding collapse.[24]

3 Influence of biocompatible materials on stent performances

Let us extend on the advantages between balloon-expandable stents which Palmaz belongs to and other types of stents, as well as comparing the various biocompatible materials used.

Before recoil even becomes a challenge, the ease of deploying the stent is essential as well as its resistance to dislodgement. For metallic stent such as the Palmaz, it requires a higher yield point value or lower modulus of elasticity. Stainless steel which is widely used achieves moderately well for yield point and radial force. However the modulus of elasticity (200 GPa) is not low, which can only be compensated with design shapes. Another metal used, Nitinol (Nickel titanium alloy) is more complex in that regards. The nickel-to-titanium ratio and the high-temperature heat treatment influences the thermal memory and the temperature at which the crystalline structure changes - which consequently changes the material flexibility and malleability. Ideally, high flexibility and malleability are desired when the stent is inserted in the implantation system (tube) at room temperature. Then, low flexibility with high rigidity (achieved with austenite crystalline structure) is desired when the stent is released in the blood vessel at around 30 degrees Celsius. These key temperature points therefore play a major role in stent performances for the nitinol material, unlike stainless steel. Dyet et al. demonstrate two distinct groups of metallic stents in terms of mechanical properties: those affected by their construction and those directly affected by the metal properties. Overall they consider high radial strength, high radio-opacity, elasticity, flexibility, good trackability as being the ideal

properties of a stent ; although none of the stents they tested or researched on (in 2000) met these criteria. The necessity of assessment of metal fatigue is also stated, since several incidents of stents breaking up occurred when submitted to repeated flexing.[29] Since 2000, many other biocompatible materials appeared and were widely tested.

Resorbable stent also called biodegradable stent which are manufactured from a material capable of dissolving or being absorbed in the organism, are becoming increasingly popular since these avoid many risks and long-term issues - permanent stents bring. A zinc-copper biodegradable metal stent developed by Zhou et al. in 2019 achieved less intrinsic elastic recoil than standard stainless steel stents; better trackability (less push force when passing through curved vessels thus reducing mechanical stimulation on them). Their stent also facilitates the recovery of vascular pulsatility (difference between blood systolic and diastolic velocities).[30] Choubey et al. performed Finite element analysis to compare predicted Von-Mises Tress, recoil, Fracture point, Ultimate Tensile strength and factor of safety for seven stent materials: Stainless Steel 316L, Cobalt Chromium L-605 Stent, biodegradable Stent (PCL), Nitinol Stent (Austenite), Elgiloy Stent, Tantalum Stent, Cobalt Chromium. Biodegradable stents made of polycaprolactone (PCL) polymer had the most significant recoil making it less ideal than the other materials. Its Ultimate Tensile strength was also the lowest. Tantalum (a transition metal) achieved poor performances too. Overall Cobalt Chromium L-605 and Nitinol (with good corrosion resistance) have the best balance and make better material for stent.[31] Other research support the use of cobalt-chromium stent.[32] Biodegradable stents remain at the center of innovation, in addition to zinc-copper or PCL, some use Fe-Mn alloys.[33]

Drug eluting stents are another important segment of research, improving standard or coated stents. Stent coatings can control biocompatibility, degradation rate, protein adsorption, and adequate formation of endothelial tissue. These are parameters making stents perform better. Coating materials are either organic polymers, biological components or inorganic coatings (e.g. nitrides). Drug eluting stents aim to reduce the overproduction of tissue in the stent site which would cause obstruction again. Antiproliferative substances 'drugs' are therefore used such as pa-

clitaxel or limus.[10]

4 Perspectives of improvement

In 2020, researchers from ETH Zurich developed a high-level precision, biocompatible and 3D-printable micro-robot which delivers drug in the organism through the blood vessels. They are now investigating the use of such micro-robots for implementation of surgical tools especially stents.[34] The same university also introduced a new technology based on '4-D printing' which allows stents to be up to forty times smaller than existing ones.[35]

To alert potential recoil, pressure monitors placed on the stent - at both ends - have been developed which informs the patient or doctor when there is a drop in blood pressure across the stent. The information is transmitted wirelessly to a reader placed on the body close to the stent site. These inexpensive small scale technologies aim to better monitor the evolution of the stents over time.[36, 37, 38]

When it comes to the blood vessel conditions (such as atherosclerosis) in the coronary artery, several research show opposing findings, conclusions and recommendations between stenting or bypass surgery. Some research demonstrate that coronary artery bypass surgery may be the best treatment option for the majority of patients having more than one blocked heart artery instead of placing stents. This highlights the need for further research and improvement of stents which may allow patients to avoid open heart surgery while obtaining the same treatment performance.[39]

The motivations of innovation is often to overcome the current limitations of stents, such as restenosis (the recurrence of narrowing of a blood vessel, sometimes from recoil), or inflammation and thrombosis (formation of blood clot) at the stent site. Other characteristics are also often desired such as biodegradability and some mechanical parameters based on the targeted blood vessel. Finding the ideal design and material doesn't suffice, when biocompatibility remains the most important constraint when designing a stent. Promising stent solutions are currently in development, the numerous stent patents filled in recent years can testify this.[10] Clinical trials and market approval requests from companies have also been rising.[35]

References

- [1] Roger Narayan. *Encyclopedia of Biomedical Engineering*. Elsevier, 2018.
- [2] Ekta Pandey, Keerti Srivastava, Saurabh Gupta, Suravi Srivastava, and Nidhi Mishra. Some biocompatible materials used in medical practices-a review. *International journal of pharmaceutical sciences and research*, 7(7):2748–2755, 2016.
- [3] Aleksandra Fortier, Vikranth Gullapalli, and Reza A Mirshams. Review of biomechanical studies of arteries and their effect on stent performance. *IJC Heart & Vessels*, 4:12–18, 2014.
- [4] D Aibibu, M Hild, and C Cherif. An overview of braiding structure in medical textile: fiber-based implants and tissue engineering. In *Advances in Braiding Technology*, pages 171–190. Elsevier, 2016.
- [5] S. Eriksson and L. Sandsjö. 12 - three-dimensional fabrics as medical textiles. In Xiaogang Chen, editor, *Advances in 3D Textiles*, Woodhead Publishing Series in Textiles, pages 305 – 340. Woodhead Publishing, 2015.
- [6] Mark B Detweiler. Dissolvable anastomosis stent and method for using the same, August 25 1992. US Patent 5,141,516.
- [7] Zhang Cong, Tang Nongxuan, Zheng Changfu, Xu Yuanwei, and Wang Tongde. Experimental study on microvascular anastomosis using a dissolvable stent support in the lumen. *Microsurgery*, 12(2):67–71, 1991.
- [8] Donald K Jones. Expandable stent having a dissolvable portion, December 12 2006. US Patent 7,147,659.
- [9] Brian K Auge, Roberto F Ferraro, Arthur R Madenjian, and Glenn M Preminger. Evaluation of a dissolvable ureteral drainage stent in a swine model. *The Journal of urology*, 168(2):808–812, 2002.
- [10] Natalia Beshchasna, Muhammad Saqib, Honorata Kraskiewicz, Łukasz Wasyluk, Oleg Kuzmin, Oana Cristina Duta, Denisa Ficai, Zeno Ghizdavet, Alexandru Marin, Anton Ficai, et al. Recent advances in manufacturing innovative stents. *Pharmaceutics*, 12(4):349, 2020.
- [11] N Majewska, MA Blaszkak, R Juszkat, M Frankiewicz, M Makalowski, and W Majewski. Patients' radiation doses during the implantation of stents in carotid, renal, iliac, femoral and popliteal arteries. *European Journal of Vascular and Endovascular Surgery*, 41(3):372–377, 2011.
- [12] K Amosova, O Iaremenko, I Matiyashchuk, P Minchenko, and N Makomela. Thu0333 frequency and nature of atherosclerotic damage of arteries in systemic lupus erythematosus. *Annals of the Rheumatic Diseases*, 72(Suppl 3):A278–A278, 2013.
- [13] J Tario and P Wallace. *Pathobiology of human disease*, 2014.
- [14] Kataria Sant Prakash Garg Monika, Aggarwal Akash Deep. Coronary atherosclerosis and myocardial infarction an autopsy study. *Journal of Indian Academy of Forensic Medicine*, 33(1):39–42, 2011.
- [15] Dhruva GA, AH Agravat, and HK Sanghvi. Atherosclerosis of coronary arteries as predisposing factor in myocardial infarction: An autopsy study. *Online Journal of Health and Allied Sciences*, 11(3 (1)), 2012.
- [16] J Golshahi, P Rajabi, and F Golshahi. Frequency of atherosclerotic lesions in coronary arteries of autopsy specimens in isfahan forensic medicine center. 2005.

- [17] YAZDI SA TABATABAEI, AR Rezaei, AZARI J BORDBAR, Aria Hejazi, Mohammad Taghi Shakeri, and SHAHRI M KARIMI. Prevalence of atherosclerotic plaques in autopsy cases with noncardiac death. 2009.
- [18] Priti Vyas, Ratigar Narangar Gonsai, Charu Meenakshi, and Meeta G Nanavati. Coronary atherosclerosis in noncardiac deaths: An autopsy study. *Journal of Mid-life Health*, 6(1):5, 2015.
- [19] Yuan Cheng Fung. What are the residual stresses doing in our blood vessels? *Annals of biomedical engineering*, 19(3):237–249, 1991.
- [20] Christopher P Cheng, Nathan M Wilson, Richard L Hallett, Robert J Herfkens, and Charles A Taylor. In vivo mr angiographic quantification of axial and twisting deformations of the superficial femoral artery resulting from maximum hip and knee flexion. *Journal of vascular and interventional radiology*, 17(6):979–987, 2006.
- [21] Jose A Diaz, Miguel Villegas, Gustavo Tamashiro, Marisa H Miceli, Daniel Enterrios, Aristobulo Balestrini, and Alberto Tamashiro. Flexions of the popliteal artery: dynamic angiography. *December*, 16:12, 2004.
- [22] Nigel B Wood, Shun Z Zhao, Andrew Zambanini, Mark Jackson, W Gedroyc, Simon A Thom, Alun D Hughes, and Xiao Yun Xu. Curvature and tortuosity of the superficial femoral artery: a possible risk factor for peripheral arterial disease. *Journal of applied physiology*, 101(5):1412–1418, 2006.
- [23] Andrew J Klein, Ivan P Casserly, John C Messenger, John D Carroll, and S-Y James Chen. In vivo 3d modeling of the femoropopliteal artery in human subjects based on x-ray angiography: Methodology and validation. *Medical physics*, 36(2):289–310, 2009.
- [24] C Dumoulin and B Cochelin. Mechanical behaviour modelling of balloon-expandable stents. *Journal of Biomechanics*, 33(11):1461 – 1470, 2000.
- [25] Benjamin Blais, Karen Carr, Sanjay P Sinha, Morris M Salem, and Daniel S Levi. Mechanical properties of low-diameter balloon expandable covered stents. *Catheterization and cardiovascular interventions : official journal of the Society for Cardiac Angiography amp; Interventions*, December 2020.
- [26] Niels Grabow, Carsten M Bünger, Katrin Sternberg, Steffen Mews, Kathleen Schmohl, and Klaus-Peter Schmitz. Mechanical properties of a biodegradable balloon-expandable stent from poly (l-lactide) for peripheral vascular applications. 2007.
- [27] Unknown author. Rebel stent radial and axial strength. <https://www.bostonscientific.com/en-US/products/stents-coronary/rebel-platinum-chromium-coronary-stent-system/rebel-stent-radial-axial-strength.html>, 2015.
- [28] Dong Bin Kim, Hyuk Choi, Sang Min Joo, Han Ki Kim, Jae Hee Shin, Min Ho Hwang, Jaesoon Choi, Dong-Gon Kim, Kwang Ho Lee, Chun Hak Lim, et al. A comparative reliability and performance study of different stent designs in terms of mechanical properties: foreshortening, recoil, radial force, and flexibility. *Artificial organs*, 37(4):368–379, 2013.
- [29] John F Dyet, William G Watts, Duncan F Ettles, and Anthony A Nicholson. Mechanical properties of metallic stents: how do these properties influence the choice of stent for specific lesions? *Cardiovascular and interventional radiology*, 23(1):47–54, 2000.

- [30] Chao Zhou, Xiangyi Feng, Zhangzhi Shi, Caixia Song, Xiaoshan Cui, Junwei Zhang, Ting Li, Egon Steen Toft, GE Junbo, Luning Wang, et al. Research on elastic recoil and restoration of vessel pulsatility of zn-cu biodegradable coronary stents. *Biomedical Engineering/Biomedizinische Technik*, 65(2):219–227, 2020.
- [31] Rahul Kumar Choubey and Sharad K Pradhan. Prediction of strength and radial recoil of various stents using fe analysis. *Materials Today: Proceedings*, 27:2254–2259, 2020.
- [32] Avinash Kumar and Naresh Bhatnagar. Finite element simulation and testing of cobalt-chromium stent: a parametric study on radial strength, recoil, foreshortening, and dogboning. *Computer Methods in Biomechanics and Biomedical Engineering*, pages 1–15, 2020.
- [33] Hendra Hermawan, Dominique Dubé, and Diego Mantovani. Degradable metallic biomaterials: design and development of fe–mn alloys for stents. *Journal of Biomedical Materials Research Part A: An Official Journal of The Society for Biomaterials, The Japanese Society for Biomaterials, and The Australian Society for Biomaterials and the Korean Society for Biomaterials*, 93(1):1–11, 2010.
- [34] PAUL HANAPHY. Zurich scientists develop 3d printed microbots for drug delivery inside the human body. *3D Printing Industry*, 2020.
- [35] SALONI WALIMBE. Innovations in cardiovascular disease treatment and the rising demand for stents. *DAIC Diagnostic and Interventional Cardiology*, 2020.
- [36] Xing Chen, Babak Assadsangabi, Daniel Brox, York Hsiang, and Kenichi Takahata. A pressure-sensing smart stent compatible with angioplasty procedure and its in vivo testing. In *2017 IEEE 30th International Conference on Micro Electro Mechanical Systems (MEMS)*, pages 133–136. IEEE, 2017.
- [37] Xing Chen, Babak Assadsangabi, York Hsiang, and Kenichi Takahata. Enabling angioplasty-ready “smart” stents to detect in-stent restenosis and occlusion. *Advanced Science*, 5(5):1700560, 2018.
- [38] Betsy DM Chaparro-Rico, Fabio Sebastiano, and Daniele Cafolla. A smart stent for monitoring eventual restenosis: Computational fluid dynamic and finite element analysis in descending thoracic aorta. *Machines*, 8(4):81, 2020.
- [39] TL Braber, RS Hermanides, and JP Ottervanger. Coronary stenting versus bypass surgery in elderly with multivessel disease: long-term mortality rate is still up for debate. *Netherlands Heart Journal*, 28(12):678–679, 2020.

Further Reading

- [A]Khalilimeybodi, A., Khoei, A.A. and Sharif-Kashani, B., 2019. Future Balloon-Expandable Stents: High or Low-Strength Materials?. *Cardiovascular Engineering and Technology*, pp.1-17.
- [B]Beier, S., Ormiston, J., Webster, M., Cater, J., Norris, S., Medrano-Gracia, P., Young, A. and Cowan, B., 2016. Hemodynamics in idealized stented coronary arteries: important stent design considerations. *Annals of biomedical engineering*, 44(2), pp.315-329.
- [C]Olivier F Bertrand, Rajender Sipehia, Rosaire Mongrain, Josep Rodés, Jean-Claude Tardif, Luc Bilodeau, Gilles Côté, Martial G Bourassa, Biocompatibility aspects of new stent technology, *Journal of the American*

College of Cardiology, Volume 32, Issue 3, 1998, Pages 562-571,

[D]Wenwang Wu, Xiaoke Song, Jun Liang, Re Xia, Guian Qian, Daining Fang, Mechanical properties of anti-tetrachiral auxetic stents, Composite Structures, Volume 185, 2018, Pages 381-392,