




# Challenges of calibrations in the mid-infrared with VISIR

Konrad R. W. Tristram  
VISIR instrument scientist

# Overview

- Part 1: Description of the Instrument
- Part 2: Calibrations in the mid-infrared


VISIR on the Cassegrain focus of UT3 and the recommissioning team



# Overview


- Part 1: Description of the Instrument
- Part 2: Calibrations in the mid-infrared

VISIR on the Cassegrain focus of UT3 and the recommissioning team



# VISIR: General overview


- VISIR = VLT spectrometer & imager for the mid-IR
- Imaging & spectroscopy between  $5\mu\text{m}$  and  $20\mu\text{m}$
- Built by CEA/DAPNIA/SAP and NFRA/ASTRON in 2004, upgraded by ESO in 2014
- Located at UT3 Cassegrain
- Three main components:
  - imager
  - spectrometer
  - calibration unit



# VISIR: the imager


## ■ Three components:

- collimator (M1 & M2)
- cold stop & filter
- objectives (M3 – M5)



# VISIR: the imager


- 26 filters between 5 and 20 $\mu$ m
- 0.045 (SF) & 0.076 (IF) pixel scale
- AGPM & 4QPM coronagraphy; SAM



# VISIR: the spectrometer

## ■ technical setup: two arms

- LR & MR prism / grating
- HR gratings



# VISIR: the spectrometer


## ■ Several resolution modes:

grating	offered wavelengths [ $\mu\text{m}$ ]	resolving power
low resolution (LR)	8 – 13 $\mu\text{m}$	$R \sim 350$ at 10 $\mu\text{m}$
medium resolution (MR)	7.5 – 9.3 $\mu\text{m}$ , 10.2 – 13.0 $\mu\text{m}$ , 17.1 – 19.0 $\mu\text{m}$	$R \sim 3200$ at 10 $\mu\text{m}$
high resolution - long slit (HR)	8.02, 12.81 and 17.03 $\mu\text{m}$	$R \sim 25000$ at 10 $\mu\text{m}$
HR - cross dispersed (HRX)	7.7 – 24.0 $\mu\text{m}$	$R \sim 25000$ at 10 $\mu\text{m}$

# VISIR: detectors


## ■ Two AQUARIUS $1024 \times 1024$ detectors

- 110 Hz frame rate
- excess low frequency noise  $\Rightarrow$  fast chopping




# VISIR: detectors

- Very good cosmetics:  
(example: LR spectro)




- Bleeding effect for bright sources:



# Overview


- Part 1: Description of the Instrument
- Part 2: Calibrations in the mid-infrared



VISIR on the Cassegrain focus of UT3 and the recommissioning team

# Observing in the MIR: transparency

■ Atmosphere transparent in three MIR bands:




# Observing in the MIR: background

■ High background due to:

- atmosphere
- telescope
- instrument


→ 4000Jy/arcsec<sup>2</sup>



Kendrew et al. 2010


# Observing in the MIR

## ■ Ambient quantities and their influence



# Chopping / nodding

- Chopping with the M2 up to 4 Hz
- Nodding with telescope at  $\sim 0.01$  Hz




# Chopping / Nodding

## ■ Two chopping modes:

- parallel
- perpendicular

## ■ Jittering also possible



# Chopping / Nodding

## ■ Issues:




- Fast vs. slow chopping (ELFN noise)
- Efficiency (settle time)
- Accuracy in positioning
- Nodding & chopping overlap for parallel chopping
- Field stabilisation – M2 interaction

## ■ Alternatives:

- Dicke switch (VISIR-AO)
- 3 point chopping (Pietrow et al. 2016)

# Image quality

- VISIR very sensitive to image quality problems
- Significant improvements by:
  - ALT-axis calibration
  - Separate controller for the 2 guiding windows when chopping




# Calibration plan

- Data self consistent (no bias / dark / flat)
- Darks only observed to verify detector health
- Observatory standards for IMG & LR spec.
  - Status verified by HealthChecker
- Static calibration data:
  - Standard stars:  
based on Cohen et al.  
426 imaging / 469 spectroscopy
  - Detector Quantum Efficiency
  - Atmospheric Emission Spectrum

DATE*: [?]		2017-01-09	2017-01-10	daytime calibs:	(may require OB grade review) [?]	Calibration action?
Raw CAL displays: [?]		SM 86 report [NLT]	SM 84 report [NLT]			
Product quality: [?]		raw	raw			
Data types:	Setup:					
SCI_IMG_CHNJ	B10.7_0.0453		ok			all ok
SCI_IMG_RCHN	J7.9_0.0453	ok	ok			all ok
	M-BAND_0.0453	ok	ok			all ok
	NEII_1_0.0453	ok	ok			all ok
	NEII_2_0.0453	ok	ok			all ok
	PAH1_0.0453	ok	ok			all ok
	Q1_0.0453	ok	ok			all ok
	Q2_0.0453	ok	ok			all ok
	Q3_0.0453	ok	ok			all ok
	SIV_2_0.0453	ok	ok			all ok
SCI_IMG_CHNB	M-BAND_0.0453		ok			all ok
SCI_SPEC_CHN	0.750_OPEN					all ok
SCI_SPEC_ECH	0.400_N_LW_GR	ok				all ok


# HR wavelength calibration

- Use atmospheric emission spectrum
- Cross correlation of observed spectrum with HITRAN model
- $\sim 1\text{\AA}$  ( $\sim 2 \text{ km/s}$ , based on a single line)



# Water vapour radiometer


- Water vapour monitor installed as part of the VISIR upgrade: RPG-LHATPRO



# Summary

## ■ Instrument description:

- M, N, Q-band imager
- coronagraphy, SAM
- LR prism, HR Echelle spectroscopy



Jupiter with VISIR (ESO PR eso1623b)

## ■ Calibrations:

- High background chopping / nodding
- Flux calibration procedures
- VISIR demanding w.r.t. image quality